



**JAWAHARLAL INSTITUTE OF POST-GRADUATE MEDICAL EDUCATION & RESEARCH**  
(Institution of National Importance Under Ministry of Health & Family Welfare, Government of India)  
Dhanvantari Nagar, Pondicherry-605 006

Tele: JIPMER  
E.mail: purchase@jipmer.edu.in  
Website: www.jipmer.edu.in

Phone: 0413-2272380 to 2272389  
Fax : 0413-2272067/66  
Purchase Section:0413-2296016



Pur.12 (11)ASV/OeT/2020

Dated: 07.11.2020

## **CORRIGENDUM**

**THE BELOW MENTIONED CORRIGENDUM REFERS TO THIS OFFICE TENDER NO. Pur.12 (11) ASV /OeT /2020 DATED: 29.10.2020: -**

In Tender Published vide Tender Notice no: JIP. Pur.12 (11)ASV/OeT/2020 Dated: 29.10.2020. The following Amendments are made in the Specification of the equipment.

<b>Tender specification Point No/ Page no</b>	<b>Read as Specification in tender document on Section VIII</b>	<b>Amendment as in Specification in tender document on Section VIII</b>
<b>Point no: 5 Page no: 51</b>	Strong Motor with at least 1300 RPM	<b>Delete</b>
<b>Point no: 8 Page no: 51.</b>	High Speed Oscillation 12000 to 21000 OPM	High Speed Oscillation within 12000 to 21000 OPM range
<b>Point no: 9 Page no: 51.</b>	Speed control facility to adjust the speed of the saw blade	<b>Delete</b>
<b>Point no: 16 Page no: 51.</b>	Air Flow Rate: 57 L/min	Air Flow Rate: 90 L/min
<b>Point no: 17 Page no: 51.</b>	Vacuum: 235 mbar	Vacuum: 1500 to 1650 mm WG
<b>Point no: 19 Page no: 51.</b>	Container Volume 35 Litre	Container Volume 7 to 15 Litre
<b>Point no: 3 Page no: 49.</b>	<b>Consumables required with the Machine</b> 3.01 Tube: 5, 3.02 Hose clamp: 5, 3.03 saw blades: 5	These are extra consumable other than full set of autopsy saw with vacuum dust collector (including two autopsy saw blades). Kindly quote for extra consumable i.e. Tube: 5, Hose clamp 5, saw blades 5

Officer – In – Charge of Purchase  
JIPMER, Puducherry