

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No.JIP/BURNS/ERRC/2015/9

Dated:06.02.2015

To

Sir,

Sub: Supply of Bipolar Machine- Limited tender- Invited – Regarding

Please quote your lowest rates for supply of the Bipolar Machine -2Nos as per the list attached in Annexure-1, subject to the following terms and conditions:

1. Rates should be quoted only for the items which are available in stock and can be supplied immediately on receipt of order.
2. Only 5% CST is applicable
3. No insurance charges are payable as per the rules in the Government. As such, the firms before quoting should take into consideration all the risks in the transit and then furnish Limited Tender which should cover insurance charges also. If any point is raised as regards insurance charges after orders are issued, the same will not be entertained and the firms thereafter should effect the supply at their own cost.
4. Rates should be quoted F.O.R. Pondicherry, Extra packing; forwarding charges etc. should not be quoted.
5. Delivery is required urgently. Tenders should please state the guarantee delivery period they can offer. As delivery date is essential for the contract, this should be strictly adhered to by the successful tenders.
6. No supply which is not according to the specifications and not meeting our requirement will be accepted.
7. The Director shall have the right of rejecting the Limited Tender in whole or part without assigning any reason therefore.
8. In case of high precision instruments the firms should give a guarantee certificate for their satisfactory performance.
9. Limited Tenders should be sent only by Registered Post. Hand Quotations are not accepted. Please furnish your Limited.Tender in a sealed cover super scribing as:

Limited Tender for EQUIPMENT Bipolar Machine

Enquiry No. : No.JIP/BURNS/ERRC/2015/9 Dated 06.02.2015

Due Date: 20.02.2015.

Limited Tender should reach this office on or before 20.02.2015.

Limited Tender received after the due date will be summarily rejected.

- NOTE:**
1. Quotation may please be address to:
The Director, JIPMER, Puducherry-6.
 2. The Sealed Cover (quotation) may please be sent directly to:
The Head, Dept. of Plastic Surgery, S.S. Block,
IV Floor, JIPMER, Puducherry-6.

Yours faithfully,

For DIRECTOR

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No.JIP/BURNS/ERRC/2015/9

DUE DATE: 20.02.2015

Dated:06.02.2015

NAME OF THE EQUIPMENT: BIPOLAR MACHINE -2Nos

SPECIFICATION:

- Should be an **advanced microprocessor based electro surgical unit** with **digital display** for all power settings.
- Should have facility **Under water cutting** Diathermy
- Should have **facility to operate without** stabilizer
- Should have Memory of previous activated power settings – last activated power settings are available on “Power On”
- The machine should have a facility for power setting in **Bipolar Mode** – Bipolar cut 70 watts, Force 70 watts
- The machine should have a facility for power setting in **Monopolar Cut** – Pure cut 200 watts, Low cut 200 watts, Blend cut 100 watts
- The machine should have a facility for power setting in **Monopolar Coagulation** – Fulgurate 100 watts, Desiccate 100 watts
- **Independent Mono-polar & Bipolar Output** Can be used **without Any Switch over from the Machine** :- This facility Enables the operating surgeon to perform surgery without any Hassle.
- Should have **Patient Plate Contact Quality Monitoring system** - With This Silicon Dual Type Patient Plate Contact Quality Monitoring Takes Place , the moment the contact between Plate & Patient reduces it stops the HF delivery with an audio visual indications
- Should have **Time out function** facility in both Foot Switch & Hand Pencil to avoid the continuous delivery of power
- Should have **3 – Pedal explosion proof / water proof Foot Switch**
- Should have Audible / visual indication on cutting & coagulation
- Should have **self test** facility provided
- Should construction in accordance with **IEC 60601 & IEC 60601-2-2**
- Should have **Safety standards certificates like CE or S-Mark or FDA**
- Should have **ISO 13485:2003** certificate

Standard Accessories:

- | | |
|---|-------|
| 1. Three Pedal Foot Switch with Autolock plug | - 1no |
| 2. Bipolar forceps with silicon Cable | - 1no |
| 3. Bipolar forceps non – stick | - 1no |
| 4. Silicon cable for Bipolar Forceps | - 1no |
| 5. Monopolar handle & four electrodes | - 1no |
| 6. Handswitch pencil - Disposable | - 1no |
| 7. Silicon patient plate Single type with cable | - 1no |
| 8. Trolley | - 1no |

Note:

*** Demonstration of the equipment is essential**

*** Warrantee Three Years**

*** AMC for next five years**

***Please quote preferably in Indian Rupees.**

(DR.RAVIKUMAR CHITTORIA)