

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No.JIP/BURNS/ERRC/2015/9

Dated:05.02.2015

To

Sir,

Sub: Supply of Defibrillator with Crash Trolley- Limited tender- Invited – Regarding

Please quote your lowest rates for supply of the Defibrillator with Crash Trolley as per the list attached in Annexure-1, subject to the following terms and conditions:

1. Rates should be quoted only for the items which are available in stock and can be supplied immediately on receipt of order.
2. Only 5% CST is applicable
3. No insurance charges are payable as per the rules in the Government. As such, the firms before quoting should take into consideration all the risks in the transit and then furnish Limited Tender which should cover insurance charges also. If any point is raised as regards insurance charges after orders are issued, the same will not be entertained and the firms thereafter should effect the supply at their own cost.
4. Rates should be quoted F.O.R. Pondicherry, Extra packing; forwarding charges etc. should not be quoted.
5. Delivery is required urgently. Tenders should please state the guarantee delivery period they can offer. As delivery date is essential for the contract, this should be strictly adhered to by the successful tenders.
6. No supply which is not according to the specifications and not meeting our requirement will be accepted.
7. The Director shall have the right of rejecting the Limited Tender in whole or part without assigning any reason therefore.
8. In case of high precision instruments the firms should give a guarantee certificate for their satisfactory performance.
9. Limited Tenders should be sent only by Registered Post. Hand Quotations are not accepted.

Please furnish your Limited.Tender in a sealed cover super scribing as:

Limited Tender for EQUIPMENT Defibrillator with Crash Trolley

Enquiry No. : No.JIP/BURNS/ERRC/2015/9 Dated 05.02.2015

Due Date: 19.02.2015.

Limited Tender should reach this office on or before 19.02.2015.

Limited Tender received after the due date will be summarily rejected.

- NOTE:**
1. Quotation may please be address to:
The Director, JIPMER, Puducherry-6.
 2. The Sealed Cover (quotation) may please be sent directly to:
The Head, Dept. of Plastic Surgery, S.S. Block,
IV Floor, JIPMER, Puducherry-6.

Yours faithfully,

For DIRECTOR

ANNEXURE-I

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No..JIP/BURNS/ERRC/2015/9

DUE DATE: 19.02.2015

Dated:05.02.2015

NAME OF THE EQUIPMENT: DEFIBRILLATOR WITH CRASH TROLLEY

SPECIFICATION:

- Light weight, portable and rugged Bi Phasic defibrillator
- Should have a high resolution LCD display of 8.4”.
- Graphical Display of waveform and numeric display of HR.
- Should have facility for Synchronised & manual defibrillation.
- Facility of ECG pickup from paddles in case of ECG electrodes connected are not connected to the Defibrillator.
- Adult and paediatric paddles should be integrated.
- Selection of energy: 2 to 200 joules.
- The alarm limit setting should be independent and continuous for Low and High alarms with provision for up-down counting.
- The internal Safety discharge facility.
- Battery type: Lead Acid Battery, 12V 4.5 AH
- Battery Charging Time: 12 hours from zero charge.
- Battery Capacity: 3 -4 hours of continuous monitoring.
- Encoder for easy menu selection
- Auto shutdown on low battery charge
- Mains / Battery charge indication on LCD as well as front panel.
- CE Approved.

Optional Features:

- Integrated thermal Printer
- Chart paper [Thermoreactive Z fold 80mm]
- External Pacemaker
- NIBP
- SpO2 [Nellcor]
- USB for data storage
- AED

Note:

- * Demonstration of the equipment is essential
- * Warrantee Three Years
- * AMC for next five years
- *Please quote preferably in Indian Rupees.

(DR.RAVIKUMAR CHITTORIA)