

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No.JIP/BURNS/ERRC/2015/9

Dated:07.02.2015

To

Sir,

Sub: Supply of Power Dermatome- Limited tender- Invited – Regarding

Please quote your lowest rates for supply of the Power Dermatome– 2Nos as per the list attached in Annexure-1, subject to the following terms and conditions:

1. Rates should be quoted only for the items which are available in stock and can be supplied immediately on receipt of order.
2. Only 5% CST is applicable
3. No insurance charges are payable as per the rules in the Government. As such, the firms before quoting should take into consideration all the risks in the transit and then furnish Limited Tender which should cover insurance charges also. If any point is raised as regards insurance charges after orders are issued, the same will not be entertained and the firms thereafter should effect the supply at their own cost.
4. Rates should be quoted F.O.R. Pondicherry, Extra packing; forwarding charges etc. should not be quoted.
5. Delivery is required urgently. Tenders should please state the guarantee delivery period they can offer. As delivery date is essential for the contract, this should be strictly adhered to by the successful tenders.
6. No supply which is not according to the specifications and not meeting our requirement will be accepted.
7. The Director shall have the right of rejecting the Limited Tender in whole or part without assigning any reason therefore.
8. In case of high precision instruments the firms should give a guarantee certificate for their satisfactory performance.
9. Limited Tenders should be sent only by Registered Post. Hand Quotations are not accepted. Please furnish your Limited.Tender in a sealed cover super scribing as:

Limited Tender for EQUIPMENTPower Dermatome

Enquiry No. : No.JIP/BURNS/ERRC/2015/9 Dated 07.02.2015

Due Date: 21.02.2015

Limited Tender should reach this office on or before 21.02.2015.

Limited Tender received after the due date will be summarily rejected.

- NOTE:**
1. Quotation may please be address to:
The Director, JIPMER, Puducherry-6.
 2. The Sealed Cover (quotation) may please be sent directly to:
The Head, Dept. of Plastic Surgery, S.S. Block,
IV Floor, JIPMER, Puducherry-6.

Yours faithfully,

For DIRECTOR

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No.JIP/BURNS/ERRC/2015/9

DUE DATE: 21.02.2015

Dated:07.02.2015

NAME OF THE EQUIPMENT: POWER DERMATOME -2NOS

SPECIFICATION:

- Should deliver smooth cutting action that produces clean, uniform grafts.
- With 5,000 – 6,000 cycles/minutes high speed cutting action.
- Precise cut thickness up to .030 inch (0.76 mm) in easily adjustable .002 inch (0.05 mm) increments.
- Width plate range from one to four inches (2.5 to 10 cm) in one inch increments.
- Safety lever to help prevent accidental activation.
- Blades should be supplied sterile without any need for lubrication.
- Electric Dermatome should have power source with 4.3 amp capability.
- Dematome Screw Driver.
- Should have Safety Standard Certificates like CE or FDA
- Should be supplied with a trolley to keep the machine

Note:

* Demonstration of the equipment is essential

* Warrantee Three Years

* AMC for next five years

*Please quote preferably in Indian Rupees.

(DR.RAVIKUMAR CHITTORIA)