

**DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY-605006**

No.JIP/PLS/ERRC/2014/10

Dated:29.12.2014

To

Sir,

**Sub: Supply of USB Video Dermatoscope with Polarised Light- Limited tender- Invited –
Regarding**

Please quote your lowest rates for supply of the USB Video Dermatoscope with Polarised Light as per the list attached in Annexure-1, subject to the following terms and conditions:

1. Rates should be quoted only for the items which are available in stock and can be supplied immediately on receipt of order.
2. Only 5% CST is applicable
3. No insurance charges are payable as per the rules in the Government. As such, the firms before quoting should take into consideration all the risks in the transit and then furnish Limited Tender which should cover insurance charges also. If any point is raised as regards insurance charges after orders are issued, the same will not be entertained and the firms thereafter should effect the supply at their own cost.
4. Rates should be quoted F.O.R. Pondicherry, Extra packing; forwarding charges etc. should not be quoted.
5. Delivery is required urgently. Tenders should please state the guarantee delivery period they can offer. As delivery date is essential for the contract, this should be strictly adhered to by the successful tenders.
6. No supply which is not according to the specifications and not meeting our requirement will be accepted.
7. The Director shall have the right of rejecting the Limited Tender in whole or part without assigning any reason therefore.
8. In case of high precision instruments the firms should give a guarantee certificate for their satisfactory performance.
9. Limited Tenders should be sent only by Registered Post. Hand Quotations are not accepted. Please furnish your Limited Tender in a sealed cover super scribing as:

Limited Tender for EQUIPMENT USB Video Dermatoscope with Polarised Light
Enquiry No. : No.JIP/PLS/ERRC/2014/10 Dated 29.12.2014

Due Date: 13.01.2015.

Limited Tender should reach this office on or before 13.01.2015.

Limited Tender received after the due date will be summarily
rejected.

- NOTE:**
1. Quotation may please be address to:
The Director, JIPMER, Puducherry-6.
 2. The Sealed Cover (quotation) may please be sent directly to:
The Head, Dept. of Plastic Surgery, S.S. Block,
IV Floor, JIPMER, Puducherry-6.

Yours faithfully,

For DIRECTOR

**DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY-605 006.**

No.JIP/PLS/ERRC/2014/10 DUE DATE: 13.01.2015 Dated:29.12.2014

NAME OF THE EQUIPMENT: USB VIDEO DERMATOSCOPE WITH POLARIZED LIGHT

SPECIFICATION:

- Image Sensor should be ¼" Sony Super HAD CCD with a main camera
- Sync System should have 2:1 Internal
- Resolution must be 480 TV Lines (Horizontal)
- Effective Pixels should be 410,000
- Video Output Signal should be 1.0V p-p Composite 75 ohm/USB2.0
- Camera Function should be Live, Capturer, TPL mode, Normal Mode
- Lens Spec should be X1 Lens, X60 triple lens, X10 Lens
- Disinfection must be UV Disinfection Lamp
- White Balance must be Automatic White Balance
- Should have dedicated software
- Should be supplied with Laptop with latest specifications and color printer
- Min Illumination should be 0.1 Lux at F2.0
- Operative Temperature should be 14F- 122F (-10C – 50 C)
- Humidity should be within 90% RH
- Power Source must be DC 5V +/- 10%, 1.5A max
- Dimensions should be 86(W) X 99.5(H) X 245(L)mm
- With option for hair kit with 200X lens and 1000X Lens
- Should be CE certified
- Accessories should include:
 - Oil stick and oil tape
 - Moisture checker
 - DC 5V adaptor
 - USB 2.0 cable
 - RCA Cable
 - Holder

Note:

*** Demonstration of the equipment is essential**

*** Warrantee Three Years**

*** AMC for next five years**

***Please quote preferably in Indian Rupees.**

(DR.RAVIKUMAR CHITTORIA)