

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No.JIP/BURNS/ERRC/2015/9

Dated:11.02.2015

To

Sir,

**Sub: Supply of Vaccum Based Patient Positioning System - Limited tender- Invited –
Regarding**

Please quote your lowest rates for supply of the Vaccum Based Patient Positioning System as per the list attached in Annexure-1, subject to the following terms and conditions:

1. Rates should be quoted only for the items which are available in stock and can be supplied immediately on receipt of order.
2. Only 5% CST is applicable
3. No insurance charges are payable as per the rules in the Government. As such, the firms before quoting should take into consideration all the risks in the transit and then furnish Limited Tender which should cover insurance charges also. If any point is raised as regards insurance charges after orders are issued, the same will not be entertained and the firms thereafter should effect the supply at their own cost.
4. Rates should be quoted F.O.R. Pondicherry, Extra packing; forwarding charges etc. should not be quoted.
5. Delivery is required urgently. Tenders should please state the guarantee delivery period they can offer. As delivery date is essential for the contract, this should be strictly adhered to by the successful tenders.
6. No supply which is not according to the specifications and not meeting our requirement will be accepted.
7. The Director shall have the right of rejecting the Limited Tender in whole or part without assigning any reason therefore.
8. In case of high precision instruments the firms should give a guarantee certificate for their satisfactory performance.
9. Limited Tenders should be sent only by Registered Post. Hand Quotations are not accepted.

Please furnish your Limited.Tender in a sealed cover super scribing as:

Limited Tender for EQUIPMENT Vaccum Based Patient Positioning System

Enquiry No. : No.JIP/BURNS/ERRC/2015/9 Dated 11.02.2015

Due Date: 25.02.2015

Limited Tender should reach this office on or before 25.02.2015.

Limited Tender received after the due date will be summarily rejected.

- NOTE:**
1. Quotation may please be address to:
The Director, JIPMER, Puducherry-6.
 2. The Sealed Cover (quotation) may please be sent directly to:
The Head, Dept. of Plastic Surgery, S.S. Block,
IV Floor, JIPMER, Puducherry-6.

Yours faithfully,

For DIRECTOR

**JIPMER TERTIARY BURN CARE CENTRE (JTBC)
DEPARTMENT OF PLASTIC SURGERY
JIPMER, PUDUCHERRY**

No.JIP/BURNS/ERRC/2015/9

DUE DATE: 25.02.2015

Dated:11.02.2015

NAME OF THE EQUIPMENT: VACCUM BASED PATIENT POSITIONING SYSTEM

SPECIFICATION:

- Should consist of vacuum pad filled with plastic beads capable of taking any shape as per positioning needs
- Should be easy to use using vacuum line / vacuum pump.
- Should be supplied with manual Vacuum Pump
- Should provide quick stable and comfortable patient positioning during all kind of surgeries without forcing the patient into a strained position.
- Should mold firmly to shape of the patient body providing uniform support to all body parts and preventing pressure injuries.
- Should facilitate maneuvering patients arm and legs without needing to change position of the system.
- Positioning pad should be made up of soft latex free PVC fabric
- Should be radiolucent and easy to disinfect using common liquid disinfectants.
- Should be supplied complete with Pads in 7 different shape and sizes
- Should be supplied with 25 Nos of disposable covers for each size.
- For Head immobilization and Neck Support the pad should be **U Shape** (40.6 X 30.5 cm)
- For Back Support the pad should be **T-Shape with flaps** (45.7 X 50.8 cm)
- For Thyroid Position the pad should be **Small, with U-shape cutout** (72.9 x 90.2 cm) -
- For Kidney Position the pad should be **Medium, with U-shape cutout** (98.3 X 90.2 cm)
- For Frog Leg Position the pad should be **Large, with U-shape cutout** – (118.6 X 90.2 cm)
- For Lateral Position and Shoulder Access the pad should be **Extra-Large, with U-shape cutout** – (87.6 X 71.1 cm)
- For Full body support the pad should be **Large rectangle** – (76.2 X 193.1 cm)

Note:

*** Demonstration of the equipment is essential**

*** Warrantee Three Years**

*** AMC for next five years**

***Please quote preferably in Indian Rupees.**

(DR.RAVIKUMAR CHITTORIA)