



Phone: 2272380 to 2272389
Fax : 0413-2272067/66
GSTIN 34AAAJJ0846M1Z4

E.mail: purchase@jipmer.edu.in
website: www.jipmer.edu.in

No. PUR.2(17)/ENDO-OPEX/2023-24/02

Dated: **22 JUN 2023**


CORRIGENDUM

“Chemiluminescence Immunoassay system on reagent rental basis” (OPEX)
e-Tender No. PUR.2(17)/ENDO-OPEX/2023-24/02, Tender ID No. 2023_JIPMP_757063_1, dt.09.06.2023 for the Department of Endocrinology, JIPMER, Puducherry - 06.

It is hereby informed that the few technical Specifications for the above item have been changed and amended technical Specification is as follows:-

Tender SL No	Existing technical specification	Amended technical specification
Specification 8, Point 5	System should have the onboard cooling facility (4 ^o -10 °C) to maintain the reagents	System should have the onboard cooling facility (5 ^o -20°C) to maintain the reagents
Specification 15, Point 3	Should have the facility for continuous random access, including loading and unloading of reagents, other consumables and samples without stopping the analyser	System should have the facility for continuous loading and unloading of samples with random access, without stopping the analyser. In addition, reagents and other consumables can be loaded during the run or should have the facility to load backup reagents packs and consumables
Specification 16, Point 3	Provision of inbuilt QC monitoring system by LJ plots and Westgard and Configurable QC based rules should be available	Provision of inbuilt QC monitoring system by LJ plots and Westgard rules should be available
Specification 17, Point 5	Real time monitoring of QC violations, and turn-around time for samples should be available	Deleted

** All other tender details remain unchanged.


OFFICER IN-CHARGE OF PURCHASE
जिपमेर / JIPMER
पुदुच्चेरी / Puducherry - 605 006