



**JAWAHARLAL INSTITUTE OF POST GRADUATE MEDICAL  
EDUCATION & RESEARCH (JIPMER)**



**GOVERNMENT OF INDIA**

(An Institution of National Importance under Ministry of Health & Family Welfare) Dhanvantri  
Nagar, Puducherry-605006

**OPEN TENDER  
ENQUIRY DOCUMENT  
FOR  
RATE CONTRACT  
FOR  
SUPPLY OF CONSUMABLES/NON-CONSUMABLES  
FOR  
DEPARTMENT OF  
RADIO DIAGNOSIS**

(This document consisting of **42** pages)

**JIPMER,  
DHANVANTARI NAGAR,  
PUDUCHERRY 605006**

**Phone: - 0413-2296016/15/47 ; Fax:- 0413-2272067/66**

**Email: [radiodiagnosisstores@gmail.com](mailto:radiodiagnosisstores@gmail.com) URL: - [www.jipmer.edu.in](http://www.jipmer.edu.in)**

## INDEX

<b>Section</b>	<b>Topic</b>	<b>Page No.</b>
<b>Section-I</b>	<b>Notice inviting Tender (NIT)</b>	<b>03</b>
<b>Section-II</b>	<b>Schedule of Requirements and Specifications</b>	<b>05</b>
<b>Section-III</b>	<b>General Instructions to Tenderers (GIT)</b>	<b>24</b>
<b>Section-IV</b>	<b>General Conditions of Contract (GCC)</b>	<b>33</b>
<b>Section-V</b>	<b>Bid Application Form</b>	<b>35</b>
<b>Section-VI</b>	<b>Manufacturer's Authorization Form</b>	<b>36</b>
<b>Section-VII</b>	<b>Tender Form</b>	<b>37</b>
<b>Section-VIII</b>	<b>Check list for Submission of Tender</b>	<b>38</b>
<b>Section-IX</b>	<b>Bid Security Declaration</b>	<b>39</b>
<b>Appendix-A</b>	<b>Self-certification format for claiming purchase preference under the "Public Procurement preference to Make in India" order</b>	<b>40</b>
<b>Section-X</b>	<b>GFR-144 (xi) compliance certificate</b>	<b>42</b>

## SECTION-I

### JAWAHARLAL INSTITUTE OF POSTGRADUATE MEDICAL EDUCATION AND RESEARCH PUDUCHERRY – 6

(Institution of National Importance under the Ministry of Health & Family Welfare,  
Govt. of India)

#### Purchase Section

No. JIP/PUR6(1)/OeT/RD/2021-22

Dated: 07.08.2021

#### NOTICE INVITING e-TENDER (e-NIT)

Sub: Supply of Consumables / Non-Consumables  
-----

E- tenders are invited from eligible and qualified domestic manufacturer's or their distributors for the RATE CONTRACT for supply of CONSUMABLES/NON-CONSUMABLES for the Department of **RADIODIAGNOSIS** for 2021– 22.

1. Scope of work: Supply of consumables/non-consumables for the Department of RADIODIAGNOSIS
2. Value of tender: Approximate Cost **Rs. 1,00,00,000**
3. Tender timelines:
  - i. Opening date & time for download of Tender document: **09.00 AM on 07.08.2021**
  - ii. Last date for receipt of pre-bid queries: **06.00 PM on 11.08.2021**
  - iii. Pre-bid queries can be made through e-mail to: **[radiodiagnosisstores@gmail.com](mailto:radiodiagnosisstores@gmail.com)**
  - iv. Opening date & time for submission of online bids: **09.00 AM on 14.08.2021**
  - v. Closing date & time for submission of online bids **06.00 PM on 04.09.2021**
  - vi. Date and Time of opening of online bids for Technical evaluation: **11.00 AM on 06.09.2021**
  - vii. Date & time of opening of Price Bid: To be intimated later.
4. Earnest Money Deposit/Bid Security: Bid security declaration must be submitted by all bidders.
5. Tender Processing Fee (Non-refundable): **Rs. 590 (including 18% GST) shall be paid through SBI collect only.**  
**<https://www.onlinesbi.com/sbicollect/icollecthome.htm>**
6. Interested bidders are advised to download the complete Tender Enquiry document from the websites **[www.jipmer.edu.in](http://www.jipmer.edu.in)** or **<https://eprocure.gov.in/eprocure/app>** for complete details.
7. The prospective bidders must register with the E-procurement system of **<https://eprocure.gov.in/eprocure/app>**. Special Instructions to the bidders for the e-submission of the bids online through this eProcurement Portal on completion of the registration process is given in **<https://eprocure.gov.in/eprocure/app>** , the bidders will

be provided user ID and password upon enrollment. In order to submit the bids electronically, bidders are required to have a valid Class 3 Digital Signature Certificate (signing and encryption/ decryption certificates).

8. Bidders are requested to read the bidders help document on e-tender web site link before proceeding for bidding.
9. Post receipt of User ID & Password, Bidders can log on for downloading & uploading tender document.
10. The bidders shall submit the required bid security declaration (as per G.I.T clause 2) before the due date and time mentioned above.
11. The online submission of tender(s) can only be done through **<https://eprocure.gov.in/eprocure/app>**
12. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **<https://eprocure.gov.in/eprocure/app>** e-portal (as described above) only.
13. Prospective bidders are advised to browse the above websites regularly before submission of their bids as any further amendments will be published in these websites only.

Asst. Officer in Charge  
Purchase Section,  
For  
Director,  
JIPMER, Puducherry

## Section-II

### SCHEDULE OF REQUIREMENTS AND SPECIFICATIONS

#### LIST OF ITEMS REQUIRED

1. Scope of work : Supply of Consumables/ Non-Consumables for the department of **RADIODIAGNOSIS** as detailed below

Sl. No.	Item Description	Quantity	Units
1.000	<b>PART A - GENERAL CONSUMABLES</b>		
1.001	Ultrasound gel 250 ml	<b>10000</b>	Pieces
1.002	Sterile Ultrasound probe covers with securing bands (150 cm length)	<b>400</b>	Pieces
1.003	Sterile Ultrasound probe covers with securing bands (50 cm length)	<b>300</b>	Pieces
1.004	Lead numbers 0 to 9 (Thick)	<b>300</b>	Pieces
1.005	Lead side mark R and L	<b>100</b>	Pieces
1.006	MRI-compatible ECG Electrodes for use in MRI systems upto 1.5 Tesla	<b>100</b>	Pieces

1.007	MRI-compatible ECG Electrodes for use in MRI systems upto 3 Tesla	<b>100</b>	Pieces
1.008	High Pressure contrast injecting tube - 150 cm	<b>6000</b>	Pieces
1.009	Contrast media for MRI (Macro cyclic ionic, cyclic non-ionic or linear ionic contrast agents) 10 ml	<b>4000</b>	Pieces
1.010	Contrast media for MRI (Macrocyclic ionic agent) 10 ml	<b>500</b>	Pieces
1.011	Intravascular radiological iodinated contrast media (non-ionic Low osmolar of range 500-650 mOsm/kg & 300 mg-Iodine/ml) 100 ml	<b>14000</b>	Pieces
1.012	Intravascular radiological iodinated contrast media (non-ionic Low osmolar of range 500-650 mOsm/kg & 300 mg-Iodine/ml) 50 ml	<b>4500</b>	Pieces
1.013	Intravascular radiological iodinated contrast media (non-ionic iso- osmolar of range 270-320 mOsm/kg & 320 mg-Iodine/ml) 100 ml	<b>500</b>	Pieces
1.014	Intravascular radiological iodinated contrast media (non-ionic , low osmolar,400 mg-Iodine/ml) 100 ml for angiographic applications	<b>2000</b>	Pieces

1.015	Intravascular radiological iodinated contrast media (non-ionic iso- osmolar of range 270-320 mOsm/kg & 270mg-Iodine/ml) 100 ml	<b>500</b>	Pieces
1.016	Intravascular (microbubble-based) ultrasound contrast agent for Liver, renal, thyroid and breast applications.	<b>200</b>	Pieces
1.017	Water soluble Iodinated oral contrast for CT 30 ml vial	<b>500</b>	Pieces
1.018	Intravascular radiological iodinated contrast media - Sodium diatrizoate Meglumine diatrizoate 60% - 20 ml	<b>1000</b>	Pieces
1.019	Micronized Barium sulphate 95% IL sus).	<b>200</b>	Pieces
1.020	Dental X ray films 3" X 4" (150 Pcs/Pk)	<b>30</b>	Pieces
1.021	Co Axial Automatic Lung Biopsy gun with needles of introducer length 15cm /gauge 18 throw length 2 to 2.5 cm with introducer.	<b>50</b>	Pieces
1.022	Co Axial Automatic Lung Biopsy gun with needles of introducer length 15cm /gauge 20 throw length 2 to 2.5 cm with introducer.	<b>100</b>	Pieces

1.023	Co-Axial Automatic Biopsy gun with Needle of throw length 10 mm 20 Guage 15 cms Length with introducer.	<b>100</b>	Pieces
1.024	Semi automatic Co-Axial Biopsy gun with Needle of length 15 cm throw length 1 cm, 20 G with introducer.	<b>100</b>	Pieces
1.025	Co-Axial Automatic Biopsy gun with Needle of throw length 10 mm 18 Guage 15 cms Length with introducer.	<b>100</b>	Pieces
1.026	Semi automatic Co-Axial Biopsy gun with Needle of length 15 cm with adjustable throw lengths of both 1 cm & 2 cm, 20 G with introducer .	<b>50</b>	Pieces
1.027	Semi automatic Co-Axial Biopsy gun with Needle of length 20 cm with adjustable throw lengths of both 1 cm & 2 cm, 20 G with introducer.	<b>50</b>	Pieces
1.028	Semi automatic Co-Axial Biopsy gun with Needle of length 15 cm with adjustable throw lengths of both 1 cm & 2 cm, 18 G with introducer.	<b>50</b>	Pieces
1.029	Automatic biopsy gun length 15 cm, 18G with adjustable throw lengths of 1cm and 2 cm.	<b>50</b>	Pieces



1.030	Automatic biopsy gun length 15 cm, 20G with adjustable throw lengths of 1cm and 2 cm.	<b>50</b>	Pieces
1.031	Vertebral bone Biopsy needle introducer length 15 cm /gauge 13 (J Needle)	<b>10</b>	Pieces
1.032	Vertebral bone Biopsy needles introducer length 15cm /gauge 13 ( Murphy's Needle)	<b>10</b>	Pieces
1.033	Ackerman's bone biopsy needle set 12 G needle & 16.8 cm length Should be FDA/CE/DGCI approved	<b>5</b>	Pieces
1.034	Cope mandril 0.018 inch diameter kink resistant nitinol guide wire	<b>30</b>	Pieces
1.035	Ultrathane/Polyurethane Percutaneous Malecot type drainage catheter 6F,8F,10F, 12F with plastic stylet ,length 40 cm or above accepts guide wire upto 0.038 diameter with urobag connector	<b>100</b>	Pieces
1.036	Ultrathane/Polyurethane multipurpose Percutaneous Pigtail type drainage catheter 6F,8F,10F, 12F,14F,16F with plastic stylet, with 5 or more side holes length 40 cm or above accepts guide wire upto 0.038 diameter with urobag connector	<b>100</b>	Pieces

1.037	25 cm length ultrathane / polyurethane multipurpose pigtail drainage catheter needle set with hydrophilic coating, minimum 4 side holes with stainless steel trocar stylet and stiffening canula. Size 8F, 10F, 12F, 14F,16F. Should be FDA/CE/DGCI approved	50	Pieces
1.038	50 cm length ultrathane/polyurethane multipurpose pigtail drainage catheter needle set with hydrophilic coating, with minimum 4 side holes with stainless steel trocar stylet and stiffening canula. Size 8F, 10F, 12F, 14F. Should be FDA/CE/DGCI approved.	25	Pieces
1.039	Micro puncture access set with 21 G needle 4 cm Length 0.018 40 Cm angled tip Guidewire 5 Fr 10 cm sheath	10	Pieces
1.040	HSG Catheter with size of 5 Fr and 7 fr with balloon.	10	Pieces
1.041	<b>Neff percutaneous access set</b> designed for the introduction of 038 guide wire in biliary interventional procedures comprising of 21/22 G,15 cm length Chiba needle, 0.018 inch diameter kink resistant nitinol guide wire, coaxial introducer sheath 4 ID *6 OD. Should be FDA/CE/DCGI approved.	50	Pieces
1.042	<b>PTBD set</b> with 18G chiba needle KMP catheter 5Fr 40 cm length accepting guide wire upto 0.038 inch; 035*150 cm hydrophilic coated angled tip guide wire with torque device; 035*150 cm curved 3 mm J tip amplatz extra stiff guide wire with gradual transition to flexible distal tip: 6,8,10,12 Fr dilators. Should be FDA/CE/DCGI approved.	50	Pieces
1.043	<b>Chiba biopsy needles</b> 20 cm length 18G & 20G. Should be FDA/CE/DCGI approved.	40	Pieces
1.044	50 cm length Ultrathane Ring biliary drainage catheter having 32 side holes. Size 8F, 10F, 12 F and it should accept guide wire diameter 0.038 inch. Should be FDA/CE/DCGI approved	50	Pieces
1.045	Amplatz Extra-Stiff PTFE-Coated Stainless Steel support wire guide - J Tip/curved tips Size 145 cm. Should be FDA/CE/DCGI approved	50	Pieces

1.046	<b>KMP catheter</b> 5 Fr,40cm length with curved tip accepts guide wire of diameter 0.035 inch	<b>25</b>	Pieces
1.047	<b>Dilators</b> of various sizes 8, 10, 12 & 14 Fr accepting guidewire of diameter 0.038 inch	<b>50</b>	Pieces
2.000	<b>PART B - VASCULAR INTERVENTIONAL CONSUMABLES</b>		
2.048	<b>Femoral arterial introducer set</b> with sheath, dilator with radiopaque tip, short stiff/teflon guidewire and wire introducer. Tapering design at sheath-dilator tip for smooth penetration. Cross-cut hemostasis valve to prevent reflux and air aspiration. Anti-kinking radiopaque sleeve. Snap-on/click-off dilator lock to avoid backout and one-hand handling. Size 4 Fr to 9 Fr. Should be FDA/CE/DCGI approved	<b>50</b>	Pieces
2.049	30 cm long kink resistance coil reinforced guiding sheath with <b>Tuohy-borst</b> valve without any metallic component. Size 6Fr, 7 Fr , 8 fr. Should be FDA/CE/DCGI approved	<b>2</b>	Pieces
2.050	55 cm long kink resistance coil reinforced guiding sheath with <b>Tuohy-borst</b> valve. Size 5 Fr ,6Fr, 7 Fr , 8 Fr Should be FDA/CE/DCGI approved	<b>2</b>	Pieces
2.051	70 cm long kink resistance coil reinforced guiding sheath with <b>Tuohy-borst</b> valve. Size 5 Fr, 6Fr, 7 Fr, 8 Fr Should be FDA/CE/DCGI approved	<b>2</b>	Pieces
2.052	90 cm long kink resistance coil reinforced, hydrophilic coated guiding sheath with <b>Tuohy-Borst</b> valve & have 018 & 038 dilators. Size 5 Fr, 6 Fr, 7 Fr, 8 Fr Should be FDA/CE/DCGI approved	<b>2</b>	Pieces
2.053	6 French kink resistance, coil reinforced, low- friction PTFE coated lumen <b>Balkin sheath</b> with 180 degree curve for contralateral approach. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces

2.054	Femoral arterial sheath <b>Ansel design</b> . 6 Fr Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.055	Femoral arterial <b>Puncture needle</b> 16G, 18G. Should be FDA/CE/DCGI approved	<b>50</b>	Pieces
2.056	<b>Micro-puncture introducer set</b> with Silhouette transitionless tip. 4Fr, 5 Fr Should be FDA/CE/DCGI approved.	<b>10</b>	Pieces
2.057	<b>Pedal access introducer set</b> with 21 G -ecotip needle for ultrasound visualisation. 4 Fr, 5Fr Should be FDA/CE/DCGI approved.	<b>10</b>	Pieces
2.058	100 cm long high-flow high-torque <b>Vertebral</b> catheter with radiopaque tip having sixteen stainless steel wire braid. 4 Fr, 5 Fr. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.059	100 cm long high-flow high-torque <b>Berenstein</b> catheter with radiopaque tip having sixteen stainless steel wire braid. 4 Fr, 5 Fr. Should be FDA/CE/DCGI approved	<b>5</b>	Pieces
2.060	100 cm long high-flow high-torque <b>Judkin's Right</b> coronary (JR) catheter braided with radiopaque tip . 4 Fr, 5 Fr. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.061	100 cm long high-flow high-torque <b>Headhunter</b> braided catheter with radiopaque tip . 5 Fr. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.062	100 cm long high-flow high-torque <b>SIM 1</b> tip configuration braided catheter with radio-opaque tip. 4 Fr, 5 Fr. Should be FDA/CE/DCGI approved	<b>20</b>	Pieces
2.063	100 cm long high-flow high-torque <b>SIM 2</b> tip configuration braided catheter with radio-opaque tip. 4Fr, 5 Fr. Should be FDA/CE/DCGI approved	<b>20</b>	Pieces

2.064	65 cm long high-flow high-torque <b>Cobra 1</b> tip configuration braided catheter with radiopaque tip. 4 Fr, 5 Fr. Should be FDA/CE/DCGI approved	5	Pieces
2.065	65 cm long high-flow high-torque <b>Cobra 2</b> tip configuration braided catheter with radiopaque tip. 4 Fr, 5 Fr. Should be FDA/CE/DCGI approved	20	Pieces
2.066	90 cm long high-flow high-torque <b>RUC</b> tip configuration braided catheter with radiopaque tip. 5 Fr. Should be FDA/CE/DCGI approved	25	Pieces
2.067	100 cm long high-flow high-torque <b>multipurpose</b> braided catheter with radiopaque tip. 4 Fr, 5 Fr. Should be FDA/CE/DCGI approved	5	Pieces
2.068	150 cm long <b>support</b> catheter with braided construction with hydrophilic coating, low profile from tip to hub and contains radiopaque bands at the tip and at 5, 10 & 15 cm. 014 & 035 Compatible. Should be FDA/CE/DCGI approved	10	Pieces
2.069	100 cm long high flow <b>IMA (internal mammary)</b> catheter with radiopaque tip 4Fr, 5Fr. Should be FDA/CE & DCGI approved.	10	Pieces
2.070	100 cm long high-flow high-torque <b>MPA</b> catheter with radiopaque tip 4Fr, 5Fr. Should be FDA/CE & DCGI approved.	10	Pieces
2.071	65 cm <b>Pigtail</b> catheter 035 compatible with 5 sideports. 5Fr Should be FDA/CE/DCGI approved	10	Pieces
2.072	110 cm long high flow <b>pigtail</b> catheter with 12 side holes having sixteen stainless steel wire braid. 5 Fr, 6Fr. Should be FDA/CE/DCGI approved	10	Pieces
2.073	035 × 150 cm hydrophilic coated <b>angled</b> tip wire guide with torque device. Should be FDA/CE/DGCI approved	100	Pieces

2.074	035 × 150 cm hydrophilic coated <b>Straight</b> tip wire guide with torque device. Should be FDA/CE/DGCI approved	25	Pieces
2.075	035 × 260 cm hydrophilic coated stiff shaft <b>straight</b> tip wire guide with torque device. Should be FDA/CE/DGCI approved	10	Pieces
2.076	035 × 260 cm hydrophilic coated <b>angled</b> tip wire guide with torque device. Should be FDA/CE/DGCI approved	25	Pieces
2.077	035 × 180 cm curved , 3 mm <b>J-tip</b> amplatz Extra stiff wire guide with gradual transition to flexible distal tip. Should be FDA/CE/DGCI approved	25	Pieces
2.078	035 × 150 cm curved , 3 mm <b>J-tip</b> amplatz Extra stiff wire guide with gradual transition to flexible distal tip. Should be FDA/CE/DGCI approved	25	Pieces
2.079	035 × 260 cm curved , 3 mm <b>J-tip</b> amplatz Extra stiff wire guide with gradual transition to flexible distal tip. Should be FDA/CE/DGCI approved	10	Pieces
2.080	Microguidewire <b>014</b> indicated for use in distal catheterization, selective, and superselective embolization in small, distal, angular, and sensitive peripheral and neurological vasculatur. Nitinol core with shape memory, flexibility, and kink resistance for swift device exchange. Should be FDA/CE/DCGI approved.	10	Pieces
2.081	Ergonomic design Inflation device set with Latch mechanism on threaded plunger, luminescent dial displays pressure in both atm and psi. Maximum pressure of 40 atm & Barrel volume of 20 ml. Should be FDA/CE/DCGI approved	1	Pieces
2.082	<b>Occlusion Balloons</b> for temporary occlusion. Mention inflated balloon size and catheter/sheath compatibility. Should be FDA or CE & DCGI approved	10	Pieces

2.083	<b>Microballoon</b> for occlusion to use in forward embolization like TACE. Should be FDA or CE & DCGI approved	<b>10</b>	Pieces
2.084	Peripheral Microcatheter with wire 150 cm nylon, hydrophilic coated, superselective microcatheter with full length stainless steel braid & has uniform diameter from proximal to distal tip. The 15 cm distal tip to be hand shaped without heat. <b>Specify Lipidol/DSMO/Alcohol compatibility</b> that can also accommodate embolic particles. <b>Should be quoted with appropriate compatible stylet/guidewire as a single package.</b> Should be FDA/CE/DCGI approved	<b>50</b>	Pieces
2.085	<b>Carotid &amp; Renal PTA Balloon 0.018</b> Compatible, 6F Compatible, Length of Balloon 20 to 80 mm, Hydrophilic coated, Rapid Exchange system. Should be FDA or CE & DCGI approved.	<b>25</b>	Pieces
2.086	<b>Renal (balloon mounted) stent 0.014 compatible</b> , with high radial strength and rapid exchange delivery system Available diameters 4 mm to 7 mm. Mention diameter range & length available. Should be FDA or CE & DCGI approved	<b>10</b>	Pieces
2.087	<b>Peripheral PTA balloon (035') 0.035</b> wire compatible, smooth transition to the wire, flexible distal end of the catheter permit optimal tracking through tortuous anatomy, mention the sizes available for all lesion type. Mention diameter range & length available. Should be FDA or CE & DCGI approved	<b>10</b>	Pieces
2.088	Peripheral PTA balloon (014') Large internal lumen design facilitates reduced deflation time, hydrophilic balloon coating, mention the sizes available for all lesion type. Mention diameter range & length available Tapered balloon if available should be quoted separately. Should be FDA or CE & DCGI approved.	<b>10</b>	Pieces
2.089	Drug Coated PTA balloon Catheter design Over the Wire (OTW), Mention balloon material, coating & eluting formulation. Mention usable shaft length & diameter. Introducer sheath compatibility 5F/6F Guidewire 0.035'. Mention Nominal pressure & burst pressure. Mention diameter range & length available. Should be FDA or CE & DCGI approved.	<b>5</b>	Pieces
2.090	<b>Peripheral stent (balloon mounted stent) 0.035</b> wire compatible, should have delivery system & compatible wire. Mention composition & strut design, should have minimal shortening. Mention diameter range & length available. Should be FDA or CE & DCGI approved	<b>10</b>	Pieces

2.091	<b>Peripheral Self Expanding stent 0.035</b> wire compatible, Mention Stent Material. Should have expansion force that resists compression while providing wall apposition Secure deployment System secures the stent to essentially eliminate premature deployment or jumping. Mention diameter range & length available. Should be FDA or CE & DCGI approved.	<b>10</b>	Pieces
2.092	<b>Stent graft (covered Stent)</b> for carotid and peripheral vessels. Should be FDA or CE & DCGI approved.	<b>10</b>	Pieces
2.093	<b>Abdominal aortic aneurysm graft set</b> along with introducer sheath, stiff wires (2 no.s), balloon and marker pigtail. Should be FDA or CE & DCGI approved	<b>5</b>	Pieces
2.094	<b>Thoracic aortic aneurysm graft set</b> along with introducer sheath, stiff wires (1 no.), balloon and marker pigtail. Should be FDA or CE & DCGI approved	<b>5</b>	Pieces
2.095	0.035 wire compatible Self expanding metallic stents 10 mm diameter and length 60 mm,80 mm,100 mm ,120 mm <b>for biliary intervention</b> . Mention Stent Material.Should have expansion force that resists compression while providing wall apposition Secure deployment System secures the stent to essentially eliminate premature deployment or jumping. Mention diameter range & length available. Should be FDA or CE & DCGI approved.	<b>10</b>	Pieces
2.096	<b>Inferior Vena cava Filter set</b> should contain both Jugular and Femoral delivery options with a 7Fr- 65 cm length introducer sheath with multipurpose dilator includes radiopaque sizing bands and flushing sideports. Filter length 48mm Should be FDA/CE/DCGI approved	<b>2</b>	Pieces
2.097	<b>IVC Filter retrieval set</b> that should comprises of long snare and coaxial 12Fr introducer sheath without haemostatic valve Should be FDA/CE/DCGI approved	<b>2</b>	Pieces
2.098	<b>Particulate embolic agents Polyvinyl alcohol (PVA)</b> <b>Particles</b> sizes of various ranges - 50-150 $\mu$ , 150-300 $\mu$ , 250-355 $\mu$ , 300-500 $\mu$ , 500-710 $\mu$ , 710-1000 $\mu$ supplied sterile in stopper – sealed bottles. <b>If any size range is not available, it should be mentioned explicitly.</b> Should be FDA/CE/DCGI approved	<b>100</b>	Pieces
2.099	<b>Gelfoam sponge</b> for intravascular embolization (mention size availale). Should be FDA/CE/DCGI approved	<b>25</b>	Pieces



2.100	<b>Gelfoam sheets</b> for intravascular embolization. Should be FDA or CE & DCGI approved	<b>25</b>	Pieces
2.101	<b>Embospheres</b> of various sizes 100-300, 300 - 500, 500 - 700 microns in pre-loaded syringes 1 or 2 ml. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.102	Super Absorbant Polymer Microspheres for Drug (Doxorubicin) Adsorption for TACE. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.103	Nester fibered pushable coils <b>35'</b> different sizes. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.104	Nester fibered pushable coils <b>18'</b> different sizes. Should be FDA/CE/DCGI approved	<b>50</b>	Pieces
2.105	<b>035</b> Platinum Fibered Super soft <b>Detachable</b> coils of various size. Should be FDA/CE/DCGI approved	<b>10</b>	Pieces
2.106	<b>018</b> Platinum Fibered Super soft <b>Detachable</b> coils of various size. Should be FDA/CE/DCGI approved	<b>25</b>	Pieces
2.107	Neuro Guiding catheters 0.070 inch inner diameter, braided guiding catheter for Carotid access, 5F inner catheter compatible, various shapes. Should be FDA or CE & DCGI approved.	<b>10</b>	Pieces
2.108	<b>Distal access guiding catheter</b> 0.071 inch and above inner diameter, straight & multi-purpose tip, end-to- end braided with flexible distal shaft with hydrophilic coating. Should be FDA/CE /DCGI approved	<b>5</b>	Pieces
2.109	<b>Intermediate or Distal Delivery catheter</b> for intracranial access/support (ID 0.05 to 0.06 inch) for AVM/ aneurysm embolization. Should be FDA or CE & DCGI approved	<b>10</b>	Pieces

2.110	Straight tip/Preshaped 45* & 90* endhole single- lumen Microcatheter for super selective catheterization with 0.017 ID 2.1/1.7F OD Nitinol braided 150 cm usable length dual radiopaque markers permitting 0.014 guide wire & DMSO Compatible. Should be FDA or CE & DCGI approved	25	Pieces
2.111	Neuro microcatheter flow-directed Super-selective high flow directed braided micro catheter OD 2.7/1.3F usable length 165 cm with Hydrophilic coating for AVM treatment. (Specify if the catheter is braided/non-braided and the specific material of the catheter hydrophilic coating used). Should be FDA or CE & DCGI approved	10	Pieces
2.112	Super-selective high flow directed braided micro catheter superior navigation durability fully reinforced detachable tip OD 1.9/0.63 F usable length 165 cm with Hydrophilic coating for AVM treatment (Specify if the catheter is braided/non-braided and the specific material of the catheter hydrophilic coating used). Should be FDA or CE & DCGI approved	25	Pieces
2.113	Neuro microguidewire Soft and extra support neuro-guidewire, nitinol- stainless steel hybrid wire, distal tip core <b>0.012</b> inch OD, distal 1.5 cm shapeable, re-shapeable 'N' times, proximal docking system. Should be FDA or CE & DCGI approved	25	Pieces
2.114	High torque, stainless steel with a radiopaque distal platinum coil tip 0.008, 200 cm hydrophilic guide wire with torque device for use with infusion catheters. Should be FDA or CE & DCGI approved	10	Pieces
2.115	Liquid embolic material Ethylene vinyl alcohol copolymer (EVAP)-based liquid embolic agent with dimethyl sulfoxide (DMSO) & tantalum powder. Mention whether it is premixed, need for shaker for mixing or pre-loaded in syringe. [If it needs shaker it should be provided for every case without extra cost]. Should be FDA or CE & DCGI approved	5	Pieces
2.116	n-butyl cyanacrylate (NBCA) glue 0.5 ml vials for direct intravascular use. Should be FDA or CE & DCGI approved	50	Pieces
2.117	<b>Lipiodol</b> 10 ml Ampoule	50	Pieces

3.000	Carotid Stent with distal embolic protection device (3.118 & 3.119) should be quoted together and price comparison will be taken on total cost of both		
3.118	<b>Carotid Endovascular self-expanding stents</b> (straight & tapered) with delivery system for use in carotid & vertebral arteries (extracranial use) & dural sinuses (rapid exchange self-expandable nitinol stent) with MR compatibility distal and proximal radiopaque markers. Mention clearly diameter/length range available, guiding catheter (7F or 8F) compatibility, crossing profile and 014 wire compatibility. Should be FDA or CE & DCGI approved.	10	Pieces
3.119	<b>Distal embolic protection device</b> for carotid stenting - filter type, affixed to guidewire, 014 or 0.018 guidewire compatible fast exchange system, specify deployment & retrieval catheter size. Mention crossing profile of deployment & retrieval catheters. Specify the diameters available. Should be FDA or CE & DCGI approved.	10	Pieces
3.120	Intracranial balloon - Double & Co-axial lumen, 014 compatible, contrast dependant inflation & deflation, guidewire independent, compliant, with distal tip marker. Mention sizes available. [If super/extra-compliant double lumen balloon available, to be quoted separately]. Also mention availability of proximal marker for coiling. Should be able to use for coiling as a microcatheter. Mention whether contrast tip needs to be sealed before usage. Should be FDA or CE & DCGI approved.	5	Pieces
3.121	Intracranial stent Vascular reconstruction device with markers. Mention material of stent and no. of markers. Mention sizes available in diameter x length. Flexible design for vessel conformability. Should be flared ends for stability. Coiling should be possible through the struts. 75 % should be retrievable. Should be FDA or CE & DCGI approved.	5	Pieces
3.122	Flow-diverter <b>with compatible delivery catheter and guidewire</b> should quoted as a single package Flow diverter - Resheathable Flow Diverter with advanced (4-Flex or similar) technology, Cobalt Chromium/Platinum Tungsten braided mesh design embolization device for treatment of intracranial large, giant, wide-necked, blister, failed-treatment & fusiform aneurysms. Should be FDA or CE & DCGI approved.	5	Pieces

3.123	<b>Aneurysm occlusion device</b> that works like a flow diverter - covered implant/device with radio-opaque markers at both ends <b>mounted on balloon-catheter</b> . Mention the make – metallic and/or membrane-based. Should be able to deliver over 014 wire and through standard guiding catheter (6F 070”). Mention sizes (diameter x length) available. Should be FDA or CE & DCGI approved	5	Pieces
4.000	Mechanical Thrombectomy/retriever Device + delivery microcatheter + guidewire (All items should quoted as a single package rate and all materials should be compatible in size and other aspects) with each other and be usable for all routine cases.(4.124,4.125,4.126)		
4.124	<b>Mechanical Thrombectomy/retriever Device</b> Stent retriever Technology, closed cell design & closed end to prevent clot migration, Stent retriever of size 4 mm x 20 mm or more with effective clot capture length of 20 mm, Non-detachable stent retriever with multiple pass profile. Compatible with 0.021 inch inner diameter catheter. Mention the distal radio-opaque length. Should be FDA or CE & DCGI approved.	5	Pieces
4.125	Straight tip endhole single-lumen <b>Microcatheter</b> for super selective catheterization with 0.021 ID braided permitting any 0.014" guide wire, DMSO- Compatible & used with Stent/retriever thrombectomy device. Should be FDA or CE & DCGI approved.	5	Pieces
4.126	200 cm & above <b>guidewire</b> for navigation and positioning of Microcatheter in neurovasculature with good visibility. Tip shapeable & reshapable. One to One Torque transmission. Should be FDA or CE & DCGI approved.	5	Pieces
4.127	<b>Balloon Guiding Catheter</b> with Balloon tip 8F or 9F in size. Design for arresting flow in Thrombectomy procedures. Dual Lumen. Length of 80 or 95 cm. Should be FDA or CE & DCGI approved.	5	Pieces
5.000	Coils for intracranial use (aneurysms & others) 1. All coils to have detachment time of less than 5 seconds (the least time preferred). Mention the detachment mechanism (mechanical, electrical or otherwise).2. Complete detachment (cables, etc) should be provided without any extra cost for set of 10 coils in each category. Describe the no. of coils that can be used for the detachment system.		
5.128	<b>Complex coil:</b> Complex shape coils, stretch resistant, 12 system or less, thinner core wire, variable or random loop design with 2D & 3D conformation to aneurysm shape. Same coil should be usable for both framing and filling. Mention the size range available. Should be FDA or CE & DCGI approved.	20	Pieces

5.129	<b>3D coils (10 system)</b> Pre-shaped framing coils, stretch resistant, Mention specific features, if any (like smaller first loop for atraumatic aneurysm entry, etc). Mention size range available. Should be FDA or CE & DCGI approved	20	Pieces
5.130	<b>3D coils (18 system)</b> Pre-shaped framing coils, stretch resistant, Mention specific features, if any (like smaller first loop for atraumatic aneurysm entry, etc). Mention size range available. Should be FDA or CE & DCGI approved.	10	Pieces
5.131	<b>2D coils (18 system)</b> – for intracranial use in parent artery occlusion and carotico-cavernous fistula. Should be FDA or CE & DCGI approved	25	Pieces
5.132	<b>Bio-active coils – (10 or 18 system)</b> – specify constituents, sizes available and to quote separately if 10 and 18 system available. Should be FDA or CE & DCGI approved	10	Pieces
5.133	<b>Vertebroplasty kit</b> contains bone biopsy needle, bone cement, liquid and syringes Should be FDA/CE/DCGI approved	2	Pieces
5.134	<b>Transjugular liver biopsy set</b> (to be quoted together) Tranjugular Liver access and biopsy set having Quick core biopsy needle, Check-flo sheath, Straight catheter, Beacon tip MPA catheter and dilator. Should be FDA or CE & DCGI approved.	10	Pieces
5.135	<b>Microvascular plug with delivery system</b> (sheath/catheter) PTFE Covered Nitinol design, full control over deployment. fully resheathable after complete deployment, mention length & width. Specify microcatheter size compatibility. Mention mode of detachment (mechanical or electrical), detachment system/cable should be quoted together. Should be FDA or CE & DCGI approved	10	Pieces
5.136	<b>Y connector push-pull type</b> with 2-way & 3-way tuohy-burst adjustable valves	200	Pieces

5.137	Luer lock syringe 10 cc	<b>100</b>	Pieces
5.138	Luer lock syringe 3 cc	<b>100</b>	Pieces
5.139	Inj Sodium tetradecyl sulphate 3% 30 Mg /2 MI	<b>100</b>	Pieces
5.140	Injection Nimodipine 10 mg/ml 50 MI Vial	<b>25</b>	Pieces
5.141	Injection Papverine 30 mg/ml 2 MI Vial	<b>25</b>	Pieces
5.142	Inj. Abciximab 2 mg/ml	<b>10</b>	Pieces
5.143	Inj.Thrombin (Human lyophilized) – 500 IU with water for injection	<b>10</b>	Pieces
5.144	Bedside INR Measurement Kits	<b>10</b>	Pieces
5.145	Arterial Pressure monitoring Transducers (Disposable)	<b>25</b>	Pieces
5.146	Thyroid Collars Radiation Protection	<b>10</b>	Pieces

5.147	Radiation Protective Eye Goggles	10	Pieces
5.148	Cath-Lab Cotton Dress Shirt with three pockets and Pyjama with pockets Large and medium Sized Light Blue Colour with label of "DSA JIPMER" on both the shirt and the pyjama and Size printed	100	Pieces
5.149	Disposable Cath-Lab Gowns Light Blue Colour full size and elastic wrist bands with water/ blood impenetrable front and hand shielding	200	Pieces
5.150	One set of high grade stainless steel surgical instruments for minor surgical procedure. Must include the following instruments – Sponge forceps 2 Nos, Towel clamp – 4, Hemostatic Crile forceps straight – 4, hemostatic Crile forceps curved – 6, Allis tissue forceps – 2, Mayo Hegar needle holder 6 inches – 1, Mayo Hegar needle holder 7 inches – 1, Mayo scissors straight – 1, Metzenbaum scissors curved – 1, Small curved scissors – 1, small straight scissors – 1, thumb forceps – 2, adsons toothed forceps – 2, Knife handle #7 – 1, knife handle #4 – 1, Weitlaner self retaining retractor sharp 5 inches – 1, double hook retractor – 2, right angle forceps – 1	5	Pieces
5.151	(S)Vinyl layered LEAD FREE aprons superior quality imported types for radiation protection during interventional procedures. COVERING BOTH FRONT AND BACK TWO PIECE SKIRT AND BLOUSE TYPE MEDIUM SIZE	4	Pieces
5.152	(S)Vinyl layered LEAD FREE aprons superior quality imported types for radiation protection during interventional procedures. COVERING BOTH FRONT AND BACK TWO PIECE SKIRT AND BLOUSE TYPE LARGE SIZE	4	Pieces

#### **OTHER TERMS AND CONDITION**

Samples are mandatory for all the above mentioned items for evaluation and should be submitted within 10 days of opening the bid and if samples not submitted the tender will not be accepted. All implantable devices should be licensed for sale in India by DCGI, and should be USFDA approved.

- All implantable devices should be supplied only by companies that has supplied to major hospitals (500 beds) in India or abroad (proof to be enclosed)
- Items which come in packs - mention price per unit and pack size.
- Patient size variable implantable devices and consumables (e.g. Stents/balloons/coils/ etc), the supplier should exchange the desired size if one size is exhausted so as to avoid device-patient mismatch till the last item is used
- Items supplied should have an expiry date of more than 24 months from date of supply.

## Section-III

### **GENERAL INSTRUCTIONS TO TENDERERS**

#### **1. Period of contract:**

The contract shall initially be for a period of **one year** that may be extended to **6 months**. The rates approved shall remain unchanged during the period of contract.

#### **2. Bid Security (Earnest Money Deposit)**

- i. In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender.
- ii. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.
- iii. The bidder hereby declares that they accept the condition that if they withdraw or modify their bids during period of validity etc., they will be suspended for a period of two years from participating in any tender invited by JIPMER, Puducherry with effect from the date of their withdrawal or modification of their bid.

#### **3. Eligibility conditions of bidders:**

- i. The Tenderer must be a domestic manufacturer that is a 'Class-I local supplier' or a 'Class-II local supplier' as defined under Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as further amended by orders of even number dated 28.05.2018, 29.05.2019, 04.06.2020 and 16.09.2020. In case the manufacturer does not quote directly, they may authorize an agent as per proforma of Manufacturer authorization form as given in the Tender enquiry document to quote and enter into a contractual obligation.
- ii. In compliance with order (Public Procurement No.1) No. 6/18/2019-PPD dated 23rd July 2020 issued by the Public Procurement Division, Dept. of Expenditure, Min of Finance under Rule 144(xi) of GFR 2017 any bidder 'from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority.
  - a. "Bidder" (including the term 'tenderer', 'consultant' or 'service provider' in certain contexts) means any person or firm or company, including any member of a consortium or joint venture (that is an association of several persons, or firms or companies), every artificial juridical person not falling in any of the descriptions of bidders stated hereinbefore, including any agency branch or office controlled by such person, participating in a procurement process.
  - b. "Bidder from a country which shares a land border with India" for the purpose of this Order means: -



- I. An entity incorporated, established or registered in such a country; or
- II. A subsidiary of an entity incorporated, established or registered in such a country; or
- III. An entity substantially controlled through entities incorporated, established or registered in such a country; or
- IV. An entity whose beneficial owner is situated in such a country; or
- V. An Indian (or other) agent of such an entity; or
- VI. A natural person who is a citizen of such a country; or
- VII. A consortium or joint venture where any member of the consortium or joint venture falls under any of the above

c. The beneficial owner for the purpose of (5.2.2) above will be as under:

- I. In case of a company or Limited Liability Partnership, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has a controlling ownership interest or who exercises control through other means.

Explanation—

- "Controlling ownership interest" means ownership of or entitlement to more than twenty-five per cent, of shares or capital or profits of the company;
  - "Control" shall include the right to appoint majority of the directors or to control the management or policy decisions including by virtue of their shareholding or management rights or shareholders agreements or voting agreements;
- II. In case of a partnership firm, the beneficial owner is the natural person(s) who, whether acting alone or together, or through one or more juridical person, has ownership of entitlement to more than fifteen percent of capital or profits of the partnership.
  - III. In case of an unincorporated association or body of individuals, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has ownership of or entitlement to more than fifteen percent of the property or capital or profits of such association or body of individuals.
  - IV. Where no natural person is identified under (a) or (b) or (c) above, the beneficial owner is the relevant natural person who holds the position of senior managing official;
  - V. In case of a trust, the identification of the beneficial owner(s) shall include the identification of the author of the trust, the trustee, the beneficiaries with 15% or more interest in the trust and any other natural person exercising ultimate effective control over the trust through a chain of control or ownership.

d. An Agent is a person employed to do any act for another, or to represent another in dealings with third person.

e. In case of turnkey contracts, the successful bidder shall not be allowed to sub-contract works to any contractor from a country which shares a land border with India unless such contractor is registered with the Competent Authority.

- f. Bidders must certify compliance with the above mentioned provisions in the tender form as per Section VII.
- g. The bidder should have successfully completed the delivery of the item of at least 25% of the quantity required in this tender to a government hospital in the last 3 years.
- h. The bidder or the manufacturer must have an average annual turnover of at least Rs.1,00,00,000/-(Rupees one crore only) each during the last three financial years i.e. 2017-18, 2018-19, 2019-20.
- i. Bidder should have ISO Certification.

#### **4. Purchase Preference**

- i. The Procurement of goods and services under this e-tender will be regulated as per the applicable provisions of Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as further amended by orders of even number dated 28.05.2018, 29.05.2019, 04.06.2020 and 16.09.2020. Salient portions of the order are reproduced in Appendix A of this tender document by way of information. Bidders are advised to see the original orders and satisfy themselves that they qualify to participate in the tender. Bidders who are claiming eligibility to participate in this tender must submit a certificate in format given in Appendix A along with documentary evidence in support of their claim wherever necessary failing which their bid will be summarily rejected. The purchaser reserves the right to give preference to the 'Class-I local supplier'.
- ii. The Purchaser reserves the right to give the purchase preference to small-scale sectors, Micro and small scale enterprises etc. as per the instruction in vogue while evaluating, comparing and ranking the responsive Tenders as given in the MSMED Act 2006 reproduced below:
  - a. In exercise of powers conferred in Section 11 of the Micro, Small and Medium Enterprises Development (MSMED) Act 2006, the Government has notified a new Public Procurement Policy for Micro & Small Enterprises effective from 1st April 2012. The policy mandates that 25% of procurement of annual requirement of goods and services by all Central Ministries/Public Sector Undertakings will be from the micro and small enterprises. The Government has also earmarked a sub-target of 4% procurement of goods & services from MSEs owned by SC/ST entrepreneurs out of above said 25% quantity.
  - b. In accordance with the above said notification, the participating Micro and Small Enterprises (MSEs) in a tender, quoting price within the band of L1+15% would also be allowed to supply a portion of the requirement by bringing down their price to the L1 price, in a situation where L1 price is from someone other than an MSE. Such MSEs would be allowed to supply up to 25% of the total tendered value. In case there are more than one such eligible MSE, the 25% supply will be shared equally. Out of 25% of the quantity earmarked for supply from MSEs, 5% quantity is earmarked for procurement from MSEs owned by SC/ST entrepreneurs. However, in the event of failure of such MSEs to participate in the Tender process or meet the

tender requirements and the L1 price, the 5% quantity earmarked for MSEs owned by SC/ST entrepreneurs will be met from other participating MSEs.

- c. The MSEs fulfilling the prescribed eligibility criteria and participating in the Tender shall enclose with their tender a copy of their valid registration certificate with District Industries Centres or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir Board or National Small Industries Corporation or any other body specified by Ministry of Micro and Small enterprises in support of their being an MSE, failing which their tender will be liable to be ignored.
- d. Special provision for Micro and Small Enterprise owned by women: – Out of the total annual procurement from Micro and Small Enterprises, 3 per cent from within the 25 per cent target shall be earmarked for procurement from Micro and Small Enterprises owned by women.
- e. Note: “If the bidder is a MSME, it shall declare in the bid document the Udyog Aadhar Memorandum Number issued to it under the MSMED Act, 2006. If a MSME bidder do not furnish the UAM Number along with bid documents, such MSME unit will not be eligible for the benefits available under Public Procurement Policy for MSEs Order 2012.

## **5. Demonstration of sample**

The tenderer must give a sample within **ten** days of the closing of online submission of bids failing which the bids will be rejected.

## **6. Instructions for the filling the tender form**

- i. E-Tender form shall be completed in all respect, signed in full and stamped at appropriate places and initialed and stamped on all remaining pages. Incomplete or e-tenders without tender processing fee, bid security declaration, Make-in-India self-certification as in Appendix-A only by original manufacturer, GFR 144 (xi) compliance certificate as in Section-X shall be treated as invalid.
- ii. Bidders have to ensure that all the documents are properly filled.
- iii. Conditional tenders are liable to be rejected.
- iv. Bids received and found valid will be evaluated by JIPMER to ascertain the complete work/services under the specification and documents. The bidder should take care to submit all the information sought by JIPMER in prescribed formats.
- v. Incomplete bids, bids in paper format, conditional bids, telephonic bids or tenders submitted after the due date and time will not be considered and summarily rejected. Vendors are, therefore, advised to submit their bids well on time.
- vi. **The bidder can quote for one or more items mentioned in the list. Bidder has to give all details (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) mentioned in BOQ for all quoted items, failure of that the bid will be rejected summarily.**

## 7. Submission of tenders

The bidders must ensure that they submit the **on-line bids** within the scheduled closing date & time.

## 8. Late Tender:

There is NO PROVISION of uploading late tender beyond stipulated date & time in the e-tendering system.

## 9. Alteration and Withdrawal of Tender

- i. The bidder, after submitting its bid, is permitted to alter/modify its bid, within the deadline for submission of bids. Alterations/modifications to bids received after the prescribed deadline will not be possible on the e-tender portal.
- ii. No tender should be withdrawn or modified after the deadline for submission of tender and before expiry of the tender validity period. If a bidder withdraws or modifies the tender during this period, it will result in the tenderer being suspended from participating in all tenders invited by JIPMER for a period of two years with effect from the date of withdrawal, as per the terms and conditions of bid security declaration furnished by the bidder in its tender.

## 10. Preparation of e-tenders

This is a Two-Bid Tender system, consisting of the **Techno-Commercial Bid and Price Bid** that are to be uploaded in the prescribed formats in the e-tendering portal. The tender(s) shall only be submitted online as mentioned below:

### I. **Techno-commercial Bid shall comprise**

- a. Fee Cover
  - i. **E-tender Processing fee** of Rs.590/- (Rupees five hundred and ninety only) inclusive of 18% GST payment **receipt** duly self-attested and rubber stamped should be uploaded.
  - ii. Scanned copy (100 or 200 DPI) in pdf format of **Bid Security Declaration** as per format attached in Section-IX
- b. In the cover named “Prequal/Technical” the scanned copy in pdf format of the following documents are to be uploaded:
  - i. The **Bidder’s Profile** as in Section-VI, **Tender Form** as in Section-VII must be downloaded duly filled signed and stamped, and an **Authorization letter** for signing tender documents if a person other than the Owner, Partner, Managing Director is signing the documents, must be uploaded as a single PDF file.

- ii. **Copies of Supply orders/Completion certificate** in support of Eligibility condition 1 and 2 and **Copies of abridged Annual report of last 03 years** (Income tax return acknowledgement, Assets and Liabilities, Balance sheet and Profit & Loss Account) must be uploaded as a single PDF file
- iii. **Manufacturer's Authorization letter** in company letterhead in format in given Section-VI and, in case the bidder is empaneled by the Competent Authority under **GFR 144 (xi) (mandatory)** a copy of the same or **GFR 144(xi) compliance certificate** as in Section-X must be uploaded.
- iv. Copy of Self Certified **GST registration certificate** and Copy of **PAN Card** must be uploaded as a single PDF file.
- v. **Bank Details** (Beneficiary name, Bank name, Account number, IFSC code, Branch address on letterhead) and **Check list** as in Section-VIII in the prescribed format duly filled and signed must be uploaded as a single PDF file.
- vi. Self-certification for supporting the claim to be a local supplier under the "Public Procurement preference to Make in India" order in format as in **Appendix-A (mandatory)** and any other document that the bidder wishes to submit as a single PDF file.
- vii. **A file mentioning the list of items without price** for which bidder is quoting must be uploaded as a single PDF file.

## **II. Price Bid:**

Prices are to be quoted in the prescribed Price Bid format provided in the e-tender portal using the BOQ template only. The price should be quoted for the **accounting unit** indicated in the e-tender document.

### **Note:**

- i) The bidder has to be diligent while filling up the Techno-Commercial Bid and Price Bid provided in prescribed formats and must not tamper with the contents of the sheets.
- ii) Bidders must ensure that the documents uploaded in pdf format are legible.
- iii) It is the responsibility of bidder to go through the Tender document to ensure furnishing all required documents in addition to above, if any.
- iv) The Make in India self-certification as in Appendix-A must be given and uploaded by original manufacture. The distributor can upload the Appendix-A given by original manufacture.
- v) ITE- Item-wise Eligibility Sheet should be downloaded, the items that the bidder wishes to quote must be selected as "Eligible", and this "ITE file" must also be uploaded for the price bid to be considered by the system. The selected items will be displayed once uploaded and the bidder can verify that all items he wishes to quote for, are present in the list
- vi) A person signing (manually or digitally) the tender form or any documents forming

part of the contract on behalf of another shall be deemed to warrant that he has authority to bind such other persons and if, on enquiry, it appears that the persons so signing had no authority to do so, the purchaser may, without prejudice to other civil and criminal remedies, cancel the contract and hold the signatory liable for all cost and damages.

- vii) A bid, which does not fulfill any of the above requirements and/ or give evasive information/reply against any such requirement, shall be liable to be ignored.
- viii) Tender sent by fax/telex/cable shall be ignored.

#### **11. Digital Signing of Tender**

The tenderers shall submit their tenders as per the instructions contained as above. Tenders shall be uploaded with all relevant tender documents in the prescribed format. The relevant tender documents should be uploaded by an authorized person having Class 3 digital signature certificate.

#### **12. Tender currencies.**

The tender shall be quoted only in INR.

#### **13. Additional information and instruction on GST:**

If the Tenderer desires to ask for GST or any other taxes to be paid extra, the same must be specifically stated. In the absence of any such stipulation, the price will be taken inclusive of such taxes and no claim for the same will be entertained later. The rate of GST quoted in the tender shall be taken for price comparison. However, the rate of GST quoted in the tender or the actual rate of GST applicable, whichever is lower shall be payable by the purchaser. The supplier can charge a higher GST than quoted in the tender only if the rate of GST was revised by the government after the tender closing date.

#### **14. Bid Security (Earnest Money Deposit)**

- i. In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit<sup>tl</sup>, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender.
- ii. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration<sup>l</sup> on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.
- iii. The bidder hereby declares that they accept the condition that if they withdraw or modify their bids during period of validity etc., they will be suspended for a period of two years from participating in any tender invited by JIPMER, Puducherry with effect from the date of their withdrawal or modification of their bid.

#### **15. Tender opening**

- i. The Tender Inviting Authority will open the e-tenders at the specified date and time and at the specified place as indicated in the NIT. In case the specified date of tender opening falls on / is subsequently declared a holiday or closed day for the Tender Inviting Authority, the tenders will be opened at the appointed time and

place on the next working day.

- ii. Authorized representatives of the tenderers, who have submitted tenders on time, may attend the tender opening provided they bring with them letters of authority from the corresponding tenderers. The tender opening official(s) will prepare a list of the representatives attending the tender opening. The list will contain the representatives' names & signatures and corresponding tenderers' names and addresses.
- iii. This being a Two-Bid Tender system, the **Techno-Commercial Tenders** are to be opened in the first instance, at the prescribed time and date as indicated in NIT. These Tenders shall be scrutinized and evaluated by the competent committee/ authority with reference to parameters prescribed in the tender document. During the Techno- Commercial Tender opening, the tender opening official(s) will read the salient features of the tenders like brief description of the goods offered, delivery period, bid security declaration and any other special features of the tenders, as deemed fit by the tender opening official(s). Thereafter, in the second stage, the Price bids of only the Techno - Commercially acceptable offers (as decided in the first stage) shall be opened for further scrutiny and evaluation on a date notified after the evaluation of the Techno-Commercial tender.

## **16. Scrutiny and evaluation of tenders**

### **A. Basic Principle**

Tenders will be evaluated on the basis of the terms & conditions already incorporated in the tender enquiry document, based on which tenders have been received and the terms, conditions etc. mentioned by the tenderers in their tenders. No new condition will be brought in while scrutinizing and evaluating the tenders.

### **B. Scrutiny of Tenders**

- i. The Purchaser will examine the Tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished and, whether the documents uploaded are in legible form.
- ii. The Purchaser's determination of a Tender's responsiveness is to be based on the contents of the tender itself without recourse to extrinsic evidence.
- iii. The tenders will be scrutinized to determine whether they are complete and meet the essential and important requirements, conditions etc. as prescribed in the tender document. The tenders, which do not meet the basic requirements, are liable to be treated as non-responsive and will be summarily ignored.
- iv. The following are some of the important aspects, for which a tender shall be declared non-responsive during the evaluation and will be ignored;
  - i. Tender validity is shorter than the required period.
  - ii. Non-submission of Bid Security Declaration.
  - iii. Non-submission of receipt of tender processing fee.
  - iv. Non-submission of self-certification in format as given in Appendix-A only by original manufacturer, for determining eligibility to participate in the tender under the "Public Procurement preference to Make in India" order.
  - v. Non-submission of GFR-144 (xi) compliance certificate.

- vi. Tenderer has not agreed to give the required performance security of required amount in an acceptable form.
- vii. Non-submission of samples within ten days of the closing of online submission of bids
- viii. Poor/ unsatisfactory past performance.
- ix. Tenderers who stand de-registered/banned/blacklisted by any Central Govt. Ministries/Departments/Hospitals/Institutes.
- x. Tenderer is not eligible as per tender conditions.
- xi. Tenderer has not quoted for the entire quantity as specified in the List of Requirements/ BOQ for the item quoted.
- xii. Non-submission of all details of quoted items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark).
- xiii. Tenderer has not agreed to other essential condition(s) specially incorporated in the tender enquiry, like delivery terms, delivery schedule, terms of payment, liquidated damages clause, warranty clause, dispute resolution mechanism applicable law.

#### **17. Minor Informality/Irregularity/Non-Conformity**

If during the preliminary examination, the purchaser find any minor informality and/or irregularity and/or non-conformity in a tender, the purchaser may waive the same provided it does not constitute any material deviation and financial impact and, also, does not prejudice or affect the ranking order of the tenders. Wherever necessary, the purchaser will convey its observation on such 'minor' issues to the tenderer by registered/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.

#### **18. Award of work**

- i. The selection of the agency will be at the sole discretion of the JIPMER who reserves its right to accept or reject any or all the proposals without assigning any reason thereof.
- ii. The lowest bid will be evaluated on the basis of annual estimated amount as quoted by bidders in the quotation including taxes.
- iii. Upon evaluation of offers the decision on the award of contract will be intimated to the successful bidder.
- iv. The annual estimate is given only as an indication. The actual quantity procured may increase or decrease. No assurance is given that the quantity stated will actually be procured.



## Section-IV

### GENERAL CONDITIONS OF CONTRACT

#### 1. **Price of goods**

The rate quoted in the e-tender will be fixed for the whole contract period.

#### 2. **Technical Specifications and Standards**

The Goods & Services to be provided by the supplier under this contract shall conform to the technical specifications mentioned in “Technical Specification” under Section II.

#### 3. **Terms of Delivery**

- i. Goods shall be delivered by the supplier within 30 days of issue of supply order. Please note that the time shall be the essence of the contract.
- ii. Any unexcused delay by the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all of the following sanctions:
  - a) Imposition of liquidated damages,
  - b) Forfeiture of its performance security and
  - c) Termination of the contract for default.

#### 4. **Liquidated Damages**

If the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods, installation, commissioning and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 11. Since the Liquidated damages are in virtue of non-performance of services, it will attract GST or any other applicable taxes which in turn shall be deducted from the bidder.

#### 5. **Performance Security**

The successful bidder shall have to deposit an amount 3% of the value of contract as Performance Security Deposit (PSD) within two weeks after award of contract, through SBI Collect available on JIPMER website. In the event of any failure /default of the supplier with or without any quantifiable loss to the purchaser, the amount of the performance security is liable to be forfeited.

Subject to the condition mentioned above the Performance Security will be released without any interest to the supplier on completion of the supplier's all contractual obligations

including the warranty obligations and extension of time (with or without Liquidated Damages).

#### **6. Payment Procedure**

The contractor shall submit bill in triplicate upon satisfactory supply of goods. Bill must be raised based on the rate quoted in e-tender. Every effort shall be made to ensure that the payment will be made within 45 days of submission of bill. No advance payment will be considered. TDS will be deducted as per provision of Income Tax Act, GST Acts and other statutes as relevant.

#### **7. Risk Clause**

The contractor shall at all times have standby arrangements for carrying out the work under the contract, in case of any failure of the existing arrangements. JIPMER reserves the right for termination of the contract at any time by giving 30 days written notice, if the items delivered are found to be unsatisfactory and also has the right to award the contract to the next higher bidder willing to supply the item at the cost, risk and responsibilities of contractor and excess expenditure incurred on account of this will be recovered by JIPMER from the contractor's Performance Security Deposit or pending bills or by raising a separate claim.

#### **8. Termination clause:**

During the period of agreement if it is found that the agency is not providing proper services, the JIPMER reserves rights to make the vendor forfeit the security deposit deposited with JIPMER or part thereof in favour of JIPMER and agreement will be terminated after giving 30 days' notice. Furthermore, in such situations, tender can be allotted to second lowest bidder and the difference in cost shall be recovered from the earlier vendor who is breach of the contract. In addition in case it is found that the supplier is charging by fraudulent means or indulging in criminal activities the contract will be terminated immediately.

#### **9. Jurisdiction of the courts**

Jurisdiction of the courts for settlement of disputes:- Jurisdiction for the settlements of disputes if any is Puducherry only.

In above mentioned conditions Director JIPMER reserves all the rights.

**Section-V**

**BIDDER'S PROFILE**

**This form duly filled and signed by authorized representative of the bidder and the scanned copy must be uploaded online**

1.	Name & Designation of the contact person	
2.	Name and Address of the Tenderer	
3.	Phone No a) Land line number (functional between 9 am and 5pm)	
4.	Mobile No of contact person (available from 9am to 6pm)	
5.	Email ID of the Tenderer	
6.	Email ID of the contact person	
7.	Local supplier/Distributor in Chennai/ Puducherry or any other place (complete address must be written)	
8.	Manufacture Name	
9.	Manufacture Address	
10	Whether Tenderer is registered MSE Manufacture the product quoted. (If registered MSE, submit copy of the Udyog Aadhaar certificate or Equivalent Certificate)	Yes / No
<p>If there is any change in the above details, I will immediately intimate you by speed post or fax or email</p> <p>I.....hereby declare that the details given above are true to the best of my knowledge and I have thoroughly read and understood the terms and conditions of the tender and shall abide by the rules,</p> <p style="text-align: right;">Signature (Name and Designation &amp; Seal)</p> <p>Dated:</p>		

**NB: This form must be duly filled in by an authorized person**

**Section-VI**  
**MANUFACTURER'S AUTHORISATION FORM**  
**(Letterhead)**

The Director,  
JIPMER, Puducherry

Dear Sir,

Ref: Your TE document No: \_\_\_\_\_ dated: \_\_\_\_\_

We, \_\_\_\_\_ who are proven and reputable manufacturers of \_\_\_\_\_ (name and description of the goods offered in the Tender) having factories at \_\_\_\_\_, hereby authorize Messrs. \_\_\_\_\_ (name and address of the agent) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also state that we are not participating directly in this tender for the following reason(s):  
\_\_\_\_\_. (Please provide reason here).

We further confirm that no supplier or firm or individual other than Messrs \_\_\_\_\_ (name and address of the above agent) is authorised to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us. We also hereby extend our full warranty, CMC/AMC as applicable as per the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document. We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorised agent.

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly”

Yours faithfully,  
[Signature with date, name, designation and Email]  
for and on behalf of Messrs \_\_\_\_\_  
[Name & address of the manufacturers]

Note:

- (1) This letter of authorization should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.
- (2) Original letter may be sent.
- (3) The purchaser reserves the right to verify this document with its signatory.

**SECTION-VII**  
**TENDER FORM**  
**(On Firms' Letter Head)**

To  
The Director  
JIPMER, Puducherry 605006

Date \_\_\_\_\_

Ref. Your TE document No. \_\_\_\_\_ dated \_\_\_\_\_

We, the undersigned have examined the above mentioned TE document, including amendment/corrigendum No. \_\_\_\_\_, dated \_\_\_\_\_ (if any), the receipt of which is hereby confirmed. We now offer to supply and deliver \_\_\_\_\_ (Description of goods and services) in conformity with your above referred document **for the sum as shown in the price schedules attached herewith and made part of this tender**. If our tender is accepted, we undertake to supply the goods and perform the services as mentioned above, in accordance with the delivery schedule specified in the List of Requirements.

We further confirm that, if our tender is accepted, we shall provide you with a performance security of required amount in an acceptable form in terms of GCC, Special Conditions of Contract", for due performance of the contract.

We have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and we certify that this bidder is not from such a country/ from such a country and has been registered with the Competent Authority and a copy of the valid registration by the Competent Authority is attached as evidence of the same (Strike out what is not applicable). In case there are Turnkey works to be carried out this bidder will not sub-contract any work to a contractor from such countries unless such contractor is registered with the Competent Authority. We hereby certify that this bidder fulfills all requirements in this regard and is eligible to be considered.

We agree to keep our tender valid for acceptance as required in the GIT, Special Instructions to Tenderers" or for subsequently extended period, if any, agreed to by us. We also accordingly confirm to abide by this tender up to the aforesaid period and this tender may be accepted any time before the expiry of the aforesaid period. We further confirm that, until a formal contract is executed, this tender read with your written acceptance thereof within the aforesaid period shall constitute a binding contract between us.

We further understand that you are not bound to accept the lowest or any tender you may receive against your above-referred tender enquiry.

We confirm that we do not stand deregistered/banned/blacklisted by any Govt. Authorities in the last 7 years.

We confirm that we fully agree to the terms and conditions specified in above mentioned TE document, including amendment/ corrigendum if any.

(Signature with date)  
(Name and designation)

Duly authorized to sign tender for and on behalf of

**Section-VIII**

**CHECK LIST FOR SUBMISSION OF TENDER**

(To be filled by the tenderer and submitted along with the bid)

Sl. No.	Particular	Yes / No
1	i. Tender Processing Fee (Mandatory for all bidders) ii. Bid security declaration	
2	Copies of Supply orders/Completion certificate in support of Eligibility condition 2(ii).	
3	Copy of PAN Card	
4	Copy of ISO Certification of bidder, ISI/CE Certification of each product wherever asked	
5	Copies of last three years Income Tax Return statement with Balance sheets & Profit & Loss A/c i.e. 2017-18, 2018-19, 2019-20.	
6	Copy of GST Registration Certificate	
7	Authorization letter for signing tender documents if a person other than the Owner, Partner, Managing Director is signing/uploading the tender	
8	Tender form as in Section – VII duly signed and stamped	
9	Manufacturer's authorization form in format as in Section VI and	
10	Self-certification for supporting the claim to be a local supplier under the "Public Procurement preference to Make in India" order in format as in Appendix A	
11	A copy bidder's empanelment by the Competent Authority under GFR 144 (xi) or GFR 144 (xi) compliance certificate	
12	A PDF file containing list of all items quoted by the bidder without price bid in technical cover	
13	All details of the items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) quoted by the bidder.	
14	Any other document(s) enclosed (To be specified)	

I/We certify that the information furnished above is true and correct. The terms and conditions are acceptable to us and have the authority to bid a tender.

Signature of the owner/  
Managing Partner/Director

Name:

Seal:

Date:

Place:

**SECTION – IX**

**Bid Security Declaration  
(To printed on the Firm's letterhead)**

(In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.)

I/We hereby declare that we accept the condition that if I/We withdraw or modify our bids during period of validity etc., we will be suspended for a period of two years from participating in all tenders invited by JIPMER, Puducherry with effect from the date of our withdrawal or modification of our bid.

(Signature with date)

(Name and designation)

Duly authorized to sign this tender and declaration

for and on behalf of \_\_\_\_\_ (Name of the Bidder's Firm) Seal

## Appendix-A

### Self-certification format for claiming purchase preference under the “Public Procurement preference to Make in India” order

As per the order issued by

(i) Department of Industrial Policy and Promotion (DIPP) vide No. P-45021/2/2017-BE-II dated 15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017-B.E.-II dated 29.05.2019, Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020 and Order No. P-45021/2/2017-PP (BE-II) dated 16.09.2020; and

(ii) Department of Pharmaceuticals vide No. F- 31026/36/2016-MD dated 18.05.2018 and the subsequent orders thereof; The purchaser reserves the right to give preference to the local supplier.

A local supplier (definition of “local supplier” is given in clause 2 of the aforesaid order of DIPP as amended from time to time) has to submit the following along with their e-tender(s) failing which their bid will be evaluated without considering such preference mentioned in the DIPP order dated 15.06.2017 further amended on 28.05.2018, 25.09.2019 and 04.06.2020:

a. The local supplier at the time of e-tender, bidding or solicitation shall be required to provide self-certification that the item offered meets the minimum local content and shall give details of the location(s) at which the local value addition is made in the format in Annexure A.

“Certified that the following items quoted has more than 50% or 20% to 50% of value addition in India at the location(s) mentioned against each and is eligible for purchase preference as per the Govt. of India “Public Procurement preference to Make in India” order Dt.15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017-B.E.-II dated 29.05.2019, and Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020.

Sl. No	Name of the item	Details of the location(s) at which the local value addition was made.

Authorized Signature:

Name:

Designation:

b. In cases of procurement for a value in excess of Rs. 10 Crore. the local supplier shall be required to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.

c. Minimum Local Content: ‘Class-I local supplier’ means a supplier or service provider, whose goods, services or works offered for procurement, has local content equal to or more than 50%, as defined



under the Order. “Class-II local supplier” means a supplier or service provider, whose goods, services or works offered for procurement, has local content more than 20% but less than 50%, as defined under this Order.

d. Margin of Purchase Preference: The margin of purchase preference shall be 20%.

e. Manufacture under license/technology collaboration agreements with phased indigenization are exempted from meeting the stipulated local content if the product is being manufactured in India under a license from a foreign manufacturer who holds intellectual property rights and where there is a technology collaboration agreement/transfer of technology agreement for indigenous manufacture of a product developed abroad with clear phasing of increase in local content

f. Decisions on complaints relating to implementation of this Order shall be taken by the competent authority which is empowered to look into procurement-related complaints relating the procuring entity.

g. A constituted committee with internal and external experts will examine for independent verification of self-declarations and auditor’s/accountant’s certificates on random basis and in the case of complaints.

h. A fees of Rs.10000/- in the form of demand draft favoring The Director, JIPMER, payable at Puducherry, is required to be deposited with complaints for verification of local content.

i. False declarations will be breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.

j. A supplier who has been debarred by any procuring entity for violation of this Order shall not be eligible for preference under this Order for procurement by any other procuring entity for the duration of the debarment. The debarment for such other procuring entities shall take effect prospectively from the date on which it comes to the notice of other procurement entities.

Officer in Charge  
Purchase Section  
JIPMER Puducherry  
For Director, JIPMER

**SECTION – X**

**GFR-144 (xi) compliance certificate  
(To printed on the Firm's letterhead)**

Tender No:

**GFR-144(xi) compliance certificate** (as per order F.No. 6/18/2019-PPD, Ministry of Finance, GOI)

I have read the clauses regarding restrictions under GFR144(xi) on procurement from a bidder of a country which shares a land border with India. I certify that ....., the vendor

- is not such a country
- is from a country and has been registered with a competent authority (attached evidence of valid registration).

(Select one of the above and strike off the other)

I hereby certify that we fulfill all requirement in this regard and is eligible to be considered for the procurement on CPP portal.

Thanking you.

Authorized Signatory