



**JAWAHARLAL INSTITUTE OF POST GRADUATE MEDICAL
EDUCATION & RESEARCH (JIPMER)**



GOVERNMENT OF INDIA

(An Institution of National Importance under Ministry of Health & Family Welfare) Dhanvantri
Nagar, Puducherry-605006

**OPEN TENDER
ENQUIRY DOCUMENT
FOR
RATE CONTRACT
FOR
SUPPLY OF CONSUMABLES/NON-CONSUMABLES
FOR
DEPARTMENT OF
MEDICAL GASTROENTEROLOGY**

(This document consisting of **48** pages)

**JIPMER,
DHANVANTARI NAGAR,
PUDUCHERRY 605006**

Phone:- 0413-2296016/15/47 ; Fax:- 0413-2272067/66

Email: mgeijipmer@gmail.com URL:- www.iipmer.edu.in

INDEX

Section	Topic	Page No.
Section-I	Notice inviting Tender (NIT)	03
Section-II	Schedule of Requirements and Specifications	05
Section-III	General Instructions to Tenderers (GIT)	56
Section-IV	General Conditions of Contract (GCC)	65
Section-V	Bid Application Form	67
Section-VI	Manufacturer's Authorization Form	68
Section-VII	Tender Form	69
Section-VIII	Check list for Submission of Tender	70
Section-IX	Bid Security Declaration	71
Appendix-A	Self-certification format for claiming purchase preference under the "Public Procurement preference to Make in India" order	72
Section-X	GFR-144 (xi) compliance certificate	74

SECTION-I

JAWAHARLAL INSTITUTE OF POSTGRADUATE MEDICAL EDUCATION AND RESEARCH PUDUCHERRY – 6

(Institution of National Importance under the Ministry of Health & Family Welfare,
Govt. of India)

Purchase Section

No. JIP/PUR6(1)/OeT/MGE/2021-22

Dated: 01.07.2021

NOTICE INVITING e-TENDER (e-NIT)

Sub: Supply of Consumables / Non-Consumables

E- tenders are invited from eligible and qualified domestic manufacturer's or their distributors for the RATE CONTRACT for supply of CONSUMABLES/NON-CONSUMABLES for the Department of **Medical Gastroenterology** for 2021– 22.

1. Scope of work: Supply of consumables/non-consumables for the Department of Surgery
2. Value of tender: Approximate Cost **Rs. 1,00,00,000**
3. Tender timelines:
 - i. Opening date & time for download of Tender document: **06.00 PM on 01.07.2021**
 - ii. Last date for receipt of pre-bid queries: **12.00 Noon on 07.07.2021**
 - iii. Pre-bid queries can be made through e-mail to: mgejipmer@gmail.com.
 - iv. Opening date & time for submission of online bids: **09.00 AM on 13.07.2021**
 - v. Closing date & time for submission of online bids **12.00 Noon on 03.08.2021**
 - vi. Date and Time of opening of online bids for Technical evaluation: **11.00 AM on 04.08.2021**
 - vii. Date & time of opening of Price Bid: To be intimated later.
4. Earnest Money Deposit/Bid Security: Bid security declaration must be submitted by all bidders.
5. Tender Processing Fee (Non-refundable): **Rs. 590 (including 18% GST) shall be paid through SBI collect only.**
6. Interested bidders are advised to download the complete Tender Enquiry document from the websites www.jipmer.edu.in or <https://eprocure.gov.in/eprocure/app> for complete details.
7. The prospective bidders must register with the E-procurement system of <https://eprocure.gov.in/eprocure/app>. Special Instructions to the bidders for the e-submission of the bids online through this eProcurement Portal on completion of the registration process is given in <https://eprocure.gov.in/eprocure/app> , the bidders will be provided user ID and password upon enrollment. In order to submit the bids

electronically, bidders are required to have a valid Class 3 Digital Signature Certificate (signing and encryption/ decryption certificates).

8. Bidders are requested to read the bidders help document on e-tender web site link before proceeding for bidding.
9. Post receipt of User ID & Password, Bidders can log on for downloading & uploading tender document.
10. The bidders shall submit the required bid security declaration (as per G.I.T clause 2) before the due date and time mentioned above.
11. The online submission of tender(s) can only be done through **<https://eprocure.gov.in/eprocure/app>**
12. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **<https://eprocure.gov.in/eprocure/app>** e-portal (as described above) only.
13. Prospective bidders are advised to browse the above websites regularly before submission of their bids as any further amendments will be published in these websites only.

Asst. Officer in Charge
Purchase Section,
For
Director,
JIPMER, Puducherry

Section-II

SCHEDULE OF REQUIREMENTS AND SPECIFICATIONS

LIST OF ITEMS REQUIRED

1. Scope of work : Supply of Consumables/ Non-Consumables for the department of **Medical Gastroenterology** as detailed below

Sl. No.	Item Description	Quantity	Units
1.000	2	4	5
1.001	<p>Single use biopsy Forceps (UGI)</p> <ul style="list-style-type: none"> • Should have oval fenestrated cup to enable large biopsy fragments • Should have ribbed sheath to aid insertion & channel protection • Should have Unique SwingJaw mechanism to aid tangential biopsies • Suitable for 2.8 mm channel • Working length 1650 mm 	350	PIECE
1.002	<p>Single use biopsy Forceps (COLON)</p> <ul style="list-style-type: none"> • Should have oval fenestrated cup to enable large biopsy fragments • Should have ribbed sheath to aid insertion & channel protection • Should have Unique SwingJaw mechanism to aid tangential biopsies • Suitable for 2.8 mm channel • Working length 2300 mm 	150	PIECE
1.003	<p>Single use biopsy Forceps (SMALL BOWEL)</p> <ul style="list-style-type: none"> • Should have oval fenestrated cup to enable large biopsy fragments • Should have ribbed sheath to aid insertion & channel protection • Should have Unique SwingJaw mechanism to aid tangential biopsies • Suitable for 2.8 mm channel • Working length 2700 mm or more 	10	PIECE
1.004	<p>Sclerotherapy Injection Needle 21 G</p> <p>Specifications</p> <ol style="list-style-type: none"> 1. Needle with catheter, presterilized and disposable 2. Should have a stiff sheath for prevention of kinking and enhanced penetration into target. 3. Should have an ergonomic handle with positive click action. 4. Length of the needle should be 4 mm 5. Working length: 1650 mm compatible with 2.8 mm working channel 	70	PIECE

1.005	<p>Sclerotherapy Injection Needle 23 G Specifications</p> <ol style="list-style-type: none"> 1. Needle with catheter, presterilized and disposable 2. Should have a stiff sheath for prevention of kinking and enhanced penetration into target. 3. Should have an ergonomic handle with positive click action. 4. Length of the needle should be 4 mm 5. Working length: 2700 mm compatible with 2.8 mm working channel 	10	PIECE
1.006	<p>Sclerotherapy Needles for Colonoscopy/UGI use Specifications</p> <ol style="list-style-type: none"> 1. It should have star shaped inner catheter design to enable needle actuation in tortuous anatomy and ensuring minimum friction 2. Needle Gauge should be 25 G with maximum needle extension of 4 mm. 4. Outer sheath diameter: 2.3 mm 5. Working length: 200 cm 	10	PIECE
1.007	<p>Sclerotherapy Needles for Colonoscopy/UGI use Specifications</p> <ol style="list-style-type: none"> 1. It should have star shaped inner catheter design to enable needle actuation in tortuous anatomy and ensuring minimum friction 2. Needle Gauge should be 25 G with maximum needle extension of 4 mm. 4. Outer sheath diameter: 1.8 mm 5. Working length: 240 cm 	10	PIECE
1.008	<p>Esophageal Varices Ligator band Specifications</p> <ol style="list-style-type: none"> 1. Should contain a barrel with 6 bands 2. Bands should be totally latex free or made of natural rubber latex for prevention of allergy 3. Bands should be prepared through machine moulding to reduce the risk of post banding rebleed. 4. Should have a clear band to signal when only one band remains to be deployed 	250	PIECE
1.009	<p>Esophageal Varices Ligator (With Handle) Specifications</p> <p>Multiband ligator set for endoscopic variceal ligation, set should include</p> <ol style="list-style-type: none"> 1. Barrel with 6 bands 2. Bands should be totally latex free or made of natural rubber latex for prevention of allergy and should be prepared through machine moulding. 3. Should have a clear band to signal when only one band remains to be deployed 4. Trigger cord and loading catheter should be provided. 5. Precision control handle and irrigation adapter compatible for use with 2.8 mm accessory channel 5. Rotation of 90 degree of the handle should result in deployment of the band 	30	PIECE

1.010	<p>Glue for gastric variceal hemostasis Specifications</p> <ol style="list-style-type: none"> 1. Should contain N butyl Cyanoacrylate glue. 2. Should contain 1 ml per ampoule. 	75	AMPOULE
1.011	<p>Polidoconol Injection Specifications:</p> <ol style="list-style-type: none"> a. 2 ml vial containing 3% polidoconol for endoscopic sclerotherapy of esophageal varices. 	15	AMPOULE
1.012	<p>Endoscopic bipolar electrohemostasis catheter Specifications:</p> <ol style="list-style-type: none"> 1. Should have a rounded gold distal tip 2. Catheter shaft should be 7 Fr, firm to reduce kinking and allow en face as well as tangential tamponade. 3. Probe should be combined with 25 G sclerotherapy needle 4. Should have an integrated channel for manual or mechanical irrigation 5. Working length should be 210 cm 6. Compatible with minimum working channel of 2.8 mm 	5	PIECE
1.013	<p>Endoscopic bipolar electrohemostasis catheter Specifications:</p> <ol style="list-style-type: none"> 1. Should have a rounded gold distal tip 2. Catheter shaft should be 7 Fr, firm to reduce kinking and allow en face as well as tangential tamponade. 3. Should have an integrated channel for manual or mechanical irrigation 5. Working length should be 300 cm 6. Compatible with minimum working channel of 2.8 mm 	5	PIECE
1.014	<p>Endoscopic Bipolar Coagulation Forceps Specifications:</p> <ol style="list-style-type: none"> a. Should have less impact of deeper tissue thermal degeneration b. Enables long-term prevention of re-bleeding c. Based on its flexibility, it can be used for different applications such as EMR/ESD, gastrointestinal bleeding and papilla resection d. Should have minimum length of 2200 mm and capable of using in 2.8 mm working channel 	5	PIECE
1.015	<p>Endoscopic Electrosurgical haemostasis forceps: Specifications</p> <ol style="list-style-type: none"> 1. Should have a precise rotatable design for precise approach & targeting 2. The forceps cup should have a anti slip construction to grasp bleeding point securely for faster and reliable hemostasis <ol style="list-style-type: none"> a. Upper GI [Tapered tip for precise approach] <ol style="list-style-type: none"> 1. Working length: 1650 mm Cup size: 5 mm b. Lower GI [smaller distal cup to suit colon's thin wal] <ol style="list-style-type: none"> 1. Working length: 2300 mm Cup size: 4 mm 	4	PIECE

1.016	<p>Endoscopic Hemostatic clips Specifications:</p> <ol style="list-style-type: none"> 1. Should have a 360° bidirectional rotation for precise clip placement. 2. Should have a reopening capability up to multiple times. 3. Catheter size: 7 Fr, Length: 230 cm 4. Should have a distinct anchoring tips and Nitinol reinforced jaws. 5. Widest span: 16 mm. 	30	PIECE
1.017	<p>Individual single use clips Specifications:</p> <ol style="list-style-type: none"> 1. Standard clip arm length with 135-degree exterior jaws angle 2. Should be sterilized, disposable and come in a cartridge for loading. 3. Should be compatible for loading on to a rotatable reusable applicator 	40	PIECE
1.018	<p>Hemospray for non-variceal bleed management Specifications:</p> <ol style="list-style-type: none"> 1. Should contain a natural mineral powder 2. Should contain a preassembled device handle containing the CO2 cartridge. 3. Should have a knob for activation and deactivation of the CO2 cartridge. 4. Should contain a valve which when opened makes it ready for deployment of hemostatic powder. 5. Should have a trigger which upon depressed for 1-2 seconds deploys the powder. <p>A - 7 Fr *230cm catheter B - 10 Fr *230cm catheter</p>	5	PIECE
1.019	<p>Metallic guide wire for use in endoscopic dilation Specifications</p> <ol style="list-style-type: none"> 1. Guidewire should be made of stainless steel 2. Should have coiled spring distal tip with the tip shape progressively flexible. 3. Wire guide length should be 200 cm or more with markings on the guidewire to assess the depth of insertion. 	1	PIECE
1.020	<p>Controlled radial expansion balloon for endoscopic dilation a. 6-8 mm b. 8-10 mm c. 10-12 mm d. 12-15 mm e. 15-18 mm Specifications</p> <ol style="list-style-type: none"> 1. Multi diameter balloon - three distinct sizes at three distinct pressure 2. Balloon made of Pebax, a non-compliant material 3. Should have a rectilinear shoulder of balloon for endoscopic visualization and evaluation during procedure 4. Balloon length of 5.5 cm, 240 cm long catheter compatible with 2.8 mm working channel. 	40	PIECE

1.021	<p>Percutaneous endoscopic gastrostomy set with feeding tubes sizes, a. 20 Fr. b. 24 Fr. Specifications 1. Percutaneous endoscopic gastrostomy set for pull technique 2. Should contain Pivotal bolster, insertion wire, cold snare, universal adapter, safety sharps and miscellaneous accessories for peg tube placement.</p>	15	PIECE
1.022	<p>PEG replacement Tubes Specifications 1. Feeding tube size 18 Fr. & 24 Fr. 2. Length 24.6 cm 3. Balloon capacity 20 cc</p>	10	PIECE
1.023	<p>Foreign body retrieval basket 1. Should have a hard wire basket for removal of foreign bodies 2. Should have a rotatable pin vise 360 degree 3. Should have a shape memory 4. Sheath size should be 7 Fr and working length 200cm 5. Should have a dual lumen A. Size 2x4 cm B. Size 3x6 cm</p>	3	PIECE
1.024	<p>Foreign body retrieval forceps Specifications 1. Jaw opening diameter should be more than 7 mm 2. Sheath diameter should be 2.3 mm and working length of 230 cm 3. The forceps should have an alligator jaw with a rat tooth design for good grasping of foreign bodies in the gut.</p>	2	PIECE
1.025	<p>Foreign body/ Polyp retrieval net Specifications: a. Should be flexible and a durable basket b. Should have a octagon design to enable spring like action c. Designed to open maximum capacity and have memory points to maintaining its shape even after several passes d. Should be made of transparent netting material to improve visibility e. Net size: A - 4x5.5 cm & 230 cms B - 2.5x4.5 cm & 230 cms C - 2x4.5 cm & 230 cms</p>	5	PIECE
1.026	<p>Foreign body retrieval prongs Specifications 1. Should be 360 degrees rotatable 2. Should have an opening diameter of 35 mm 3. Sheath diameter should be 2.3 mm and working sheath length of 230 cm.</p>	2	PIECE

1.027	<p>Rubber tipped grasping forceps for sharp foreign bodies: Specifications</p> <ol style="list-style-type: none"> 1. Should have a working length of 190 cm and an opening width of 4.8 mm 2. Must be compatible with channel size - 2 mm 3. Should be reusable and autoclavable 4. The rubber at the tip should be latex free and must grasp sharp objects for easy removal. 	2	PIECE
1.028	<p>Overtubes for extraction of upper gastrointestinal foreign bodies Specifications</p> <ol style="list-style-type: none"> 1. Compatible with scope diameter 9.8 to 11.7 mm 2. Tapered inner tube 3. Reinforced outer tube with detachable insufflations cap 4. Gastric length of 50 cm. 	1	PIECE
1.029	<p>Overtubes for extraction of upper gastrointestinal foreign bodies Specifications</p> <ol style="list-style-type: none"> 1. Compatible with scope diameter 8.6 to 10 mm 2. Tapered inner tube 3. Reinforced outer tube with detachable insufflations cap 4. Gastric length of 50 cm. 	1	PIECE
1.030	<p>Inflation syringe for endoscopic controlled radial expansion balloon dilators Specifications</p> <ol style="list-style-type: none"> 1. Should possess a syringe with the pressure gauge for use with endoscopic balloon dilators 2. Should be compatible for attachment with the inflation handle. 	5	PIECE
1.031	<p>Inflation handle for endoscopic controlled radial expansion balloon dilators Specifications</p> <ol style="list-style-type: none"> 1. Should possess a handle with a provision for attaching the inflation syringe with the pressure gauge. 2. Should contain a knob which can be adjusted forwards and backwards for activation of the inflation syringe. 	2	PIECE
1.032	<p>Single use Distal attachment Cap Specifications</p> <ol style="list-style-type: none"> 1. Should be compatible with olympus endoscopes, colonoscopes & enteroscopes 2. Should maintain the appropriate distance between endoscope Tip and Target lesion 3. The edge of distal attachment should have a round shape to reduce damage to tissue. 4. Should have a 3 mm side hole on the sleeve that ensures any fluid on the inside to be drained easily. 5. Distal end from scope after attachment should be 4mm <p>A. Outer diameter of the cap: compatible for use with olympus GF - H170 gastroscopes B. Outer diameter of the cap: compatible for use with olympus GF - H170 colonoscopes C. Outer diameter of the cap: compatible for use with olympus GF - H170 enteroscopes</p>	3	PIECE

1.033	<p>Reusable black cap for Distal attachment to endoscope Specifications</p> <ol style="list-style-type: none"> 1. Should be compatible with olympus magnification endoscopes with near focus 2. Should be autoclavable 3. Should be black in colour to prevent halation. 4. Should have an outer diameter of 11.45 mm for enabling good fit to the tip of the magnification endoscopes. 5. Should maintain the appropriate distance between endoscope Tip and Target lesion during electronic chromoendoscopy 6. The edge of distal attachment should have a round shape to reduce damage to tissue. 7. Should have a side hole on the sleeve that ensures any fluid on the inside to be drained easily. 	1	PIECE
1.034	<p>Reusable spray catheter for chromoendoscopy Specifications</p> <ol style="list-style-type: none"> 1. Should have a screw type catheter with a stylet. 2. Should be sterile and autoclavable 3. Should be kink resistant for smooth insertion. 4. Should have a sprinkler mechanism for wider and evenly distributed spraying for clear visualization. <p>A - Working length of 1650 mm compatible with 2.8 mm channel. B- Working length of 2400 mm compatible with 2.8 mm channel</p>	3	PIECE
1.035	<p>Detachable snare Specifications</p> <ol style="list-style-type: none"> 1. Opening width: 30 mm 2. Sterile and disposable 3. Should have Loop materials with firm grip 4. Compatible with applicator in item 36 	10	PIECE
1.036	<p>Detachable snare applicator</p> <ol style="list-style-type: none"> 1. Sterile and disposable. 2. Working length 2300 mm 3. Working channel: 2.8 mm 4. Ligation pressure adjustable by controlling handle. 5. Compatible for snares in item 35 	1	PIECE
1.037	<p>Polypectomy snare for use in enteroscopy Specifications</p> <ol style="list-style-type: none"> 1. Should have wide opening oval Snare 2. Should have thick braided wires for maximum coagulation and control 3. Should possess an integrated handle with measurement marking 4. Wire Thickness should be 0.47 mm and loop diameter should be 25 mm 5. Should be suitable for 2.8 mm channel 6. Working length should be 3000 mm and reusable 	3	PIECE

1.038	<p>Polypectomy snare Specifications</p> <ol style="list-style-type: none"> 1. Should have wide opening oval snare 2. Should have thick braided wires for maximum coagulation and control 3. Should possess an integrated handle with measurement marking 4. Wire Thickness should be 0.47 mm and loop diameter should be 25 mm 5. Should be suitable for 2.8 mm channel 6. Working length should be 2300 mm and intended for single use 	5	PIECE
1.039	<p>Polypectomy Snare</p> <ol style="list-style-type: none"> a. Loop width: 27 mm b. Loop width: 30 mm <p>Specifications</p> <ol style="list-style-type: none"> 1. Should possess a short throw three ring handle 2. Sheath diameter should be 2.4 mm 3. Working length should be 240 cm 4. Should have a flexible braided wire loop 5. The inner diameter of the sheath should have lubricant coated to minimize friction during extension and retraction of the snare. 	10	PIECE
1.040	<p>Rotatable polypectomy snare Specifications</p> <ol style="list-style-type: none"> 1. Loop width 20 mm 2. Working length 195 Cm 3. Mid catheter rotation activator - Gives snare loop rotation for proper positioning 4. It should have star shaped inner catheter design for reduced friction 5. Should possess the torque technology which provides smooth, predictable rotation even when scope is in torque position 	5	PIECE
1.041	<p>Nasojejunal Tube Specifications</p> <ol style="list-style-type: none"> 1. Should be radio-opaque, tissue compatible, and made of polyurethane 2. Should be 14 Fr and 120 cm long 3. Should have a hollow flexible guidewire and nasal fixation tape 	100	PIECE

1.042	<p>Over the scope clip system: Specifications: a. Must contain an applicator cap with a mounted clip, a thread and handwheel for clip release b. Caps should be available in different sizes (11, 12 & 14 mm) & different depths (3 or 6 mm) for capturing different amount of tissue. c. The clips should be available in different sizes adapted to cap size & different shapes of teeth for different indications like compression effect, anchoring and closure of gastric lumen. d. Should provide two instruments for grasping or anchoring and help in approximation of indurated tissue 1. 8.5 - 11 mm cap, 6 mm depth and teeth with small spikes 2. 10.5 - 12 mm cap, 6 mm depth and teeth with small spikes 3. 11.5 - 14 mm cap, 6 mm depth and teeth with small spikes</p>	12	PIECE
1.043	<p>Over the scope clip aids: Specifications: a. Must contain an anchor for better approximation of tissue especially when indurated especially for thin tissue. b. Should be applicable in fistulae and chronic ulcers. c. Should provide approximation of gaping edges of the lesion. A - Tissue anchor; B - Tissue grasper</p>	2	PIECE
1.044	<p>Polyethylene Glycol for bowel preparation for colonoscopy Specifications Each pkt weighs 137.15 g 1. Potassium chloride 2. Sodium bicarbonate 3. Polyethylene glycol (118 g) 4. Sodium chloride 5. Sodium sulphate</p>	200	PIECE
1.045	<p>Single use splinting tube for use in single balloon enteroscope <ul style="list-style-type: none"> • Should be made up of latex-free hydrophilic coated silicone • Should have radiopaque materials in the distal end of the splinting tube's tip for effective positioning confirmation under fluoroscopy • Must be Sterile and Disposable • Outer Diameter : 13.2 mm. Inner Diameter : 11.0 mm • Working Length : 1320mm • Comatible for use along with enteroendoscope and balloon control unit. </p>	8	PIECE

1.046	<p>ERCP Sphincterotome for use with enteroscope Specifications</p> <ol style="list-style-type: none"> 1. Should have taper tip with the catheter size of 6-5 Fr 2. Length of the catheter should be 320 cm 3. Should allow 0.035"guidewire and compatible with 2.8 mm channel 4. Cutting wire should be 25 mm in length with proximal insulated tip for protection 	3	PIECE
1.047	<p>Triple lumen Needle knife Specifications</p> <ol style="list-style-type: none"> 1. 5 Fr, 320 cm length catheter 2. 3-5 mm cutting needle compatible with 0.035 inch guidewire 3. Needle control lock in handle 	2	PIECE
1.048	<p>Endoscopic Biliary Dilatation Balloon for enteroscope assisted ERCP Specifications</p> <ol style="list-style-type: none"> 1. Should be a Hydrophilic dilatation balloon 2. Catheter size should be 6.8 - 4.5 Fr. 3. Length of the balloon catheter should be 320 cm 4. Should allow 0.035"guidewire and compatible with 2.8 mm channel <p>A. 6 mm diameter balloon B. 8 mm diameter balloon C. 10 mm diameter balloon</p>	6	PIECE
1.049	<p>Stone extraction Balloon for enteroscope assisted ERCP</p> <ol style="list-style-type: none"> 1. 3 inflation diameters in one device to meet specific clinical needs 2. Ease of injection with contrast exiting the distal tip 3. Radiopaque bands enhance fluoroscopic visualization of the extraction balloon 4. Reference marks on syringe indicate each inflation diameter 5. SIZES: a. 8.5//12/15 6. Length of the catheter should be 275 cm or more 	3	PIECE
1.050	<p>ERCP high performance guide wire for enteroscope assisted ERCP Specifications:</p> <ol style="list-style-type: none"> 1. Should have a 25 cm hydrophilic end 2. Radio-opaque coil spring for fluoroscopic visibility 3. Nitinol core for kink resistance and shape retention 4. Length of the wire should be 600 cm 5. Should be 0.035" and compatible with 2.8 mm channel 	3	PIECE
1.051	<p>ERCP cannula for enteroscope assisted ERCP Specifications</p> <ol style="list-style-type: none"> 1. Should have a pre curved type tapered tip catheter of size 5.5 - 4.5 Fr 2. Should have a double lumen and a length of 320 cm 3. Should allow 0.035"guidewire and compatible with 2.8 mm channel 4. Should have radiopaque distal tip markings 	3	PIECE

1.052	<p>Biliary Dilation catheter, bougie type for enteroscope assisted ERCP</p> <p>Specifications</p> <ol style="list-style-type: none"> 1. The length of the catheter should be 275 cm or more. 2. 7 Fr catheter with radiopaque tapered tip to 4 Fr 3. Length of tapering part should be 3 cm 4. Should allow 0.035" guidewire and compatible with 2.8 mm channel 	2	PIECE
1.053	<p>One action Stent introduction system for enteroscope assisted ERCP</p> <p>Specifications</p> <ol style="list-style-type: none"> 1. Should have a 10 Fr pushing catheter 2. Should have a 6 Fr guiding catheter 3. Should contain four radiopaque bands 5 cm apart 4. Length of the system should be 320 cm 5. Should allow 0.035" guidewire and compatible with 2.8 mm channel 	2	PIECE
1.054	<p>Sclerotherapy Needles for enteroscopy use</p> <p>Specifications</p> <ol style="list-style-type: none"> 1. Should be compatible with 2.8 mm channel 2. Catheter size should be 7 Fr. 3. Catheter length should be 320 cm. 4. Maximum needle extension possible should be 4 mm <p>A - 23 G B - 25 G</p>	3	PIECE
1.055	<p>Small bowel capsule</p> <p>Specifications:</p> <ol style="list-style-type: none"> 1. Should have dimensions of 10 x 25 mm and frame rate of more than 5 fps 2. Operating time of 10 hours or more. 3. Field of view should be more than 155 degrees. 4. Transmission mode should be radiofrequency & image sensor should be preferably complementary metal oxide silicon. 5. Recorder should have real time viewing option. 6. Should be FDA approved. 7. Should provide the sensor pads belt, recorder and software workstation for analysis and interpretation of images. 	15	PIECE
1.056	<p>ERCP cannula</p> <p>a. 5-4-3</p> <ol style="list-style-type: none"> 1. Distal tip should be 5 Fr 2. Should have an OD of 1.7 mm compatible with 0.018 inch guidewire <p>b. Tapered tip</p> <ol style="list-style-type: none"> 1. Distal tip should be 5.5 to 4.5 Fr 2. Catheter length should be 200 cm and compatible with 0.035 guidewire 3. Should possess two lumen 4. Should have radiopaque distal tip markings 	10	PIECE

1.057	<p>Dedicated sphincterotome for minor papilla cannulation Specifications</p> <ol style="list-style-type: none"> 1. Should be a 6 Fr Sphincterotome tapered to 3 Fr at the tip of length 6 mm 2. Should have a braided cutting wire of length 15 mm 3. Should contain over the guidewire injection port 4. Should be compatible with 0.025" guidewire 	2	PIECE
1.058	<p>Triple lumen sphincterotome Specifications</p> <ol style="list-style-type: none"> 1. Sphincterotome should contain a precurved distal tip for easier positioning of cutting wire. 2. Distal tip length should be 3 mm and tapered to 4.5 Fr 3. Cutting wire should be 20 mm in length 4. Should contain separate lumens for guide wire, cutting wire and contrast injection 5. Unique safety coating at proximal half of cutting wire and endoscopic markers assist with positioning the cutting wire 6. Should contain a friction handle to ensure maintaining the bowed position and stable cutting wire orientation 7. Should contain radiopaque tip markings for optimal visibility under fluoroscopy 8. Should contain working length of 170 cms and compatible with 0.035 inch guidewire 	70	PIECE
1.059	<p>ERCP hydrophilic guidewire Specifications</p> <ol style="list-style-type: none"> a. Angled tip (0.025") b. Straight tip (0.025") <ol style="list-style-type: none"> 1. Should have a 7 cm highly flexible hydrophilic tip for excellent passability 2. Should contain special fluorine coatings for effortless manipulation 3. Should possess a wider core design of the wire giving enhanced one-to-one torqueability 4. Exceptional endoscopic and fluoroscopic visibility 5. Should have provision for measurement over the wire 6. Should contain working length of 450 cm 	35	PIECE
1.060	<p>ERCP standard straight guidewire Specifications</p> <ol style="list-style-type: none"> 1. Should have a hydrophilic tip of 5 cm. 2. Shaft should be made of nitinol to prevent kinking and enable good wire manipulation and control. 2. Should have a coating to promote smooth tracking and tactile sensation of the wire during exchange 4. Should contain working length of 450 cm <ol style="list-style-type: none"> A. .035 inch and B. 0.025 inch 	40	PIECE

1.061	<p>ERCP high performance guide wire Specifications: 1. Should have an atraumatic flexible tip design. 2. Should have a calibrated spring coil tip for gentle contact to avoid complications. 3. Spring coil should be made up of platinum for 4 cm for increasing fluoroscopic visibility 3. Should have a hydrophilic coating for 5 cm or more to reduce friction and improves negotiation through challenging anatomy 4. Nitinol core for kink resistance, shape retention and enhanced pushability 5. Should provide one to one torqueability for precise wire rotation to enable selective duct cannulation and duct navigation. 6. Length should be 480 cm A. 0.035" guidewire with straight tip B. 0.035" guidewire with angled tip C. 0.025" guidewire with straight tip D. 0.025" guidewire with angled tip</p>	40	PIECE
1.062	<p>Ultra thin Guidewire Specifications 1. 018 Guidewire for trauma less easy passage 2. 480 cm catheter length 3. Tip with gold coating for better visibility. 4. Should be made of Triton alloy core for superior pushability & rail support. 5. Straight shapable tip for selective cannulation.</p>	2	PIECE
1.063	<p>Biliary cytology brush Specifications 1. Working length should be 1900 mm 2. Should have a brush diameter of 3 mm 3. Brush length should be 10 mm or more 4. Should have a distal wire guided design 5. Should have a locking cap on the control section to allow the brush to be locked.</p>	10	PIECE
1.064	<p>Biliary plastic stents (Straight) 1. 5 Fr (a-5cm, b-7cm, c-9cm, d-12 cm) 2. 7 Fr (a-5cm, b-7cm, c-10cm, d-12cm, e-15 cm) 3. 8.5 Fr (a-10cm, b-12cm, c-14 cm) 4. 10 Fr (a-7cm, b-10cm, c-12cm, d-14 cm) Specifications 1. Minimizes migration with proximal and distal flaps. 2. Smooth cannulation facilitated by tapered tip. 3. Adapts to bile duct due to gentle curved shape. 4. Pushability and softness delivered by polyethylene material.</p>	160	PIECE

1.065	<p>Biliary plastic stents (Double pig tail) 1. 5 Fr (a-5 cm, b-7 cm) 2. 7 Fr (a-5 cm, b-7 cm, c-10 cm, d-12 cm) 3.10 Fr (a-5 cm, b-7 cm, c-10 cm, d-12 cm)</p> <p>Specifications 1. Minimizes migration with double pigtail configuration 2. Smooth cannulation facilitated by tapered tip 3. Pushability and softness delivered by polyethylene material</p>	120	PIECE
1.066	<p>Pancreatic plastic stents (Single pig tail) 1. 5 Fr (a-3cm, b-5cm, c-7cm, d-9cm) 2. 7 Fr (a-5cm, b-7cm, c-9cm, d-11cm) 3. 8.5 Fr (a-5cm, b-7cm, c-9cm)</p> <p>Specifications 1. Minimize migration with duodenal pigtail (and optional ductal flap) 2. Drainage enhanced by spiral side holes on select stents 3. Smooth cannulation facilitated by tapered tip 4. Accurate positioning enhanced by radiopaque markers on select stents 5. Pushability and softness delivered by polyethylene material</p>	55	PIECE
1.067	<p>Biliary Stone Extraction Balloon a. 9-12 mm b. 12-15 mm</p> <p>Specifications 1. "Squared shoulder" balloon is designed to facilitate stone removal. 2. Triple-lumen Retrieval Balloon 3. Catheter enables front or back loading of guidewire and offers injection-above or injection-below options to support physician preference and facilitate procedural needs.</p>	25	PIECE
1.068	<p>Endoscopic Stone Retrieval Basket a.2 cm b.3 cm</p> <p>Specifications 1. Can be used for performing mechanical lithotripsy using a lithotripter - Gun 2. Silver tip technology to avoid basket impaction 3. Wire guided, Pre cursor Alloy material</p>	10	PIECE
1.069	<p>One action Stent introduction system</p> <p>Specifications 1. Should have a 10 Fr pushing catheter 2. Should have a 6 Fr. guiding catheter which should be tapered enabling dilation of the stricture prior to stent placement. 3. Four radiopaque bands 5 cm apart</p>	15	PIECE

1.070	<p>Stent pushing Catheter a.8.5 Fr b.7 Fr d.5 Fr Specifications 170 cm pushing catheter Compatible with 0.035 inch guidewire</p>	20	PIECE
1.071	<p>Mechanical Lithotripter for biliary stones Specifications 1. Two strong wire dormia baskets with sheath with open basket diameter of 3 cm. 2. Rotatable basket available in new bullet tip style. 3. Inner and outer sheath can be preassembled to maximize safety and aid cannulation 4. Working length 1950 mm 5. Should be compatible for use with 4.2 mm working channel</p>	5	PIECE
1.072	<p>Endoscopic Biliary Dilatation Balloon a. 4 mm diameter b. 6 mm diameter c. 8 mm diameter d. 10 mm diameter Specifications 1. Should have a hydrophilic dilatation balloon 2. The tip should be 4Fr and the length of balloon should be 4 cm 3. Should have a facility for rapid exchange system with two radioopaque positioning markers</p>	20	PIECE
1.073	<p>Biliary Dilation catheter, bougie type a. 6 Fr b. 7 Fr c. 8.5 Fr d. 10 Fr Specifications 1. 200 cm catheter 2. 6 Fr radiopaque tapered tip and 3. Length of tapering part should be 3 cm</p>	15	PIECE
1.074	<p>Cystotome for Endoscopic Cystogastrostomy Specifications 1. Should have a 10 Fr outer catheter & 5 Fr inner catheter 2. Should contain a proximal & distal electrode 3. 10 Fr diathermic ring size & 0.038" needle knife</p>	2	PIECE
1.075	<p>Cystotome for Endoscopic Cystogastrostomy or EUS guided biliary drainage Specifications 1. Should have a 6 Fr catheter with a diathermic tip 2. Should have a length of 180 cm and allow passage of 0.035" guidewire 3. Should contain one HF connector</p>	3	PIECE

1.076	<p>EUS Fine Needle Biopsy a. 19 G b. 22 G Specifications: 1. Should be made up of Nitinol to retain its shape without any deformation even after multiple passes. This helps ensure the needle follows a predictable and consistent trajectory. 2. Coil sheath with multi layered construction to facilitates excellent force transmission from the handle to the needle tip and Deep tissue can be punctured regardless of scope angle. 3. Echogenic needle tip for enhanced visibility for Controlled Puncture 4. Side hole option to collect the sample within the needle. 5. Sharp Needle Tip: Menghini-type tip with sharp edges supports smooth puncture even at an oblique angle. 6. Adjustable sheath length- Improved Handle Maneuverability for Convenient and controlled handling</p>	20	PIECE
1.077	<p>EUS Fine Needle Biopsy a. 19 G Specifications: 1. Should be made up of Nitinol to retain its shape without any deformation even after multiple passes. This helps ensure the needle follows a predictable and consistent trajectory. 2. Coil sheath with multi layered construction to facilitates excellent force transmission from the handle to the needle tip and Deep tissue can be punctured regardless of scope angle. 3. Echogenic needle tip for enhanced visibility for Controlled Puncture 4. Sharp Needle Tip: Menghini-type tip with sharp edges supports smooth puncture even at an oblique angle. 5. Adjustable sheath length- Improved Handle Maneuverability for Convenient and controlled handling</p>	3	PIECE
1.078	<p>EUS Fine Needle Biopsy a. 25 G Specifications: 1. Should possess core trap to receive tissue into the needle. 2. Should have a reverse bevel for shearing material from lesion during retrograde needle movement. 3. High definition needle for ultrasound visibility. 4. Should contain safety lock ring, integrated sheath adjuster 5. Sheath size 4.8 - 5.2 Fr & needle extension: 0-8 cm 6. Metal Luer lock for secure anchoring to the echoendoscope</p>	15	PIECE

1.079	<p>EUS Fine Needle Biopsy : Needle size: 20 G Specifications: 1. Should possess core trap to receive tissue into the needle. 2. Should have a forward bevel for shearing material from lesion during needle movement. 3. High definition needle for ultrasound visibility. 4. Should contain safety lock ring, integrated sheath adjuster 5. Should have a coiled sheath size 4.8 - 5.2 Fr & needle extension: 0-8 cm 6. Metal Luer lock for secure anchoring to the echoendoscope 7. Needle should be made up of stainless steel with a menghini tip</p>	25	PIECE
1.080	<p>EUS Fine Needle for cytology Specifications: Needle size: 22 G 1. Should possess bevel tip to receive tissue into the needle. 2. Stylet should be 2 mm ball type. 3. High definition needle for ultrasound visibility with a flexible coiled sheath 4. Should contain safety lock ring, integrated sheath adjuster 5. Sheath size 4.8 - 5.2 Fr & needle extension: 0-8 cm 6. Metal Luer lock for secure anchoring to the echoendoscope. 7. Needle should be made of stainless steel.</p>	25	PIECE
1.081	<p>EUS Needle for EUS guided drainage procedures: Needle size: 19 G Specifications: 1. Should possess bevel stylet tip to easy punch needle. 2. Atraumatic tip designed to allow 0.35 wire guide 3. High definition needle for ultrasound visibility with coiled sheath 4. Should contain safety lock ring, integrated sheath adjuster 5. Sheath size 4.8 - 5.2 Fr & needle extension: 0-8 cm 6. Metal Luer lock for secure anchoring to the echoendoscope</p>	10	PIECE

1.082	<p>EUS flexible FNB needle a. 22 G b. 25 G Specifications: 1. Should have three pronged tip to provide stability at puncture and have fully formed heels designed to maximize tissue capture and minimize fragmentation 2. Cobalt chromium needle assembly 3. Should have superior needle penetration, pushability, durability and kink resistant. 4. The three points provide stability at puncture while the high quality, fully formed heels are designed to maximize tissue capture and minimize fragmentation</p>	50	PIECE
1.083	<p>Lumen apposing metal stent for pseudocyst drainage Specifications 1. Should be made of Nitinol 2. Should have flexible and fine mesh tubular prosthesis 3. Should have radiopaque markers, three at both ends and two in the middle. 4. Upon deployment the body of the stent should impart an outward radial force while both heads impart a horizontal inward force, allowing transmural drainage. 5. Should be fully silicone covered facilitating removal. 6. Should have a retrieval String for repositioning or easy removal. A - 12 mm B - 16 mm diameter</p>	4	PIECE
1.084	<p>Lumen apposing SEMS Specifications: a. Should have an electrocautery enhanced Stent delivery system, one step diathermic ring & cut wire for easy access into target tissue b. Lumen apposing SEMS & perpendicular flanges for securing tissue layers and prevent migration c. Saddle length: 8 mm & 10 mm d. Stent diameter A - 8 mm, B - 10 mm C - 15 mm, D - 20 mm e. Compatible with 3.7 mm or more working channel of EUS scope</p>	2	PIECE
1.085	<p>Sterile water for irrigation of manometry catheter Specifications: 1. Should be apyrogenic 2. Should be compatible for irrigation of water perfused channels of the manometry system 3. Should be available as 3 L containers</p>	10	PIECE
1.086	<p>Probes for pH and Impedance System Specifications: a. Single channel pH probe. 1. Should be compatible with the Zephyr pH & impedance recorder</p>	50	PIECE
1.087	<p>Probes for pH and Impedance System Specifications: a. 6 channel impedance & 2 channel pH probe 1. Should be compatible with the Zephyr pH & impedance recorder</p>	25	PIECE

1.088	<p>Buffer solutions for PH and Impedance testing. Specifications:</p> <ol style="list-style-type: none"> 1. Should contain buffer solution with pH of 4 ± 0.05 which is stable at 25-degree C 2. Should contain 250 ml of the solution per bottle for calibration and be compatible with the Zephyr pH & impedance recorder 	2	PIECE
1.089	<p>Buffer solutions for PH and Impedance testing. Specifications:</p> <ol style="list-style-type: none"> 1. Should contain buffer solution with pH of $7 + 0.05$ which is stable at 25-degree C 2. Should contain 250 ml of the solution per bottle for calibration and be compatible with the Zephyr pH & impedance recorder 	2	PIECE
1.090	<p>Argon plasma coagulation (APC) probe: Specifications:</p> <ol style="list-style-type: none"> 1. Should be a flexible probe. 2. Should have an integrated membrabe filter and connecting cable so that it will be able to Plug and Play 3. Should be compatible with the existing APC generator and electrosurgical unit of the department. No preparation of the probe or cable should be required <ol style="list-style-type: none"> 1. Axial openings for the axial plasma beam Dimensions: 2.3 mm Length: 2.2 m 2. Lateral openings for the side fire conical plasma beam Dimensions: 2.3 mm Length: 2.2 m 3. Circular openings for the circumferential plasma beam Dimensions: 2.3 mm Length: 2.2 m 4. Axial openings for the axial plasma beam Dimensions: 1.5 mm Length: 3 m 5. Axial openings for the axial plasma beam Dimensions: 3.2 mm Length: 2.2 m 	12	PIECE
1.091	<p>Neutral electrode safety system foot plate for electrocautery Specifications</p> <ol style="list-style-type: none"> 1. Disposable foot plate 2. Contact area 168 cm² 3. 3 m connecting cord 4. Compatible with ERBE VIO system 	200	PIECE
1.092	<p>Endoscopic Bite block Specifications</p> <ol style="list-style-type: none"> 1. Large lateral openings to enhance digital guidance when advancing the endoscope and provides easy access for additional suctioning 2. Should be compatible with 54 Fr mechanical dilator 3. Soft, adjustable, latex-free strap enhances patient comfort 4. Must be self-retaining using a strap 	200	PIECE

1.093	<p>Endoscope tip protector: Specifications:</p> <ol style="list-style-type: none"> 1. Made of Foam 2. Inner diameter: 7-18 mm 3. Outer diameter: 4 cm 4. Length: 14 cm 	25	PIECE
1.094	<p>Disposable biopsy valves Specifications:</p> <ol style="list-style-type: none"> 1. Cover the openings to the biopsy channels. 2. Maintains scope pressure while passing devices. 3. Latex free. 4. Compatible with olympus endoscopes. 	25	PIECE
1.095	<p>Single use instrument channel adaptor for use with water irrigation pump Specifications:</p> <ol style="list-style-type: none"> 1. Should contain an adaptor for dual function of water irrigation and passage of an accessory through the instrument working channel. 2. Should be compatible with instrument channel water tube and the working channel of olympus endoscope. 	10	PIECE
1.096	<p>Single use instrument channel water tube for use with water irrigation pump Specifications:</p> <ol style="list-style-type: none"> 1. Should contain a water tube for attachment to the instrument channel adaptor. 2. Should be compatible for use with olympus endoscope and water irrigation pump. 	20	PIECE
1.097	<p>Auxillary water tube for use with water irrigation pump Specifications:</p> <ol style="list-style-type: none"> 1. Should contain a reusable water tube for attachment to the auxillary water irrigation channel of olympus endoscopes. 2. Should be compatible for use with olympus water irrigation pump. 	10	PIECE
1.098	<p>Single use endoscopy channel cleaning brushes Specifications:</p> <ol style="list-style-type: none"> 1. Should contain two channel cleaning brush on either end of a single sheath. 2. Sheath should be flexible and kink resistant and enable brushing through the channel with minimum effort. 3. Bristles should contain optimum thickness and stiffness to ensure full contact with the working channel during cleaning. 4. Should be compatible with olympus endoscopes with working channel of diameter 2 - 4 .2 mm. 5. Working length should be 220 cm or more 	50	PIECE

1.099	<p>Single use endoscopy channel opening cleaning brushes Specifications:</p> <ol style="list-style-type: none"> 1. Should contain channel opening cleaning brush. 2. Sheath should be flexible and kink resistant. 3. Bristles should contain optimum thickness and stiffness during cleaning. 4. Should be compatible with olympus endoscopes with working channel of diameter 1.2 - 6 mm. 	25	PIECE
1.100	<p>Disposable Endoscopy Cleaning brushes</p> <ol style="list-style-type: none"> 1. Able to clean accessory channels and endoscope valves, valve housings and connectors. 2. Rigid yet flexible channel brush shaft to allow for ease of passage. 3. Covered brush tip to protect endoscope channel. 4. Ergonomically designed double ended valve cleaning brush to maximise cleaning. 5. Brush size – 6.5-5 mm 6. Valve brush size - 10, 6-4.5mm 7. Brush length - 240 cm 	25	PIECE
1.101	<p>Xenon lamp for light source of video endoscope processor Specifications:</p> <p>A. 100 W -</p> <ol style="list-style-type: none"> 1. Should have 400 hours of life 2. Compatible with light source of Pentax endoscope processor. 	1	PIECE
1.102	<p>Xenon lamp for light source of video endoscope processor Specifications:</p> <p>A. 300 W xenon arc lamp - 1 no</p> <ol style="list-style-type: none"> 1. Should have 500 hours of life 2. Compatible with light source of Pentax endoscope processor. 	1	PIECE
1.103	<p>C-arm cover Specifications:</p> <ol style="list-style-type: none"> a. Disposable sterile cover for fluoroscopy image intensifier. b. Should be available in different sizes to fit the image intensifier. c. Should have an elastic band for easy retention. 	25	PIECE
1.104	<p>Gastrolaryngeal tube:</p> <ol style="list-style-type: none"> 1. Should have a Bite block with an endoscopy lumen compatible with duodenoscopes 13.8 mm outer diameter. 2. Should contain a proximal cuff for stabilizing the tube and blocks oro and nasopharynx and distal cuff that blocks esophageal inlet and avoids gastric ventilation. 3. Inflation line for inflation and deflation of cuffs. 4. Ventilation lumen with 15 mm connector. 5. Ventilation holes should lie in front of the larynx 	3	PIECE

1.105	<p>Balloon tube tamponade for variceal bleed Specifications: a. Should have a separate oesophageal & gastric balloon b. Must have a port for continuous suction of gastric contents, separate ports for inflating gastric & oesophageal balloon c. Sizes: A - 14 Fr; B - 16 Fr; C - 18 Fr</p>	6	PIECE
1.106	<p>Rapid urease test kit for Helicobacter pylori detection in endoscopic biopsies Specifications: a. Easy to read colour change for test positivity with superior accuracy b. Results are read within 15 mins. c. Should not require refrigeration d. Should have a long shelf life of atleast 2 years.</p>	50	PIECE
1.107	<p>Glucose test strips for blood sugar estimation – SPECIFICATIONS: 1. Compatible glucometer should be supplied free of cost to all wards, services & OT where it is required. 2. Free repair/replacement should be provided during the period of use.</p>	3000	PIECE
1.108	<p>Sealing Machine</p>	2	PIECE
1.109	<p>Partially Covered Oesophageal Metal Stent a. 10 cm b. 12 cm c. 15 cm Specifications 1. Fluoroless Stent Placement using the positioning ruler on the shaft of the delivery system. 2. Single-strand Nitinol wire is designed to maintain luminal patency whilst reducing the risk of migration. 3. Proximal and Distal release deployment with different Flair and Body sizes of Stent</p>	15	PIECE
1.110	<p>Fully Covered Removable Oesophageal Metal Stent a. 10 cm b. 12 cm c. 15 cm Specifications 1. 18.5 French coaxial style delivery system with 28 mm Flare and 23 mm Body and Reconstrainability upto 75% 2. Excellent fluoroscopic visibility with 4 RO markers with yellow transition zone and proximal reposition marker on handle. 3. Should have permalume silicone covering for preventing tumour ingrowth</p>	5	PIECE

1.111	<p>Partially Covered Oesophageal Metal Stent with anti-reflux valve a. 8 cm b. 10 cm c. 12 cm</p> <p>Specifications</p> <ol style="list-style-type: none"> 1. Fluoroless Stent Placement using the positioning ruler on the shaft of the delivery system. 2. Single-strand Nitinol wire is designed to maintain luminal patency whilst reducing the risk of migration. 3. Proximal and Distal release deployment with different Flair and Body sizes of Stent 	3	PIECE
1.112	<p>Through The Scope Duodenal Highly flexible Metal Stent – a. 6 cm b. 9 cm c. 12 cm</p> <p>Specifications</p> <ol style="list-style-type: none"> 1. Nitinol Treatment offers Complete mesh visibility under fluoroscopy and Superior radial force 2. Recapturing of the stent and repositioning is possible up to 70% stent deployment 3. Should have looped ends to minimise risk of trauma 4. Large diameter stent and flare design: 27 mm flare and 22 mm body for improved relief of obstruction and reduce the risk of migration. 5. Customised Stent for enteral use & flare design proximal to food flow 	15	PIECE
1.113	<p>Through The Scope Colonic Highly flexible Metal Stent – a. 6 cm b. 9 cm c. 12 cm</p> <p>Specifications</p> <ol style="list-style-type: none"> 1. Nitinol Treatment offers Complete mesh visibility under fluoroscopy and Superior radial force 2. Recapturing of the stent and repositioning is possible up to 70% stent deployment 3. Should have looped ends to minimise risk of trauma 4. Large diameter stent and flare design: 30 mm flare and 25 mm body for improved relief of obstruction and reduce the risk of migration. 5. Customised Stent for Colon & flare design proximal to food flow 	3	PIECE

1.114	<p>Uncovered Biliary Metal Stent – a. 6 cm b. 8 cm c. 10 cm Specifications 1. Should have fixed core, Nitinol wire design for exceptional trackability 2. 8 Fr (2.67mm) low-profile catheter to deploys 30Fr (10mm) lumen stent 3. Should have an elongated, tapered tip to ease passage through difficult strictures 4. Should have an outward radial pressure to gently maintain luminal patency and reduce the need for repeat stenting procedures</p>	20	PIECE
1.115	<p>Fully covered Removable Biliary Metal Stent a. 6 cm b. 8 cm Specifications 1. Should have platinum wire construction for greater flexibility aid placement in tortuous anatomy. 2. Should have removability indication up to 12 months after placement in benign biliary structure 3. Should possess permalume covering 4. Reconstrainable up to 80% of deployment to aid in repositioning.</p>	2	PIECE
1.116	<p>Uncovered Biliary Metal Stent – a. 6 cm b. 8 cm Specifications 1. Fixed core, Nitinol wire designed for exceptional trackability 2. 6 Fr low-profile catheter deploys 30 Fr lumen stent 3. Elongated, tapered tip has been shown to ease passage through difficult strictures 4. Outward radial pressure has been shown to gently maintain luminal patency and may reduce the need for repeat stenting procedures 5. Should enable easier deployment of side by side stent at the same time in hilar obstruction</p>	2	PIECE

1.117	<p>Uncovered Biliary stent Specifications: a. Fixed core, Nitinol wire designed for exceptional trackability b. Laser cut SEMS with a tapered 6 Fr tip stent c. Elongated, tapered tip has been shown to ease passage through difficult strictures d. Outward radial pressure has been shown to gently maintain luminal patency and may reduce the need for repeat stenting procedures e. Should enable easier deployment of side by side stent at the same time in hilar obstruction f. Dual deployment mechanism: Thumb wheel & pull grip g. Stent length should be 10 cm</p>	2	PIECE
1.118	<p>Self expandable removable metallic stent for variceal bleeding: Specifications: a. Must be made of nitinol and durable silicone covering. b. Should have a design to prevent stent migration c. Should have preloaded delivery system 20 - 28 Fr in diameter & 60 cm in length d. Atraumatic stent ends & radio opaque markers at both ends as well as in the middle of stent. e. Should have retrievable loops made of medical grade alloy at both ends facilitating easy stent removal f. Diameter of the stent should be 25 mm in the body and 30 mm at flare with a length of 135 mm</p>	3	PIECE

Samples should be given to storeroom, Department of Medical Gastroenterology within ten days of the closing of online submission of bids failing which the bids will be rejected.

Section-III

GENERAL INSTRUCTIONS TO TENDERERS

1. Period of contract:

The contract shall initially be for a period of **one year** that may be extended to **6 months**. The rates approved shall remain unchanged during the period of contract.

2. Bid Security (Earnest Money Deposit)

- i. In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender.
- ii. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.
- iii. The bidder hereby declares that they accept the condition that if they withdraw or modify their bids during period of validity etc., they will be suspended for a period of two years from participating in any tender invited by JIPMER, Puducherry with effect from the date of their withdrawal or modification of their bid.

3. Eligibility conditions of bidders:

- i. The Tenderer must be a domestic manufacturer that is a 'Class-I local supplier' or a 'Class-II local supplier' as defined under Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as further amended by orders of even number dated 28.05.2018, 29.05.2019, 04.06.2020 and 16.09.2020. In case the manufacturer does not quote directly, they may authorize an agent as per proforma of Manufacturer authorization form as given in the Tender enquiry document to quote and enter into a contractual obligation.
- ii. In compliance with order (Public Procurement No.1) No. 6/18/2019-PPD dated 23rd July 2020 issued by the Public Procurement Division, Dept. of Expenditure, Min of Finance under Rule 144(xi) of GFR 2017 any bidder 'from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority.
 - a. "Bidder" (including the term 'tenderer', 'consultant' or 'service provider' in certain contexts) means any person or firm or company, including any member of a consortium or joint venture (that is an association of several persons, or firms or companies), every artificial juridical person not falling in any of the descriptions of bidders stated hereinbefore, including any agency branch or office controlled by such person, participating in a procurement process.
 - b. "Bidder from a country which shares a land border with India" for the purpose of this Order means: -

- I. An entity incorporated, established or registered in such a country; or
- II. A subsidiary of an entity incorporated, established or registered in such a country; or
- III. An entity substantially controlled through entities incorporated, established or registered in such a country; or
- IV. An entity whose beneficial owner is situated in such a country; or
- V. An Indian (or other) agent of such an entity; or
- VI. A natural person who is a citizen of such a country; or
- VII. A consortium or joint venture where any member of the consortium or joint venture falls under any of the above

c. The beneficial owner for the purpose of (5.2.2) above will be as under:

- I. In case of a company or Limited Liability Partnership, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has a controlling ownership interest or who exercises control through other means.

Explanation—

- "Controlling ownership interest" means ownership of or entitlement to more than twenty-five per cent, of shares or capital or profits of the company;
 - "Control" shall include the right to appoint majority of the directors or to control the management or policy decisions including by virtue of their shareholding or management rights or shareholders agreements or voting agreements;
- II. In case of a partnership firm, the beneficial owner is the natural person(s) who, whether acting alone or together, or through one or more juridical person, has ownership of entitlement to more than fifteen percent of capital or profits of the partnership.
 - III. In case of an unincorporated association or body of individuals, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has ownership of or entitlement to more than fifteen percent of the property or capital or profits of such association or body of individuals.
 - IV. Where no natural person is identified under (a) or (b) or (c) above, the beneficial owner is the relevant natural person who holds the position of senior managing official;
 - V. In case of a trust, the identification of the beneficial owner(s) shall include the identification of the author of the trust, the trustee, the beneficiaries with 15% or more interest in the trust and any other natural person exercising ultimate effective control over the trust through a chain of control or ownership.

d. An Agent is a person employed to do any act for another, or to represent another in dealings with third person.

e. In case of turnkey contracts, the successful bidder shall not be allowed to sub-contract works to any contractor from a country which shares a land border with India unless such contractor is registered with the Competent Authority.

- f. Bidders must certify compliance with the above mentioned provisions in the tender form as per Section VII.
- g. The bidder should have successfully completed the delivery of the item of at least 25% of the quantity required in this tender to a government hospital in the last 3 years.
- h. The bidder or the manufacturer must have an average annual turnover of at least Rs.1,00,00,000/-(Rupees one crore only) each during the last three financial years i.e. 2017-18, 2018-19, 2019-20.
- i. Bidder should have ISO Certification.

4. Purchase Preference

- i. The Procurement of goods and services under this e-tender will be regulated as per the applicable provisions of Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as further amended by orders of even number dated 28.05.2018, 29.05.2019, 04.06.2020 and 16.09.2020. Salient portions of the order are reproduced in Appendix A of this tender document by way of information. Bidders are advised to see the original orders and satisfy themselves that they qualify to participate in the tender. Bidders who are claiming eligibility to participate in this tender must submit a certificate in format given in Appendix A along with documentary evidence in support of their claim wherever necessary failing which their bid will be summarily rejected. The purchaser reserves the right to give preference to the 'Class-I local supplier'.
- ii. The Purchaser reserves the right to give the purchase preference to small-scale sectors, Micro and small scale enterprises etc. as per the instruction in vogue while evaluating, comparing and ranking the responsive Tenders as given in the MSMED Act 2006 reproduced below:
 - a. In exercise of powers conferred in Section 11 of the Micro, Small and Medium Enterprises Development (MSMED) Act 2006, the Government has notified a new Public Procurement Policy for Micro & Small Enterprises effective from 1st April 2012. The policy mandates that 25% of procurement of annual requirement of goods and services by all Central Ministries/Public Sector Undertakings will be from the micro and small enterprises. The Government has also earmarked a sub-target of 4% procurement of goods & services from MSEs owned by SC/ST entrepreneurs out of above said 25% quantity.
 - b. In accordance with the above said notification, the participating Micro and Small Enterprises (MSEs) in a tender, quoting price within the band of L1+15% would also be allowed to supply a portion of the requirement by bringing down their price to the L1 price, in a situation where L1 price is from someone other than an MSE. Such MSEs would be allowed to supply up to 25% of the total tendered value. In case there are more than one such eligible MSE, the 25% supply will be shared equally. Out of 25% of the quantity earmarked for supply from MSEs, 5% quantity is earmarked for procurement from MSEs owned by SC/ST entrepreneurs. However, in the event of failure of such MSEs to participate in the Tender process or meet the

tender requirements and the L1 price, the 5% quantity earmarked for MSEs owned by SC/ST entrepreneurs will be met from other participating MSEs.

- c. The MSEs fulfilling the prescribed eligibility criteria and participating in the Tender shall enclose with their tender a copy of their valid registration certificate with District Industries Centres or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir Board or National Small Industries Corporation or any other body specified by Ministry of Micro and Small enterprises in support of their being an MSE, failing which their tender will be liable to be ignored.
- d. Special provision for Micro and Small Enterprise owned by women: – Out of the total annual procurement from Micro and Small Enterprises, 3 per cent from within the 25 per cent target shall be earmarked for procurement from Micro and Small Enterprises owned by women.
- e. Note: “If the bidder is a MSME, it shall declare in the bid document the Udyog Aadhar Memorandum Number issued to it under the MSMED Act, 2006. If a MSME bidder do not furnish the UAM Number along with bid documents, such MSME unit will not be eligible for the benefits available under Public Procurement Policy for MSEs Order 2012.

5. Demonstration of sample

The tenderer must give a sample within **ten** days of the closing of online submission of bids failing which the bids will be rejected.

6. Instructions for the filling the tender form

- i. E-Tender form shall be completed in all respect, signed in full and stamped at appropriate places and initialed and stamped on all remaining pages. Incomplete or e-tenders without tender processing fee, bid security declaration, Make-in-India self-certification as in Appendix-A only by original manufacturer, GFR 144 (xi) compliance certificate as in Section-X shall be treated as invalid.
- ii. Bidders have to ensure that all the documents are properly filled.
- iii. Conditional tenders are liable to be rejected.
- iv. Bids received and found valid will be evaluated by JIPMER to ascertain the complete work/services under the specification and documents. The bidder should take care to submit all the information sought by JIPMER in prescribed formats.
- v. Incomplete bids, bids in paper format, conditional bids, telephonic bids or tenders submitted after the due date and time will not be considered and summarily rejected. Vendors are, therefore, advised to submit their bids well on time.
- vi. **The bidder can quote for one or more items mentioned in the list. Bidder has to give all details (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) mentioned in BOQ for all quoted items, failure of that the bid will be rejected summarily.**

7. Submission of tenders

The bidders must ensure that they submit the **on-line bids** within the scheduled closing date & time.

8. Late Tender:

There is NO PROVISION of uploading late tender beyond stipulated date & time in the e-tendering system.

9. Alteration and Withdrawal of Tender

- i. The bidder, after submitting its bid, is permitted to alter/modify its bid, within the deadline for submission of bids. Alterations/modifications to bids received after the prescribed deadline will not be possible on the e-tender portal.
- ii. No tender should be withdrawn or modified after the deadline for submission of tender and before expiry of the tender validity period. If a bidder withdraws or modifies the tender during this period, it will result in the tenderer being suspended from participating in all tenders invited by JIPMER for a period of two years with effect from the date of withdrawal, as per the terms and conditions of bid security declaration furnished by the bidder in its tender.

10. Preparation of e-tenders

This is a Two-Bid Tender system, consisting of the **Techno-Commercial Bid and Price Bid** that are to be uploaded in the prescribed formats in the e-tendering portal. The tender(s) shall only be submitted online as mentioned below:

I. **Techno-commercial Bid shall comprise**

- a. Fee Cover
 - i. **E-tender Processing fee** of Rs.590/- (Rupees five hundred and ninety only) inclusive of 18% GST payment **receipt** duly self-attested and rubber stamped should be uploaded.
 - ii. Scanned copy (100 or 200 DPI) in pdf format of **Bid Security Declaration** as per format attached in Section-IX
- b. In the cover named “Prequal/Technical” the scanned copy in pdf format of the following documents are to be uploaded:
 - i. The **Bidder’s Profile** as in Section-VI, **Tender Form** as in Section-VII must be downloaded duly filled signed and stamped, and an **Authorization letter** for signing tender documents if a person other than the Owner, Partner, Managing Director is signing the documents, must be uploaded as a single PDF file.

- ii. **Copies of Supply orders/Completion certificate** in support of Eligibility condition 1 and 2 and **Copies of abridged Annual report of last 03 years** (Income tax return acknowledgement, Assets and Liabilities, Balance sheet and Profit & Loss Account) must be uploaded as a single PDF file
- iii. **Manufacturer's Authorization letter** in company letterhead in format in given Section-VI and, in case the bidder is empaneled by the Competent Authority under **GFR 144 (xi) (mandatory)** a copy of the same or **GFR 144(xi) compliance certificate** as in Section-X must be uploaded.
- iv. Copy of Self Certified **GST registration certificate** and Copy of **PAN Card** must be uploaded as a single PDF file.
- v. **Bank Details** (Beneficiary name, Bank name, Account number, IFSC code, Branch address on letterhead) and **Check list** as in Section-VIII in the prescribed format duly filled and signed must be uploaded as a single PDF file.
- vi. Self-certification for supporting the claim to be a local supplier under the "Public Procurement preference to Make in India" order in format as in **Appendix-A (mandatory)** and any other document that the bidder wishes to submit as a single PDF file.
- vii. **A file mentioning the list of items without price** for which bidder is quoting must be uploaded as a single PDF file.

II. Price Bid:

Prices are to be quoted in the prescribed Price Bid format provided in the e-tender portal using the BOQ template only. The price should be quoted for the **accounting unit** indicated in the e-tender document.

Note:

- i) The bidder has to be diligent while filling up the Techno-Commercial Bid and Price Bid provided in prescribed formats and must not tamper with the contents of the sheets.
- ii) Bidders must ensure that the documents uploaded in pdf format are legible.
- iii) It is the responsibility of bidder to go through the Tender document to ensure furnishing all required documents in addition to above, if any.
- iv) The Make in India self-certification as in Appendix-A must be given and uploaded by original manufacture. The distributor can upload the Appendix-A given by original manufacture.
- v) ITE- Item-wise Eligibility Sheet should be downloaded, the items that the bidder wishes to quote must be selected as "Eligible", and this "ITE file" must also be uploaded for the price bid to be considered by the system. The selected items will be displayed once uploaded and the bidder can verify that all items he wishes to quote for, are present in the list
- vi) A person signing (manually or digitally) the tender form or any documents forming

part of the contract on behalf of another shall be deemed to warrant that he has authority to bind such other persons and if, on enquiry, it appears that the persons so signing had no authority to do so, the purchaser may, without prejudice to other civil and criminal remedies, cancel the contract and hold the signatory liable for all cost and damages.

- vii) A bid, which does not fulfill any of the above requirements and/ or give evasive information/reply against any such requirement, shall be liable to be ignored.
- viii) Tender sent by fax/telex/cable shall be ignored.

11. Digital Signing of Tender

The tenderers shall submit their tenders as per the instructions contained as above. Tenders shall be uploaded with all relevant tender documents in the prescribed format. The relevant tender documents should be uploaded by an authorized person having Class 3 digital signature certificate.

12. Tender currencies.

The tender shall be quoted only in INR.

13. Additional information and instruction on GST:

If the Tenderer desires to ask for GST or any other taxes to be paid extra, the same must be specifically stated. In the absence of any such stipulation, the price will be taken inclusive of such taxes and no claim for the same will be entertained later. The rate of GST quoted in the tender shall be taken for price comparison. However, the rate of GST quoted in the tender or the actual rate of GST applicable, whichever is lower shall be payable by the purchaser. The supplier can charge a higher GST than quoted in the tender only if the rate of GST was revised by the government after the tender closing date.

14. Bid Security (Earnest Money Deposit)

- i. In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender.
- ii. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.
- iii. The bidder hereby declares that they accept the condition that if they withdraw or modify their bids during period of validity etc., they will be suspended for a period of two years from participating in any tender invited by JIPMER, Puducherry with effect from the date of their withdrawal or modification of their bid.

15. Tender opening

- i. The Tender Inviting Authority will open the e-tenders at the specified date and time and at the specified place as indicated in the NIT. In case the specified date of tender opening falls on / is subsequently declared a holiday or closed day for the Tender Inviting Authority, the tenders will be opened at the appointed time and

place on the next working day.

- ii. Authorized representatives of the tenderers, who have submitted tenders on time, may attend the tender opening provided they bring with them letters of authority from the corresponding tenderers. The tender opening official(s) will prepare a list of the representatives attending the tender opening. The list will contain the representatives' names & signatures and corresponding tenderers' names and addresses.
- iii. This being a Two-Bid Tender system, the **Techno-Commercial Tenders** are to be opened in the first instance, at the prescribed time and date as indicated in NIT. These Tenders shall be scrutinized and evaluated by the competent committee/ authority with reference to parameters prescribed in the tender document. During the Techno- Commercial Tender opening, the tender opening official(s) will read the salient features of the tenders like brief description of the goods offered, delivery period, bid security declaration and any other special features of the tenders, as deemed fit by the tender opening official(s). Thereafter, in the second stage, the Price bids of only the Techno - Commercially acceptable offers (as decided in the first stage) shall be opened for further scrutiny and evaluation on a date notified after the evaluation of the Techno-Commercial tender.

16. Scrutiny and evaluation of tenders

A. Basic Principle

Tenders will be evaluated on the basis of the terms & conditions already incorporated in the tender enquiry document, based on which tenders have been received and the terms, conditions etc. mentioned by the tenderers in their tenders. No new condition will be brought in while scrutinizing and evaluating the tenders.

B. Scrutiny of Tenders

- i. The Purchaser will examine the Tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished and, whether the documents uploaded are in legible form.
- ii. The Purchaser's determination of a Tender's responsiveness is to be based on the contents of the tender itself without recourse to extrinsic evidence.
- iii. The tenders will be scrutinized to determine whether they are complete and meet the essential and important requirements, conditions etc. as prescribed in the tender document. The tenders, which do not meet the basic requirements, are liable to be treated as non-responsive and will be summarily ignored.
- iv. The following are some of the important aspects, for which a tender shall be declared non-responsive during the evaluation and will be ignored;
 - i. Tender validity is shorter than the required period.
 - ii. Non-submission of Bid Security Declaration.
 - iii. Non-submission of receipt of tender processing fee.
 - iv. Non-submission of self-certification in format as given in Appendix-A only by original manufacturer, for determining eligibility to participate in the tender under the "Public Procurement preference to Make in India" order.
 - v. Non-submission of GFR-144 (xi) compliance certificate.

- vi. Tenderer has not agreed to give the required performance security of required amount in an acceptable form.
- vii. Non-submission of samples within ten days of the closing of online submission of bids
- viii. Poor/ unsatisfactory past performance.
- ix. Tenderers who stand de-registered/banned/blacklisted by any Central Govt. Ministries/Departments/Hospitals/Institutes.
- x. Tenderer is not eligible as per tender conditions.
- xi. Tenderer has not quoted for the entire quantity as specified in the List of Requirements/ BOQ for the item quoted.
- xii. Non-submission of all details of quoted items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark).
- xiii. Tenderer has not agreed to other essential condition(s) specially incorporated in the tender enquiry, like delivery terms, delivery schedule, terms of payment, liquidated damages clause, warranty clause, dispute resolution mechanism applicable law.

17. Minor Informality/Irregularity/Non-Conformity

If during the preliminary examination, the purchaser find any minor informality and/or irregularity and/or non-conformity in a tender, the purchaser may waive the same provided it does not constitute any material deviation and financial impact and, also, does not prejudice or affect the ranking order of the tenders. Wherever necessary, the purchaser will convey its observation on such 'minor' issues to the tenderer by registered/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.

18. Award of work

- i. The selection of the agency will be at the sole discretion of the JIPMER who reserves its right to accept or reject any or all the proposals without assigning any reason thereof.
- ii. The lowest bid will be evaluated on the basis of annual estimated amount as quoted by bidders in the quotation including taxes.
- iii. Upon evaluation of offers the decision on the award of contract will be intimated to the successful bidder.
- iv. The annual estimate is given only as an indication. The actual quantity procured may increase or decrease. No assurance is given that the quantity stated will actually be procured.

Section-IV

GENERAL CONDITIONS OF CONTRACT

1. **Price of goods**

The rate quoted in the e-tender will be fixed for the whole contract period.

2. **Technical Specifications and Standards**

The Goods & Services to be provided by the supplier under this contract shall conform to the technical specifications mentioned in “Technical Specification” under Section II.

3. **Terms of Delivery**

- i. Goods shall be delivered by the supplier within 30 days of issue of supply order. Please note that the time shall be the essence of the contract.
- ii. Any unexcused delay by the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all of the following sanctions:
 - a) Imposition of liquidated damages,
 - b) Forfeiture of its performance security and
 - c) Termination of the contract for default.

4. **Liquidated Damages**

If the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods, installation, commissioning and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 11. Since the Liquidated damages are in virtue of non-performance of services, it will attract GST or any other applicable taxes which in turn shall be deducted from the bidder.

5. **Performance Security**

The successful bidder shall have to deposit an amount 3% of the value of contract as Performance Security Deposit (PSD) within two weeks after award of contract, through SBI Collect available on JIPMER website. In the event of any failure /default of the supplier with or without any quantifiable loss to the purchaser, the amount of the performance security is liable to be forfeited.

Subject to the condition mentioned above the Performance Security will be released without any interest to the supplier on completion of the supplier's all contractual obligations

including the warranty obligations and extension of time (with or without Liquidated Damages).

6. Payment Procedure

The contractor shall submit bill in triplicate upon satisfactory supply of goods. Bill must be raised based on the rate quoted in e-tender. Every effort shall be made to ensure that the payment will be made within 45 days of submission of bill. No advance payment will be considered. TDS will be deducted as per provision of Income Tax Act, GST Acts and other statutes as relevant.

7. Risk Clause

The contractor shall at all times have standby arrangements for carrying out the work under the contract, in case of any failure of the existing arrangements. JIPMER reserves the right for termination of the contract at any time by giving 30 days written notice, if the items delivered are found to be unsatisfactory and also has the right to award the contract to the next higher bidder willing to supply the item at the cost, risk and responsibilities of contractor and excess expenditure incurred on account of this will be recovered by JIPMER from the contractor's Performance Security Deposit or pending bills or by raising a separate claim.

8. Termination clause:

During the period of agreement if it is found that the agency is not providing proper services, the JIPMER reserves rights to make the vendor forfeit the security deposit deposited with JIPMER or part thereof in favour of JIPMER and agreement will be terminated after giving 30 days' notice. Furthermore, in such situations, tender can be allotted to second lowest bidder and the difference in cost shall be recovered from the earlier vendor who is breach of the contract. In addition in case it is found that the supplier is charging by fraudulent means or indulging in criminal activities the contract will be terminated immediately.

9. Jurisdiction of the courts

Jurisdiction of the courts for settlement of disputes:- Jurisdiction for the settlements of disputes if any is Puducherry only.

In above mentioned conditions Director JIPMER reserves all the rights.

Section-V

BIDDER'S PROFILE

This form duly filled and signed by authorized representative of the bidder and the scanned copy must be uploaded online

1.	Name & Designation of the contact person	
2.	Name and Address of the Tenderer	
3.	Phone No a) Land line number (functional between 9 am and 5pm)	
4.	Mobile No of contact person (available from 9am to 6pm)	
5.	Email ID of the Tenderer	
6.	Email ID of the contact person	
7.	Local supplier/Distributor in Chennai/ Puducherry or any other place (complete address must be written)	
8.	Manufacture Name	
9.	Manufacture Address	
10	Whether Tenderer is registered MSE Manufacture the product quoted. (If registered MSE, submit copy of the Udyog Aadhaar certificate or Equivalent Certificate)	Yes / No
<p>If there is any change in the above details, I will immediately intimate you by speed post or fax or email</p> <p>I.....hereby declare that the details given above are true to the best of my knowledge and I have thoroughly read and understood the terms and conditions of the tender and shall abide by the rules,</p> <p style="text-align: right;">Signature (Name and Designation & Seal)</p> <p>Dated:</p>		

NB: This form must be duly filled in by an authorized person

Section-VI
MANUFACTURER'S AUTHORISATION FORM
(Letterhead)

The Director,
JIPMER, Puducherry

Dear Sir,

Ref: Your TE document No: _____ dated: _____

We, _____ who are proven and reputable manufacturers of _____ (name and description of the goods offered in the Tender) having factories at _____, hereby authorize Messrs. _____ (name and address of the agent) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also state that we are not participating directly in this tender for the following reason(s):
_____. (Please provide reason here).

We further confirm that no supplier or firm or individual other than Messrs _____ (name and address of the above agent) is authorised to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us. We also hereby extend our full warranty, CMC/AMC as applicable as per the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document. We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorised agent.

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly”

Yours faithfully,

[Signature with date, name, designation and Email]

for and on behalf of Messrs _____

[Name & address of the manufacturers]

Note:

- (1) This letter of authorization should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.
- (2) Original letter may be sent.
- (3) The purchaser reserves the right to verify this document with its signatory.

SECTION-VII
TENDER FORM
(On Firms' Letter Head)

To
The Director
JIPMER, Puducherry 605006

Date _____

Ref. Your TE document No. _____ dated _____

We, the undersigned have examined the above mentioned TE document, including amendment/corrigendum No. _____, dated _____ (if any), the receipt of which is hereby confirmed. We now offer to supply and deliver _____ (Description of goods and services) in conformity with your above referred document **for the sum as shown in the price schedules attached herewith and made part of this tender**. If our tender is accepted, we undertake to supply the goods and perform the services as mentioned above, in accordance with the delivery schedule specified in the List of Requirements.

We further confirm that, if our tender is accepted, we shall provide you with a performance security of required amount in an acceptable form in terms of GCC, Special Conditions of Contract”, for due performance of the contract.

We have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and we certify that this bidder is not from such a country/ from such a country and has been registered with the Competent Authority and a copy of the valid registration by the Competent Authority is attached as evidence of the same (Strike out what is not applicable). In case there are Turnkey works to be carried out this bidder will not sub-contract any work to a contractor from such countries unless such contractor is registered with the Competent Authority. We hereby certify that this bidder fulfills all requirements in this regard and is eligible to be considered.

We agree to keep our tender valid for acceptance as required in the GIT, Special Instructions to Tenderers” or for subsequently extended period, if any, agreed to by us. We also accordingly confirm to abide by this tender up to the aforesaid period and this tender may be accepted any time before the expiry of the aforesaid period. We further confirm that, until a formal contract is executed, this tender read with your written acceptance thereof within the aforesaid period shall constitute a binding contract between us.

We further understand that you are not bound to accept the lowest or any tender you may receive against your above-referred tender enquiry.

We confirm that we do not stand deregistered/banned/blacklisted by any Govt. Authorities in the last 7 years.

We confirm that we fully agree to the terms and conditions specified in above mentioned TE document, including amendment/ corrigendum if any.

(Signature with date)
(Name and designation)

Duly authorized to sign tender for and on behalf of

Section-VIII

CHECK LIST FOR SUBMISSION OF TENDER

(To be filled by the tenderer and submitted along with the bid)

Sl. No.	Particular	Yes / No
1	i. Tender Processing Fee (Mandatory for all bidders) ii. Bid security declaration	
2	Copies of Supply orders/Completion certificate in support of Eligibility condition 2(ii).	
3	Copy of PAN Card	
4	Copy of ISO Certification of bidder, ISI/CE Certification of each product wherever asked	
5	Copies of last three years Income Tax Return statement with Balance sheets & Profit & Loss A/c i.e. 2017-18, 2018-19, 2019-20.	
6	Copy of GST Registration Certificate	
7	Authorization letter for signing tender documents if a person other than the Owner, Partner, Managing Director is signing/uploading the tender	
8	Tender form as in Section – VII duly signed and stamped	
9	Manufacturer's authorization form in format as in Section VI and	
10	Self-certification for supporting the claim to be a local supplier under the "Public Procurement preference to Make in India" order in format as in Appendix A	
11	A copy bidder's empanelment by the Competent Authority under GFR 144 (xi) or GFR 144 (xi) compliance certificate	
12	A PDF file containing list of all items quoted by the bidder without price bid in technical cover	
13	All details of the items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) quoted by the bidder.	
14	Any other document(s) enclosed (To be specified)	

I/We certify that the information furnished above is true and correct. The terms and conditions are acceptable to us and have the authority to bid a tender.

Signature of the owner/
Managing Partner/Director

Name:

Seal:

Date:

Place:

SECTION – IX

**Bid Security Declaration
(To printed on the Firm's letterhead)**

(In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.)

I/We hereby declare that we accept the condition that if I/We withdraw or modify our bids during period of validity etc., we will be suspended for a period of two years from participating in all tenders invited by JIPMER, Puducherry with effect from the date of our withdrawal or modification of our bid.

(Signature with date)

(Name and designation)

Duly authorized to sign this tender and declaration

for and on behalf of _____ (Name of the Bidder's Firm) Seal

Appendix-A

Self-certification format for claiming purchase preference under the “Public Procurement preference to Make in India” order

As per the order issued by

(i) Department of Industrial Policy and Promotion (DIPP) vide No. P-45021/2/2017-BE-II dated 15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017-B.E.-II dated 29.05.2019, Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020 and Order No. P-45021/2/2017-PP (BE-II) dated 16.09.2020; and

(ii) Department of Pharmaceuticals vide No. F- 31026/36/2016-MD dated 18.05.2018 and the subsequent orders thereof; The purchaser reserves the right to give preference to the local supplier.

A local supplier (definition of “local supplier” is given in clause 2 of the aforesaid order of DIPP as amended from time to time) has to submit the following along with their e-tender(s) failing which their bid will be evaluated without considering such preference mentioned in the DIPP order dated 15.06.2017 further amended on 28.05.2018, 25.09.2019 and 04.06.2020:

a. The local supplier at the time of e-tender, bidding or solicitation shall be required to provide self-certification that the item offered meets the minimum local content and shall give details of the location(s) at which the local value addition is made in the format in Annexure A.

“Certified that the following items quoted has more than 50% or 20% to 50% of value addition in India at the location(s) mentioned against each and is eligible for purchase preference as per the Govt. of India “Public Procurement preference to Make in India” order Dt.15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017-B.E.-II dated 29.05.2019, and Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020.

Sl. No	Name of the item	Details of the location(s) at which the local value addition was made.

Authorized Signature:

Name:

Designation:

b. In cases of procurement for a value in excess of Rs. 10 Crore. the local supplier shall be required to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.

c. Minimum Local Content: ‘Class-I local supplier’ means a supplier or service provider, whose goods, services or works offered for procurement, has local content equal to or more than 50%, as defined

under the Order. “Class-II local supplier” means a supplier or service provider, whose goods, services or works offered for procurement, has local content more than 20% but less than 50%, as defined under this Order.

d. Margin of Purchase Preference: The margin of purchase preference shall be 20%.

e. Manufacture under license/technology collaboration agreements with phased indigenization are exempted from meeting the stipulated local content if the product is being manufactured in India under a license from a foreign manufacturer who holds intellectual property rights and where there is a technology collaboration agreement/transfer of technology agreement for indigenous manufacture of a product developed abroad with clear phasing of increase in local content

f. Decisions on complaints relating to implementation of this Order shall be taken by the competent authority which is empowered to look into procurement-related complaints relating the procuring entity.

g. A constituted committee with internal and external experts will examine for independent verification of self-declarations and auditor’s/accountant’s certificates on random basis and in the case of complaints.

h. A fees of Rs.10000/- in the form of demand draft favoring The Director, JIPMER, payable at Puducherry, is required to be deposited with complaints for verification of local content.

i. False declarations will be breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.

j. A supplier who has been debarred by any procuring entity for violation of this Order shall not be eligible for preference under this Order for procurement by any other procuring entity for the duration of the debarment. The debarment for such other procuring entities shall take effect prospectively from the date on which it comes to the notice of other procurement entities.

Officer in Charge
Purchase Section
JIPMER Puducherry
For Director, JIPMER

SECTION – X

**GFR-144 (xi) compliance certificate
(To printed on the Firm's letterhead)**

Tender No:

GFR-144(xi) compliance certificate (as per order F.No. 6/18/2019-PPD, Ministry of Finance, GOI)

I have read the clauses regarding restrictions under GFR144(xi) on procurement from a bidder of a country which shares a land border with India. I certify that, the vendor

- is not such a country
- is from a country and has been registered with a competent authority (attached evidence of valid registration).

(Select one of the above and strike off the other)

I hereby certify that we fulfill all requirement in this regard and is eligible to be considered for the procurement on CPP portal.

Thanking you.

Authorized Signatory