



**JAWAHARLAL INSTITUTE OF POST GRADUATE MEDICAL
EDUCATION & RESEARCH (JIPMER)**



GOVERNMENT OF INDIA

(An Institution of National Importance under Ministry of Health & Family Welfare) Dhanvantri
Nagar, Puducherry-605006

**OPEN TENDER
ENQUIRY DOCUMENT
FOR
RATE CONTRACT
FOR
SUPPLY OF CONSUMABLES/NON-CONSUMABLES
FOR THE
DEPARTMENT OF PAEDIATRICS**

(This document consisting of 57 pages)

**JIPMER,
DHANVANTARI NAGAR,
PUDUCHERRY 605006**

Phone:- 0413-2296016/15 ; Fax:- 0413-2272067/66

Email: dept-paediatrics@jipmer.edu.in URL:- www.jipmer.edu.in

INDEX

Section	Topic	Page No.
Section-I	Notice inviting Tender (NIT)	03
Section-II	Schedule of Requirements and Specifications	05
Section-III	General Instructions to Tenderers (GIT)	39
Section-IV	General Conditions of Contract (GCC)	48
Section-V	Bid Application Form	50
Section-VI	Manufacturer's Authorization Form	51
Section-VII	Tender Form	52
Section-VIII	Check list for Submission of Tender	53
Section-IX	Bid Security Declaration	54
Appendix-A	Self-certification format for claiming purchase preference under the "Public Procurement preference to Make in India" order	55
Section-X	GFR-144 (xi) compliance certificate	57

SECTION-I

JAWAHARLAL INSTITUTE OF POSTGRADUATE MEDICAL EDUCATION AND RESEARCH PUDUCHERRY – 6

(Institution of National Importance under the Ministry of Health & Family Welfare,
Govt. of India)

Purchase Section

No. JIP/PUR6(1)/OeT/Paed/2021-22

Dated: 28.07.2021

NOTICE INVITING e-TENDER (e-NIT)

Sub: Supply of Consumables / Non-Consumables

E- tenders are invited from eligible and qualified domestic manufacturer's or their distributors for the RATE CONTRACT for supply of CONSUMABLES/NON-CONSUMABLES for the Department of Paediatrics for 2021– 22.

1. Scope of work: Supply of consumables/non-consumables for the Department of Paediatrics
2. Value of tender: Approximate Cost **Rs. 1,00,00,000**
3. Tender timelines:
 - i. Opening date & time for download of Tender document: **06.00 PM on 28.07.2021**
 - ii. Last date for receipt of pre-bid queries: **06.00 PM on 03.08.2021**
Pre-bid queries can be made through e-mail to: dept-paediatrics@jipmer.edu.in
 - iv. Opening date & time for submission of online bids: **09.00 AM on 09.08.2021**
 - v. Closing date & time for submission of online bids **06.00 PM on 30.08.2021**
 - vi. Date and Time of opening of online bids for Technical evaluation: **11.00 AM on 01.09.2021**
 - vii. Date & time of opening of Price Bid: To be intimated later.
4. Earnest Money Deposit/Bid Security: Bid security declaration must be submitted by all bidders.
5. Tender Processing Fee (Non-refundable): **Rs. 590 (including 18% GST) shall be paid through SBI collect only.**
<https://www.onlinesbi.com/sbicollect/icollecthome.htm?corpID=362853>
6. Interested bidders are advised to download the complete Tender Enquiry document from the websites www.jipmer.edu.in or <https://eprocure.gov.in/eprocure/app> for complete details.
7. The prospective bidders must register with the E-procurement system of <https://eprocure.gov.in/eprocure/app>. Special Instructions to the bidders for the e-submission of the bids online through this eProcurement Portal on completion of the

registration process is given in <https://eprocure.gov.in/eprocure/app> , the bidders will be provided user ID and password upon enrollment. In order to submit the bids electronically, bidders are required to have a valid Class 3 Digital Signature Certificate (signing and encryption/ decryption certificates).

8. Bidders are requested to read the bidders help document on e-tender web site link before proceeding for bidding.
9. Post receipt of User ID & Password, Bidders can log on for downloading & uploading tender document.
10. The bidders shall submit the required bid security declaration (as per G.I.T clause 2) before the due date and time mentioned above.
11. The online submission of tender(s) can only be done through <https://eprocure.gov.in/eprocure/app>
12. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through <https://eprocure.gov.in/eprocure/app> e-portal (as described above) only.
13. Prospective bidders are advised to browse the above websites regularly before submission of their bids as any further amendments will be published in these websites only.

Asst. Officer in Charge
Purchase Section,
For
Director,
JIPMER, Puducherry

Section-II

SCHEDULE OF REQUIREMENTS AND SPECIFICATIONS

LIST OF ITEMS REQUIRED

1. Scope of work : Supply of Consumables/ Non-Consumables for the department of Pediatrics as detailed below

Sl. No.	Item Description	Quantity	Units
1	CONSUMABLE		-
1.001	Infant feeding tube - 6F	2000	Nos.
1.002	Infant feeding tube - 7F	1000	Nos.
1.003	Infant feeding tube - 8F	1000	Nos.
1.004	Infant feeding tube - 9F	2000	Nos.
1.005	Infant feeding tube - 10F	2500	Nos.
1.006	Yankauers Oral Suction tube - Pediatric Size. Sample demonstration is mandatory	50	Nos.
1.007	Disposable sterile De Lee mucous extractor	50	Nos.
1.008	Disposable sterile De Lee mucous extractor with adapter for bronchoscopy	50	Nos.
1.009	Intercostal drainage bottle (Single bottle underwater seal) - Should have atleast one litre capacity. Should be supplied along with connection tubes and facility for hanging to the patient's cot.	50	Nos.
1.010	Foleys catheter (Rubber) - 6F	200	Nos.
1.011	Foleys catheter (Silicone) - 6F	100	Nos.

1.012	Foleys catheter (Rubber) - 8F	200	Nos.
1.013	Foleys catheter (Silicone) - 8F	100	Nos.
1.014	Foleys catheter (Rubber) - 10F	200	Nos.
1.015	Foleys catheter (Rubber) - 12F	200	Nos.
1.016	Foleys catheter (Rubber) - 14F	200	Nos.
1.017	Urine collecting bag (To be attached to foleys catheter)	500	Nos.
1.018	Urine collecting bag with uroflowmeter (To be attached to foleys catheter)	500	Nos.
1.019	Neonatal urine collecting bag- adhesive type	100	Nos.
1.020	Endotracheal tube with micro-sized cuffed - Should made on advanced microthin polyurethane material (10 microns). It should be a true high volume, low pressure (HVLP) cuff, effectively sealing at an average of 10 to 11 cm H2O. Anatomically-based intubation depth mark should be there that help to place in correct position and a cuff-free subglottic zone. Precision bands (minimum of four) facilitate and confirm optimal tube placement. Should be compatible to Oral/Nasal intubation and pre-curved. Size: 3mm Internal Diameter)	200	Nos.

1.021	<p>Endotracheal tube with micro-sized cuffed - Should made on advanced microthin polyurethane material (10 microns). It should be a true high volume, low pressure (HVLP) cuff, effectively sealing at an average of 10 to 11 cm H2O. Anatomically-based intubation depth mark should be there that help to place in correct position and a cuff-free subglottic zone. Precision bands (minimum of four) facilitate and confirm optimal tube placement. Should be compatible to Oral/Nasal intubation and pre-curved. (Size: 3.5mm Internal Diameter)</p>	200	Nos.
1.022	<p>Endotracheal tube with micro-sized cuffed - Should made on advanced microthin polyurethane material (10 microns). It should be a true high volume, low pressure (HVLP) cuff, effectively sealing at an average of 10 to 11 cm H2O. Anatomically-based intubation depth mark should be there that help to place in correct position and a cuff-free subglottic zone. Precision bands (minimum of four) facilitate and confirm optimal tube placement. Should be compatible to Oral/Nasal intubation and pre-curved. (Size: 4mm Internal Diameter)</p>	200	Nos.
1.023	<p>Endotracheal tube with micro-sized cuffed - Should made on advanced microthin polyurethane material (10 microns). It should be a true high volume, low pressure (HVLP) cuff, effectively sealing at an average of 10 to 11 cm H2O. Anatomically-based intubation depth mark should be there that help to place in correct position and a cuff-free subglottic zone. Precision bands (minimum of four) facilitate and confirm optimal tube placement. Should be compatible to Oral/Nasal intubation and pre-curved. (Size: 4.5mm Internal Diameter)</p>	200	Nos.
1.024	<p>Endotracheal tube with micro-sized cuffed - Should made on advanced microthin polyurethane material (10 microns). It should be a true high volume, low pressure (HVLP) cuff, effectively sealing at an average of 10 to 11 cm H2O. Anatomically-based intubation depth mark should be there that help to place in correct position and a cuff-free subglottic zone. Precision bands (minimum of four) facilitate and confirm optimal tube placement. Should be compatible to Oral/Nasal intubation and pre-curved. (Size: 5mm Internal Diameter)</p>	200	Nos.

1.025	Endotracheal tube with micro-sized cuffed - Should made on advanced microthin polyurethane material (10 microns). It should be a true high volume, low pressure (HVLP) cuff, effectively sealing at an average of 10 to 11 cm H2O. Anatomically-based intubation depth mark should be there that help to place in correct position and a cuff-free subglottic zone. Precision bands (minimum of four) facilitate and confirm optimal tube placement. Should be compatible to Oral/Nasal intubation and pre-curved. (Size: 5.5mm Internal Diameter)	100	Nos.
1.026	Reusable flexible Laryngeal mask airway (Size-1)	5	Nos.
1.027	Reusable flexible Laryngeal mask airway (Size-1.5)	5	Nos.
1.028	Reusable flexible Laryngeal mask airway (Size-2)	5	Nos.
1.029	Reusable flexible Laryngeal mask airway (Size-2.5)	5	Nos.
1.030	Reusable flexible Laryngeal mask airway (Size-3)	5	Nos.
1.031	Guedel Oropharyngeal airway (Size - 00) Neonatal	5	Nos.
1.032	Guedel Oropharyngeal airway (Size - 0) Infant	10	Nos.
1.033	Guedel Oropharyngeal airway (Size - 1) Child	10	Nos.
1.034	Guedel Oropharyngeal airway (Size - 2) Small adult	5	Nos.
1.035	Soft Nasopharyngeal Airways (Inner diameter 3 mm)	5	Nos.
1.036	Soft Nasopharyngeal Airways (Inner diameter 3.5 mm)	5	Nos.
1.037	Soft Nasopharyngeal Airways (Inner diameter 4 mm)	5	Nos.
1.038	Soft Nasopharyngeal Airways (Inner diameter 4.5 mm)	5	Nos.

1.039	Soft Nasopharyngeal Airways (Inner diameter 5 mm)	5	Nos.
1.040	Endotracheal tube holder - Neonatal size. Sample demonstration is mandatory	5	Nos.
1.041	Endotracheal tube holder - Pediatric size. Sample demonstration is mandatory	5	Nos.
1.042	Endotracheal tube holder - Adult size. Sample demonstration is mandatory	5	Nos.
1.043	Nasal prongs for oxygen delivery - Neonatal size	100	Nos.
1.044	Nasal prongs oxygen delivery - Pediatric size	500	Nos.
1.045	HFNC nasal prongs compatible with Fischer and Plackel Airvo 2 Optiflow HFNC system (Neonatal)	50	Nos.
1.046	HFNC nasal prongs compatible with Fischer and Plackel Airvo 2 Optiflow HFNC system (Infant)	50	Nos.
1.047	HFNC nasal prongs compatible with Fischer and Plackel Airvo 2 Optiflow HFNC system (Pediatric)	50	Nos.
1.048	Non rebreathing Oxygen mask (NRM) with reservoir - Paediatric	200	Nos.
1.049	Flow inflating bag (paediatric) for resuscitation. The bag should be flow inflating with an oxygen source. The bag size should be suitable for infant and pediatric use. The bag should have a PEEP regulating mechanism. Sample demonstration is a must.	15	Nos.

1.050	Flow inflating bag (Infant) for resuscitation. The bag should be flow inflating with an oxygen source. The bag size should be suitable for infant and pediatric use. The bag should have a PEEP regulating mechanism. Sample demonstration is a must.	15	Nos.
1.051	Reusable stylet for endotracheal intubation compatible with endotracheal tube sizes 3, 3.5, 4, 4.5, 5, 5.5, 6. (Item Should be brought for demonstration)	20	Nos.
1.052	Gum elastic bougie (to aid difficult intubation) - Pediatric size (Item Should be brought for demonstration)	5	Nos.
1.053	Gum elastic bougie (to aid difficult intubation) - Neonatal size (Item Should be brought for demonstration)	5	Nos.
1.054	Paediatric Laryngoscope (macintosh) with blade sizes 0,1,2,3. (Item Should be brought for demonstration)	27	Nos.
1.055	Paediatric Laryngoscope (Miller) with blade sizes 0,1,2. (Item Should be brought for demonstration)	27	Nos.
1.056	Paediatric Fiberoptic Laryngoscope (macintosh) with blade sizes 0,1,2,3. (Item Should be brought for demonstration)	30	Nos.

1.057	Paediatric Fiberoptic Laryngoscope (miller) with blade sizes 0,1,2,3. (Item Should be brought for demonstration)	5	Nos.
1.058	Laryngoscope bulb (LED)	15	Nos.
1.059	Endotracheal tube cuff pressure monitor	5	Nos.
1.060	Emergency cricothyroidotomy kit. Seldinger technique based.	3	Nos.
1.061	Arterial catheter for intra arterial access with straight guidewire and introducer needle - 18 G . Sample demonstration is mandatory	100	Nos.
1.062	Arterial catheter for intra arterial access with straight guidewire and introducer needle - 20 G. Sample demonstration is mandatory	100	Nos.
1.063	Arterial catheter for intra arterial access with straight guidewire and introducer needle- 22 G . Sample demonstration is mandatory	100	Nos.
1.064	Arterial catheter for intra arterial access with straight guidewire and introducer needle - 24 G. Sample demonstration is mandatory	100	Nos.

1.065	Peripherally Inserted Central Venous Catheter with introducer 2F / 24G / Length 15 cm - Each kit should have Puncture needle / introducer needle with Peelable cannula.	50	Nos.
1.066	Double lumen hemodialysis catheter kit (Pediatric) including (1) Straight two lumen catheter- 6.5F - 12cm, (2) Dilator, (3) Spring guidewire (4) Introducer needle	40	Nos.
1.067	Double lumen hemodialysis catheter kit (Pediatric) including (1) Straight two lumen catheter- 8.5F-12cm, (2) Dilator, (3) Spring guidewire (4) Introducer needle	40	Nos.
1.068	Double lumen hemodialysis catheter kit (Pediatric) including (1) Straight two lumen catheter- 11.5F - 10cm, (2) Dilator, (3) Spring guidewire (4) Introducer needle	100	Nos.
1.069	Pigtail catheter -12F with guidewire, seial dilators and introducer needle (Seldinger technique) Sample demonstration is mandatory	20	Nos.
1.070	Pigtail catheter -14F with guidewire, seial dilators and introducer needle (Seldinger technique) Sample demonstration is mandatory	20	Nos.
1.071	Non-Vented NIV oronasal mask compatible with maquet dual limb breathing circuit - (Small size) with head gear	20	Nos.

1.072	Non-Vented NIV oronasal mask compatible with maquet dual limb breathing circuit - (Medium size) with head gear	20	Nos.
1.073	Heater wire for Paediatric ventilatory circuit (inspiratory limb) compatible with Fischer and paykel MR 850 humidifier (27W- 0.95m). (Item Should be brought for demonstration)	25	Nos.
1.074	Reusable heater wire adaptor (single) compatible with Fischer and paykel MR 850 humidifier. (Item Should be brought for demonstration)	13	Nos.
1.075	Reusable heater wire adaptor (dual adapter) compatible with Fischer and paykel MR 850 humidifier.(Item Should be brought for demonstration)	13	Nos.
1.076	Expiratory filter for paediatric ventilator circuit compatible with Maquet servo I ventilator. Sample demonstration is mandatory	1000	Nos.
1.077	Airway temperature probe compatible with Fischer and paykel MR 850 humidifier. (Item Should be brought for demonstration)	25	Nos.
1.078	Humidification chamber (infant) compatible with Fischer and paykel MR 850 humidifier. (Item Should be brought for demonstration)	20	Nos.

1.079	Humidification chamber (Adult compatible with Fischer and paykel MR 850 humidifier. (Item Should be brought for demonstration)	20	Nos.
1.080	Expiratory Cassette compatible with Maquet servo I ventilator. Sample demonstration is mandatory	15	Nos.
1.081	HFOV ventilator circuit compatible with sensormedics A HFOV ventilator	5	Nos.
1.082	Reusable vibrating mesh nebulizer compatible with aerogen pro portable inline nebulisation system for delivering aerosols through ventilator circuits. It should have low row residual volume in medication cup (average 0.3 mL of 3 mL dose, FPF (<5um) should be above 80%. Item should be autoclavable. Sample demonstration is mandatory	10	Nos.
1.083	Disposable defibrillator pads compatible with ZOLL defibrillator - Pediatric (<10kg)	10	Nos.
1.084	Disposable defibrillator pads compatible with ZOLL defibrillator - Adult (>10kg)	10	Nos.
1.085	Transparent I.V site Dressing Size-7X9 cms	100	Nos.
1.086	Transparent - Central Line dressing - size 10x12 cms	200	Nos.

1.087	Sterile ultrasound probe cover with sterile gel	250	Nos.
1.088	Intravenous infusion set with drip chamber and measuring set (Burette Type)	100	Nos.
1.089	Intravenous infusion set compatible with Optima VS volumetric infusion pump (I.V Set for optima infusion pump)	150	Nos.
1.090	Three way stop cock	5000	Nos.
1.091	Three way stop cock with 10 cm extn lines	5000	Nos.
1.092	Closed Luer IV Access system - Single Lumen Type - Silicone split septum. Unobstructed fluid path. No moving parts. Transparent housing. 15 cm single extension set with flow rate 3L/hr.	1000	Nos.
1.093	Closed Luer IV Access system - Bi-Lumen Type - Silicone split septum. Unobstructed fluid path. No moving parts. Transparent housing. 15 cm single extension set with flow rate 3L/hr.	1000	Nos.
1.094	Benedicts reagent (Pack of 5L)	5	Nos.
1.095	Peritoneal dialysis straight-stiff catheter (infant)	100	Nos.
1.096	Peritoneal dialysis straight-stiff catheter (pediatric)	50	Nos.
1.097	CAPD catheter swan neck double cuff 37.5 cm length (Paediatric size)	15	Nos.
1.098	Hemodialysis circuit for Paediatrics (75cm length)	80	Nos.

1.099	Arteriovenous blood tubing for hemodialysis, 1/4" diameter, with two wound dressing caps and sterri trap & two injection ports. For use in older children(>25kg) in pediatric intensive care unit(PICU)	100	Nos.
1.100	Polysulfone synthetic membrane low flux hollow fibre dialyzer with detachable headers, gamma sterilized, with effective surface area 0.3 m ²	5	Nos.
1.101	Polysulfone synthetic membrane low flux hollow fibre dialyzer with detachable headers, gamma sterilized, with effective surface area 0.7 m ²	60	Nos.
1.102	Polysulfone synthetic membrane low flux hollow fibre dialyzer with detachable headers, gamma sterilized, with effective surface area 1.0 m ²	25	Nos.
1.103	Polysulfone synthetic membrane low flux hollow fibre dialyzer with detachable headers, gamma sterilized, with effective surface area 1.3 m ²	5	Nos.
1.104	Plasma filter, Hollow fiber, Polysulfone or polyethylene, 0.3m ²	25	Nos.
1.105	Plasma filter, Hollow fiber, Polysulfone or polyethylene, 0.6m ²	25	Nos.
1.106	CRRT Paediatric circuit compatible with Prismaflex CRRT machine. (M60)(Item Should be brought for demonstration)	20	Nos.

1.107	CRRT Infant circuit compatible with Prismaflex CRRT machine. (HF20) (Item Should be brought for demonstration)	20	Nos.
1.108	CRRT Paediatric circuit compatible with Fresenius CRRT machine. (Item should be brought for demonstration)	10	Nos.
1.109	CRRT Infant circuit compatible with Fresenius CRRT machine. (Item Should be brought for demonstration)	10	Nos.
1.110	CRRT replacement fluid compatible with Fresenius and Prismaflex CRRT machine	200	Nos.
1.111	Peritoneal dialysis fluids (1.7% Dextrose)	200	Nos.
1.112	ECMO circuit compatible with Maquet Rotaflow ECMO system (Pediatric)	3	Nos.
1.113	Pediatric ECMO membrane oxygenator- Heparin coated, Microporous polypropylene/polymethylpentane fibres with blood flow rate upto 2.8L/min, Gas exchange rate upto 5.6L/min, surface area 0.8 m2 compatible with Maquet Rotaflow ECMO system (Pediatric)	3	Nos.
1.114	ACT cartridge compatible with Medtronic ACTplus automated coagulation timer	2000	Nos.

1.115	Fingertip pulseoximeter for SPO2 and pulse rate with Masimo Signal Extraction Technology and Measure-through Motion and Low Perfusion technology. It should display Perfusion index and Pleth Variability index (Pediatrics)	10	Nos.
1.116	Portable vital signs monitor for transport. Should be able to measure oxygen saturation, heart rate and ECG. Should have Clear LCD display of atleast 7 inches. Atleast 2 compatible ECG cables and Oxygen saturation probes of different sizes (Infant and Pediatric size) with cables should be provided. Sample demonstration is mandatory	4	Nos.
1.117	ETCO2 (mainstream) sensor with cable compatible with Philips Efficia CM10 multiparameter monitor	5	Nos.
1.118	ETCO2 (mainstream) sensor compatible with Schiller Truscope multiparameter monitor	5	Nos.
1.119	ETCO2 (mainstream) cable compatible with Schiller Truscope multiparameter monitor	5	Nos.
1.120	ETCO2 (mainstream) sensor with cable compatible with Maque servo ventilator	5	Nos.
1.121	Rectal temperature sensor with cable compatible with Philips Efficia CM10 multiparameter monitor	5	Nos.

1.122	Disposable Pressure Transducer with Flush device and connection tube compatible with Meditec england multiparameter monitor - Should be supplied along with appropriate IBP cable suitable for Meditec england multiparameter monitor -10 numbers atleast.	100	Nos.
1.123	Disposable Pressure Transducer with Flush device and connection tube compatible with Philips MP 20 monitor - Should be supplied along with appropriate IBP cable suitable for PHILIPS MP 20 monitor -10 numbers atleast.	250	Nos.
1.124	Disposable Pressure Transducer with Flush device and connection tube compatible with Schiller Truscope monitor - Should be supplied along with appropriate IBP cable suitable for Schiller Truscope monitor -10 numbers atleast.	100	Nos.
1.125	SpO2 probe Reusable clip type (Paediatric) compatible with Schiller Truscope monitor.(Item Should be brought for demonstration)	10	Nos.
1.126	SpO2 probe Reusable clip type (Paediatric) compatible with Nihon Kodein Vismo monitor.(Item Should be brought for demonstration)	5	Nos.
1.127	SpO2 probe reusable clip type(Paediatric) compatible with Philips Intellivue MP20 monitor. (Item Should be brought for demonstration)	10	Nos.
1.128	SpO2 probe reusable Clip type (Paediatric) compatible with L&T STAR 55 monitor.(Item Should be brought for demonstration)	10	Nos.

1.129	SpO2 probe reusable Wrap type (Infant) compatible with L&T STAR 55 monitor.(Item Should be brought for demonstration)	10	Nos.
1.130	SpO2 probe Reusable Clip type (Pediatric) compatible with Masimo Radical pulse oximeter. (Item Should be brought for demonstration)	20	Nos.
1.131	SpO2 probe Reusable Clip type (Infant) compatible with Masimo Radical pulse oximeter. (Item Should be brought for demonstration)	10	Nos.
1.132	SpO2 Connecting cable (between monitor and probe) compatible with Philips Intellivue MP20 monitor. (Item Should be brought for demonstration)	5	Nos.
1.133	SpO2 Connecting cable (between monitor and probe) compatible with Schiller Truescope multiparameter monitor. (Item Should be brought for demonstration)	5	Nos.
1.134	SpO2 Connecting cable (between monitor and probe) compatible with Meditech England multiparameter monitor. (Item Should be brought for demonstration)	5	Nos.
1.135	SpO2 Connecting cable (between monitor and probe) compatible with Masimo radical pulse oximeter. (Item Should be brought for demonstration)	5	Nos.
1.136	SpO2 Connecting cable (between monitor and probe) compatible with Nihon Kohden Vismo multiparameter monitor. (Item Should be brought for demonstration)	5	Nos.

1.137	SpO2 Connecting cable (between monitor and probe) compatible with GE carescape multiparameter monitor. (Item Should be brought for demonstration)	10	Nos.
1.138	NIBP cuff compatible with Nihon Kohden Vismo multiparameter monitor - Infant size. (Item Should be brought for demonstration)	25	Nos.
1.139	NIBP cuff compatible with Nihon Kohden Vismo multiparameter monitor - Paediatric size. (Item Should be brought for demonstration)	25	Nos.
1.140	NIBP cuff compatible with Nihon Kohden Vismo multiparameter monitor - Adult size. (Item Should be brought for demonstration)	25	Nos.
1.141	NIBP cuff compatible with Nihon Kohden Vismo multiparameter monitor - Infant size. (Item Should be brought for demonstration)	25	Nos.
1.142	NIBP cuff compatible with Nihon Kohden Vismo multiparameter monitor - Paediatric size. (Item Should be brought for demonstration)	25	Nos.
1.143	NIBP cuff compatible with Nihon Kohden Vismo multiparameter monitor - Adult size. (Item Should be brought for demonstration)	25	Nos.
1.144	NIBP cuff compatible with Philips MP 40 multiparameter monitor - Infant size. (Item Should be brought for demonstration)	25	Nos.

1.145	NIBP cuff compatible with Philips MP 40 multiparameter monitor - Paediatric size. (Item Should be brought for demonstration)	25	Nos.
1.146	NIBP cuff compatible with Philips MP 40 multiparameter monitor - Adult size. (Item Should be brought for demonstration)	25	Nos.
1.147	NIBP cuff compatible with L&T star 55 multiparameter monitor - Neonatal size. (Item Should be brought for demonstration)	30	Nos.
1.148	NIBP cuff compatible with L&T star 55 multiparameter monitor - Infant size. (Item Should be brought for demonstration)	30	Nos.
1.149	NIBP cuff compatible with Meditech England multiparameter monitor - Infant size. (Item Should be brought for demonstration)	30	Nos.
1.150	NIBP cuff compatible with Meditech England multiparameter monitor - Neonatal size. (Item Should be brought for demonstration)	30	Nos.
1.151	NIBP cuff compatible with Meditech England multiparameter monitor - Infant size. (Item Should be brought for demonstration)	30	Nos.
1.152	NIBP cuff compatible with Meditech England multiparameter monitor - Paediatric size. (Item Should be brought for demonstration)	25	Nos.

1.153	NIBP cuff compatible with Meditech England multiparameter monitor - Adult size. (Item Should be brought for demonstration)	25	Nos.
1.154	3 Lead ECG cable (Paediatric) compatible with Philips Intellivue MP20 monitor. (Item Should be brought for demonstration)	5	Nos.
1.155	EEG electrodes compatible with Nihon Kohden Neurofac EEG-1200 EEG machine	1 set	Nos.
1.156	EEG electrode paste for EEG electrodes compatible with Nihon Kohden Neurofac EEG-1200 EEG machine	2	Nos.
1.157	Lumbar puncture needle (22 G) sterile, disposable for single use	50	Nos.
1.158	Lumbar puncture needle (24 G) sterile, disposable for single use	50	Nos.
1.159	Blood Ketone measuring Strips - Compatible glucometer should be provided at free of cost. - Each strip should be covered in separate water/moister resistant folder.	1500	Nos.
1.160	Point of care INR and PT test strips- Compatible devic should be provided. Sample demonstration mandatory	500	Nos.
1.161	Procalcitonin(PCT -Q) Test kit (Immunochemical test for Beside application)	1000	Nos.

1.162	Bedside measurement of cardiac injury markers(TROP-I/T, CK-MB, Myoglobin, D-Dimer, NT - proBNP, BNP in single or combination) kit compatible with available equipment in JIPMER or the equipment etc should be provided free of cost if needed	500	Nos.
1.163	Point of care Ammonia testing device. Sample demonstration mandatory	1	Nos.
1.164	Urine Test Strip-Multipurpose (Protein, Blood, pH, Leukocte, Nitrate, Glucose, Urobilinogen, ketone,Bilirubin, Specific gravity)	1100	Nos.
1.165	Urine albumin and glucose reagent strips	1000	Nos.
1.166	Urine ketone, glucose reagent strips	1000	Nos.
1.167	Micro ESR tube (Capillary tube) - heparinized - Should be heparinised. Should withstand centrifuge of 3000 rpm for 5 minute. Inner diameter of atleast 1.2 mm. Length of atleast 75 mm	10000	Nos.
1.168	Neubauer's cell counting chamber (for RBC & WBC counting)	20	Nos.
1.169	Thermal seal packaging material (with solid paper base and a transparent plastic sheet on the top for thermal Small, Medium	25	Nos.
1.170	ECG paper compatible with Biocare IE 12 A ECG machine. Paper: 210mm x 150mm Z-fold paper (A4 size)	100	Nos.

1.171	Infantometer with Height Measuring Scale - Sleek broad acrylic base with one sliding side adjustable as per length of baby, dual scale for direct reading in cms / inches upto 90cm. Folding sides for easy storage, baby can comfortably lay on the chest of the scale. Size : minimum 18" x 7" (Length x Width). Sample demonstration is mandatory	8	Nos.
1.172	Stadiometer - Height measuring scale, floor model without weighing scale, Built should be sturdy. Measuring range: 20 - 210cm with 1mm graduation. Measuring in inches and centimeters. Sample demonstration is mandatory	5	Nos.
1.173	Wall mounted height measuring scale- Self retracting. Measuring range upto 200 cm	3	Nos.
1.174	Electronic kitchen weighing scale. Should be able measure weight in range of 1g-5000 g. Should have easy TARE function to reset scale to zero. Should have clear LCD display screen with large font. Should have auto-off feature. Sample demonstration is mandatory	6	Nos.
1.175	Infant digital weighing machine with tray for holding baby. Range upto 20 kg with accuracy of 5 g. Sample demonstration is mandatory	5	Nos.
1.176	Electronic Digital Weighing Scale - Platform type with sturdy metallic build. Weighing capacity : 150Kg with accuracy of 10g. Power supply 230 V AC, 50 Hz. Sample demonstration is mandatory	4	Nos.
1.177	Electronic (Digital) BP apparatus for Paediatrics. Should have the following features: 1. Should be supplied with minimum three cuff sizes, infant, paediatric and adult. 2. Should be supplied with A/C adapter suitable with 230V 3. Item should be brought for demonstration	25	Nos.

1.178	Handheld aneroid sphygmomanometer. Should be light weight and shock resistant. Should have a calibration warranty of atleast 5 years. Should be supplied with minimum 3 cuff sizes - Infant, Pediatric, Adult.	25	Nos.
1.179	White light torch for eye examination. Should have pupil gauge depicted on the body	25	Nos.
1.180	Pediatric vein finder deep infrared light with inbuilt Rechargeable Battery (Infant)	3	Nos.
1.181	Pilocarpine disk gels and conductivity sensors kit compatible with nanoduct sweat analysis system. (Item Should be brought for demonstration)	5	Nos.
1.182	Peak flow meter (Sample demonstration is mandatory)	8	Nos.
1.183	Pediatric Stethoscope- One-inch traditional bell (pediatric) and 3/4 inch bell (infant) combined with a floating diaphragm with antichill rim, Single-lumen tubing. (Sample demonstration is mandatory)	35	Nos.
1.184	Skinfold caliper for measuring skinfold thickness Measuring Range: 0mm to 80mm Measuring Pressure: 10gms/mm ² (constant over range) Accuracy: 99.00% - Repeatability: 0.20mm. (Sample demonstration is mandatory)	3	Nos.
1.185	Prader orchidometer (Sample demonstration is mandatory)	3	Nos.

1.186	Mini centrifuge for centrifuging 2 ml vials. Capacity: Minimum 6 vials. Maximum rotation per minute: Atleast 6000. RPM should be adjustable. Centrifuge should have inbuilt brake. (Sample demonstration is mandatory	2	Nos.
1.187	Induction stove(Cooktop) - Wattage should be atleast 2000W, Operating voltage - 220-240V, Cord length should be atleast 1 metre, Should have microcrystal plate for placing utensils, Should have touch start, auto-off program and time setting for 0-3 hours	2	Nos.
1.188	Endoscopic variceal sclerotherapy injection needle with locking facility, 23 G, presterilized and disposable	5	Nos.
1.189	Multiband ligator set of 6 bands for EVL, with barrel,6 bands and trigger cord, compatible for use with 2.8 mm accessory channel with trigger handle and irrigation port	5	Nos.
1.190	Multiband ligator set of 6 bands for EVL, with barrel,6 bands and trigger cord, compatible for use with 2.8 mm accessory channel	5	Nos.
1.191	Radio-opaque, tissue compatible, polyurethane naso-jejunal feeding tube 12 Fr and 120 cm long with hollow, flexible guidewire and Nasal fixation tape	5	Nos.
1.192	Controlled radial expansion balloon dilator, multistage with diameter 6-8 mm, 240 cm, balloon length atleast 5.5 cm, compatible with 2.8 mm biopsy channel. Indicated for esophageal dilatation, distinct diameters at 3 distinct pressures; wire guided	5	Nos.

1.193	Achalasia cardia pneumatic dilatation balloon with 10 cm long balloon, dilatation diameter 25mm, with radio opaque markers	2	Nos.
1.194	Pneumatic inflation device for achalasia cardia balloon, with pressure gauge in psi	2	Nos.
1.195	Percutaneous endoscopic gastrostomy insertion set (pull technique) size 14 with silicone feeding tube, looped inserter wire, bolster kit, drape, syringe, needle. Scalpel, prep tray, cannula	10	Nos.
1.196	PEG low profile replacement tube with antireflux valve, 16 F	10	Nos.
1.197	Colonoscopic biopsy forceps with multiple biopsy sampling function	2	Nos.
1.198	Pediatric endoscopic bite guard with less than 13 mm opening	2	Nos.
1.199	Endoscope tip protector	2	Nos.
1.200	Foreign body net retrieval device, 3X6CM oval shaped, compatible with 2.8 instrument channel scope	2	Nos.
1.201	10 Fr bipolar probe catheter for endoscopic hemostasis, 350 cm, single plug type, compatible with electrosurgical unit	2	Nos.
1.202	Endoscopy biteguard - Paediatric	10	Nos.
1.203	Endoscopy biteguard - Adult	5	Nos.
1.204	Sclerotherapy needle for variceal injection	15	Nos.

1.205	Reagent designed for lysing erythrocytes (RBC) for leucocytes (WBC) counting and differentiation and for hemoglobin determination compatible with HORIBA Microsemi automated cell counter. (Composition EDTA,2H2O-0.06%, Sodium chloride-1.4%, Di-Sodium hydrogen phosphate- 0.3%, Sodium dihydrogen phosphate-0.01%, Dodecyltrimethylammonium chloride (DTACL)- 2.4%)	10	ltrs
1.206	Cleaning reagent compatible with HORIBA Microsemi automated cell counter. (Composition Sodium chloride-0.9%, Preservative -0.09%, Proteolytic enzyme -0.1%, Rocima 553-0.01%, Detergents-0.25%, Calcium chloride -0.0294%, HcL-0.055%)	10	ltrs
1.207	Cleaning reagent compatible with HORIBA Microsemi automated cell counter. (Composition Sodium chloride-0.9%, Preservative -0.09%, Proteolytic enzyme -0.1%, Rocima 553-0.01%, Detergents-0.25%, Calcium chloride -0.0294%, HcL-0.055%)	5	ltrs
1.208	CRP measuring kit compatible with HORIBA Microsemi automated cell counter	8	packs
1.209	JRMATP - Paediatric Circuit with 5mm/22mm Angle Mount, 15 mm Ayre's T, 15 cm X 11 mm with angle Luer lock for suctioning, Corrugated Hose, with 15mm MF Hose Mount pair; 0.5 Ltr. Breathing Bag, Open tail 1.8 mtrs corrugated PVC Tubing for fresh gas supply (Bain Circuit)	50	Nos.

1.210	Tracheal suction catheter - length of atleast 50cm with a single hole at the tip. The connector end should have a side hole (whistle type) for control of suction. Size:6F	50	Nos.
1.211	Tracheal suction catheter - length of atleast 50cm with a single hole at the tip. The connector end should have a side hole (whistle type) for control of suction. Size:8F	50	Nos.
1.212	Tracheal suction catheter - length of atleast 50 cm with a single hole at the tip. The connector end should have a side hole (whistle type) for control of suction. Size:10F	100	Nos.
1.213	Closed endo tracheal suction system - should be suitable for use in neonates and children upto 13 years of age. should have minimal dead space. should be made of transparent tear proof material. Size:6F	50	Nos.
1.214	Closed endo tracheal suction system - should be suitable for use in neonates and children upto 13 years of age. should have minimal dead space. should be made of transparent tear proof material. Size:8F	50	Nos.

1.215	Closed endo tracheal suction system - should be suitable for use in neonates and children upto 13 years of age. should have minimal dead space. should be made of transparent tear proof material. Size:10F	50	Nos.
1.216	Intercostal drainage tube without trocar - Size: 12 F	20	Nos.
1.217	Intercostal drainage tube without trocar - Size: 14 F	50	Nos.
1.218	Intercostal drainage tube without trocar - Size: 16 F	50	Nos.
1.219	Intercostal drainage tube without trocar - Size: 18 F	50	Nos.
1.220	Endotracheal tube with murphy's eye - Uncuff (Size: 3 mm Internal Diameter) (One per pack)	50	Nos.
1.221	Endotracheal tube with murphy's eye - Uncuff (Size: 4mm Internal Diameter) (One per pack)	50	Nos.

1.222	Endotracheal tube with murphy's eye - Uncuff (Size: 4.5mm Internal Diameter) (One per pack)	50	Nos.
1.223	Endotracheal tube with murphy's eye - Uncuff (Size: 5mm Internal Diameter) (One per pack)	50	Nos.
1.224	Endotracheal tube with murphy's eye - Uncuff (Size: 5.5mm Internal Diameter) (One per pack)	50	Nos.
1.225	Endotracheal tube with murphy's eye - Uncuff (Size: 6mm Internal Diameter) (One per pack)	50	Nos.
1.226	Endotracheal tube with murphy's eye - Uncuff (Size: 6.5mm Internal Diameter) (One per pack)	50	Nos.
1.227	Endotracheal tube with Murphy eye with cuffed (Size: 2.5mm Internal Diameter)	25	Nos.
1.228	Endotracheal tube with Murphy eye with cuffed (Size: 3mm Internal Diameter)	200	Nos.

1.229	Endotracheal tube with Murphy eye with cuffed (Size: 3.5mm Internal Diameter)	200	Nos.
1.230	Endotracheal tube with Murphy eye with cuffed (Size: 4mm Internal Diameter)	200	Nos.
1.231	Endotracheal tube with Murphy eye with cuffed (Size: 4.5mm Internal Diameter)	200	Nos.
1.232	Endotracheal tube with Murphy eye with cuffed (Size: 5mm Internal Diameter)	200	Nos.
1.233	Endotracheal tube with Murphy eye with cuffed (Size: 5.5mm Internal Diameter)	200	Nos.
1.234	Endotracheal tube with Murphy eye with cuffed (Size: 6mm Internal Diameter)	200	Nos.
1.235	Endotracheal tube with Murphy eye with cuffed (Size: 6.5mm Internal Diameter)	100	Nos.

1.236	Tracheostomy tube uncuffed (Size 3mm)	10	Nos.
1.237	Tracheostomy tube uncuffed (Size 3.5mm)	10	Nos.
1.238	Tracheostomy tube uncuffed (Size 4mm)	10	Nos.
1.239	Tracheostomy tube uncuffed (Size 4.5mm)	2	Nos.
1.240	Tracheostomy tube uncuffed (Size 5mm)	2	Nos.
1.241	Tracheostomy tube uncuffed (Size 5.5 mm)	2	Nos.
1.242	Tracheostomy tube uncuffed (Size 6 mm)	2	Nos.
1.243	Tracheostomy tube cuffed (Size 4 mm)	15	Nos.
1.244	Tracheostomy tube cuffed (Size 4.5 mm)	15	Nos.
1.245	Tracheostomy tube cuffed (Size 5 mm)	15	Nos.

1.246	Tracheostomy tube cuffed (Size 5.5 mm)	10	Nos.
1.247	Tracheostomy tube cuffed (Size 6 mm)	10	Nos.
1.248	Nebulization set with tubing disposable - Compatible with ALL standard nebulizer compressors(Pediatric size)	500	Nos.
1.249	Oxygen delivery face mask with adapter for oxygen tube - Pediatric Size	100	Nos.
1.250	Autoclavable Silicon self inflating resuscitation bag with safety valve- Pediatric (750 ml) - Should be reusable and autoclavable. Should have oxygen reservoir. Should have a robust safety pressure release valve (pop off valve) for pressure > 40 cm H2O. The inlet outlet and expiratory valves should be durable and should not have any leaks. The equipment should be able to withstand a fall from a height of 5 feet. Should have a facility for attaching a PEEP valve (desirable)	50	Nos.
1.251	Autoclavable Silicon self inflating resuscitation bag with safety valve -Adult 1500 ml - Should be reusable and autoclavable. Should have oxygen reservoir. Should have a robust safety pressure release valve (pop off valve) for pressure > 40 cm H2O. The inlet outlet and expiratory valves should be durable and should not have any leaks. The equipment should be able to withstand a fall from a height of 5 feet. Should have a facility for attaching a PEEP valve (desirable).	30	Nos.

1.252	Autoclavable Silicon self inflating resuscitation bag with safety valve - Pediatric 300 ml - Should be reusable and autoclavable. Should have oxygen reservoir. Should have a robust safety pressure release valve (pop off valve) for pressure > 40 cm H2O. The inlet outlet and expiratory valves should be durable and should not have any leaks. The equipment should be able to withstand a fall from a height of 5 feet. Should have a facility for attaching a PEEP valve (desirable).	50	Nos.
1.253	Double Lumen Paediatric Central Venous Catheter - 3.0 F with Length 5 cm - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable). (Number should be supply as per request).	150	Nos.
1.254	Double Lumen Paediatric Central Venous Catheter - 4.0 F with Length 5 cm and 4 F - Length 8 cm. - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable). (Number should be supply as per request).	100	Nos.
1.255	Double Lumen Paediatric Central Venous Catheter - 5.0 F with Length 8 cm. - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable).	100	Nos.
1.256	Double Lumen Paediatric Central Venous Catheter - 7.0 F with Length 15 cm. - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable).	100	Nos.

1.257	Triple Lumen Paediatric Central Venous Catheter - 4.0 F with Length 5 cm and 4.0F with Length 8 cm. - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable). (Number should be supply as per request).	100	Nos.
1.258	Triple Lumen Paediatric Central Venous Catheter - 5.0 F with Length 8 cm. - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable).	100	Nos.
1.259	Triple Lumen Paediatric Central Venous Catheter - 5.5 F with Length 8 cm and 5.5 F with Length 13 cm. - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable). (Number should be supply as per request).	100	Nos.
1.260	Triple Lumen Paediatric Central Venous Catheter - 7.0 F with Length 15 to 16 cm. - Seldinger technique based. Puncture needle / introducer needle. Appropriate spring guide wire and dilator. Catheter clamp and fastener. Suture or suture less fixation should be mentioned clearly. Site connector and ECG lead (desirable).	50	Nos.
1.261	Disposable ventilator circuit (paediartic) compatible with Fischer and Paykel MR 850 Humidifier with heater wire in inspiratory limb. Sample demonstration is mandatory.	500	Nos.
1.262	Disposable ventilator circuit (Neonatal) compatible with Fischer and Paykel MR 850 Humidifier with heater wire in inspiratory limb. (Sample demonstration is mandatory)	500	Nos.

1.263	Pediatric Heat Moisture Exchange (HME) Filter - Should provide a combination of high-efficiency heat/moisture exchangers (HME) and bacterial viral filtration, rated to 0.027 microns at 99.999% or higher efficiency. Moisture output: 30 mg/H ₂ O @ tidal volume of 6 to 8 ml/kg. Weight should be less than 20g. Dead space less than 30 ml. Resistance (cmH ₂ O) should be less than 225pa @ 30 LPM. Filtration media type should be electrostat	2000	Nos.
1.264	Disposable ECG Electrodes - Neonatal Size	2000	Nos.
1.265	Disposable ECG Electrodes - Pediatric Size	2000	Nos.
1.266	High Pressure Extension Line - 150 CM Length	5000	Nos.
1.267	High Pressure Extension Line - 100 CM Length	5000	Nos.
1.268	Sodium Bicarb packets IP 600 gms & Sodium Chloride 235 gms	10	packets
1.269	Acid solution with dextrose per liter (Pack of 5 liter)	50	litres
1.270	Glucostrips to measure blood sugar value at bedside. - Compatible glucometer should be provided at free of cost.	20000	Nos.
1.271	Cover Glass 22x22mm	550	Nos.
1.272	Cover Glass 22x24mm	550	Nos.
1.273	Micro slide 75x25mm.As per ISO 8037	1000	Nos.
1.274	Non Vaccum tubes EDTA K2 TUBE 2 ml	20000	Nos.
1.275	Automatic Kidney biopsy gun 18G - 16 cm (Pediatrics)	80	Nos.
1.276	Bone marrow biospy needle pediatric size (16 G)	200	Nos.
1.277	Bone marrow biospy needle pediatric size (11 G)	100	Nos.

Samples should be given to storeroom, Department of Pediatrics within ten days of the closing of online submission of bids failing which the bids will be rejected.

Section-III

GENERAL INSTRUCTIONS TO TENDERERS

1. Period of contract:

The contract shall initially be for a period of **one year** that may be extended to **6 months**. The rates approved shall remain unchanged during the period of contract.

2. Bid Security (Earnest Money Deposit)

- i. In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender.
- ii. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.
- iii. The bidder hereby declares that they accept the condition that if they withdraw or modify their bids during period of validity etc., they will be suspended for a period of two years from participating in any tender invited by JIPMER, Puducherry with effect from the date of their withdrawal or modification of their bid.

3. Eligibility conditions of bidders:

- i. The Tenderer must be a domestic manufacturer that is a 'Class-I local supplier' or a 'Class-II local supplier' as defined under Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as further amended by orders of even number dated 28.05.2018, 29.05.2019, 04.06.2020 and 16.09.2020. In case the manufacturer does not quote directly, they may authorize an agent as per proforma of Manufacturer authorization form as given in the Tender enquiry document to quote and enter into a contractual obligation.
- ii. In compliance with order (Public Procurement No.1) No. 6/18/2019-PPD dated 23rd July 2020 issued by the Public Procurement Division, Dept. of Expenditure, Min of Finance under Rule 144(xi) of GFR 2017 any bidder 'from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority.
 - a. "Bidder" (including the term 'tenderer', 'consultant' or 'service provider' in certain contexts) means any person or firm or company, including any member of a consortium or joint venture (that is an association of several persons, or firms or companies), every artificial juridical person not falling in any of the descriptions of bidders stated hereinbefore, including any agency branch or office controlled by such person, participating in a procurement process.
 - b. "Bidder from a country which shares a land border with India" for the purpose of this Order means: -

- I. An entity incorporated, established or registered in such a country; or
- II. A subsidiary of an entity incorporated, established or registered in such a country; or
- III. An entity substantially controlled through entities incorporated, established or registered in such a country; or
- IV. An entity whose beneficial owner is situated in such a country; or
- V. An Indian (or other) agent of such an entity; or
- VI. A natural person who is a citizen of such a country; or
- VII. A consortium or joint venture where any member of the consortium or joint venture falls under any of the above

c. The beneficial owner for the purpose of (5.2.2) above will be as under:

- I. In case of a company or Limited Liability Partnership, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has a controlling ownership interest or who exercises control through other means.

Explanation—

- "Controlling ownership interest" means ownership of or entitlement to more than twenty-five per cent, of shares or capital or profits of the company;
 - "Control" shall include the right to appoint majority of the directors or to control the management or policy decisions including by virtue of their shareholding or management rights or shareholders agreements or voting agreements;
- II. In case of a partnership firm, the beneficial owner is the natural person(s) who, whether acting alone or together, or through one or more juridical person, has ownership of entitlement to more than fifteen percent of capital or profits of the partnership.
 - III. In case of an unincorporated association or body of individuals, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has ownership of or entitlement to more than fifteen percent of the property or capital or profits of such association or body of individuals.
 - IV. Where no natural person is identified under (a) or (b) or (c) above, the beneficial owner is the relevant natural person who holds the position of senior managing official;
 - V. In case of a trust, the identification of the beneficial owner(s) shall include the identification of the author of the trust, the trustee, the beneficiaries with 15% or more interest in the trust and any other natural person exercising ultimate effective control over the trust through a chain of control or ownership.

d. An Agent is a person employed to do any act for another, or to represent another in dealings with third person.

e. In case of turnkey contracts, the successful bidder shall not be allowed to sub-contract works to any contractor from a country which shares a land border with India unless such contractor is registered with the Competent Authority.

- f. Bidders must certify compliance with the above mentioned provisions in the tender form as per Section VII.
- g. The bidder should have successfully completed the delivery of the item of at least 25% of the quantity required in this tender to a government hospital in the last 3 years.
- h. The bidder or the manufacturer must have an average annual turnover of at least Rs.1,00,00,000/-(Rupees one crore only) each during the last three financial years i.e. 2017-18, 2018-19, 2019-20.
- i. Bidder should have ISO Certification.

4. Purchase Preference

- i. The Procurement of goods and services under this e-tender will be regulated as per the applicable provisions of Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as further amended by orders of even number dated 28.05.2018, 29.05.2019, 04.06.2020 and 16.09.2020. Salient portions of the order are reproduced in Appendix A of this tender document by way of information. Bidders are advised to see the original orders and satisfy themselves that they qualify to participate in the tender. Bidders who are claiming eligibility to participate in this tender must submit a certificate in format given in Appendix A along with documentary evidence in support of their claim wherever necessary failing which their bid will be summarily rejected. The purchaser reserves the right to give preference to the 'Class-I local supplier'.
- ii. The Purchaser reserves the right to give the purchase preference to small-scale sectors, Micro and small scale enterprises etc. as per the instruction in vogue while evaluating, comparing and ranking the responsive Tenders as given in the MSMED Act 2006 reproduced below:
 - a. In exercise of powers conferred in Section 11 of the Micro, Small and Medium Enterprises Development (MSMED) Act 2006, the Government has notified a new Public Procurement Policy for Micro & Small Enterprises effective from 1st April 2012. The policy mandates that 25% of procurement of annual requirement of goods and services by all Central Ministries/Public Sector Undertakings will be from the micro and small enterprises. The Government has also earmarked a sub-target of 4% procurement of goods & services from MSEs owned by SC/ST entrepreneurs out of above said 25% quantity.
 - b. In accordance with the above said notification, the participating Micro and Small Enterprises (MSEs) in a tender, quoting price within the band of L1+15% would also be allowed to supply a portion of the requirement by bringing down their price to the L1 price, in a situation where L1 price is from someone other than an MSE. Such MSEs would be allowed to supply up to 25% of the total tendered value. In case there are more than one such eligible MSE, the 25% supply will be shared equally. Out of 25% of the quantity earmarked for supply from MSEs, 5% quantity is earmarked for procurement from MSEs owned by SC/ST entrepreneurs. However, in the event of failure of such MSEs to participate in the Tender process or meet the

tender requirements and the L1 price, the 5% quantity earmarked for MSEs owned by SC/ST entrepreneurs will be met from other participating MSEs.

- c. The MSEs fulfilling the prescribed eligibility criteria and participating in the Tender shall enclose with their tender a copy of their valid registration certificate with District Industries Centres or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir Board or National Small Industries Corporation or any other body specified by Ministry of Micro and Small enterprises in support of their being an MSE, failing which their tender will be liable to be ignored.
- d. Special provision for Micro and Small Enterprise owned by women: – Out of the total annual procurement from Micro and Small Enterprises, 3 per cent from within the 25 per cent target shall be earmarked for procurement from Micro and Small Enterprises owned by women.
- e. Note: “If the bidder is a MSME, it shall declare in the bid document the Udyog Aadhar Memorandum Number issued to it under the MSMED Act, 2006. If a MSME bidder do not furnish the UAM Number along with bid documents, such MSME unit will not be eligible for the benefits available under Public Procurement Policy for MSEs Order 2012.

5. Demonstration of sample

The tenderer must give a sample within **ten** days of the closing of online submission of bids failing which the bids will be rejected.

6. Instructions for the filling the tender form

- i. E-Tender form shall be completed in all respect, signed in full and stamped at appropriate places and initialed and stamped on all remaining pages. Incomplete or e-tenders without tender processing fee, bid security declaration, Make-in-India self-certification as in Appendix-A only by original manufacturer, GFR 144 (xi) compliance certificate as in Section-X shall be treated as invalid.
- ii. Bidders have to ensure that all the documents are properly filled.
- iii. Conditional tenders are liable to be rejected.
- iv. Bids received and found valid will be evaluated by JIPMER to ascertain the complete work/services under the specification and documents. The bidder should take care to submit all the information sought by JIPMER in prescribed formats.
- v. Incomplete bids, bids in paper format, conditional bids, telephonic bids or tenders submitted after the due date and time will not be considered and summarily rejected. Vendors are, therefore, advised to submit their bids well on time.
- vi. **The bidder can quote for one or more items mentioned in the list. Bidder has to give all details (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) mentioned in BOQ for all quoted items, failure of that the bid will be rejected summarily.**

7. Submission of tenders

The bidders must ensure that they submit the **on-line bids** within the scheduled closing date & time.

8. Late Tender:

There is NO PROVISION of uploading late tender beyond stipulated date & time in the e-tendering system.

9. Alteration and Withdrawal of Tender

- i. The bidder, after submitting its bid, is permitted to alter/modify its bid, within the deadline for submission of bids. Alterations/modifications to bids received after the prescribed deadline will not be possible on the e-tender portal.
- ii. No tender should be withdrawn or modified after the deadline for submission of tender and before expiry of the tender validity period. If a bidder withdraws or modifies the tender during this period, it will result in the tenderer being suspended from participating in all tenders invited by JIPMER for a period of two years with effect from the date of withdrawal, as per the terms and conditions of bid security declaration furnished by the bidder in its tender.

10. Preparation of e-tenders

This is a Two-Bid Tender system, consisting of the **Techno-Commercial Bid and Price Bid** that are to be uploaded in the prescribed formats in the e-tendering portal. The tender(s) shall only be submitted online as mentioned below:

I. **Techno-commercial Bid shall comprise**

- a. Fee Cover
 - i. **E-tender Processing fee** of Rs.590/- (Rupees five hundred and ninety only) inclusive of 18% GST payment **receipt** duly self-attested and rubber stamped should be uploaded.
 - ii. Scanned copy (100 or 200 DPI) in pdf format of **Bid Security Declaration** as per format attached in Section-IX
- b. In the cover named “Prequal/Technical” the scanned copy in pdf format of the following documents are to be uploaded:
 - i. The **Bidder’s Profile** as in Section-VI, **Tender Form** as in Section-VII must be downloaded duly filled signed and stamped, and an **Authorization letter** for signing tender documents if a person other than the Owner, Partner, Managing Director is signing the documents, must be uploaded as a single PDF file.

- ii. **Copies of Supply orders/Completion certificate** in support of Eligibility condition 1 and 2 and **Copies of abridged Annual report of last 03 years** (Income tax return acknowledgement, Assets and Liabilities, Balance sheet and Profit & Loss Account) must be uploaded as a single PDF file
- iii. **Manufacturer's Authorization letter** in company letterhead in format in given Section-VI and, in case the bidder is empaneled by the Competent Authority under **GFR 144 (xi) (mandatory)** a copy of the same or **GFR 144(xi) compliance certificate** as in Section-X must be uploaded.
- iv. Copy of Self Certified **GST registration certificate** and Copy of **PAN Card** must be uploaded as a single PDF file.
- v. **Bank Details** (Beneficiary name, Bank name, Account number, IFSC code, Branch address on letterhead) and **Check list** as in Section-VIII in the prescribed format duly filled and signed must be uploaded as a single PDF file.
- vi. Self-certification for supporting the claim to be a local supplier under the "Public Procurement preference to Make in India" order in format as in **Appendix-A (mandatory)** and any other document that the bidder wishes to submit as a single PDF file.
- vii. **A file mentioning the list of items without price** for which bidder is quoting must be uploaded as a single PDF file.

II. Price Bid:

Prices are to be quoted in the prescribed Price Bid format provided in the e-tender portal using the BOQ template only. The price should be quoted for the **accounting unit** indicated in the e-tender document.

Note:

- i) The bidder has to be diligent while filling up the Techno-Commercial Bid and Price Bid provided in prescribed formats and must not tamper with the contents of the sheets.
- ii) Bidders must ensure that the documents uploaded in pdf format are legible.
- iii) It is the responsibility of bidder to go through the Tender document to ensure furnishing all required documents in addition to above, if any.
- iv) The Make in India self-certification as in Appendix-A must be given and uploaded by original manufacture. The distributor can upload the Appendix-A given by original manufacture.
- v) ITE- Item-wise Eligibility Sheet should be downloaded, the items that the bidder wishes to quote must be selected as "Eligible", and this "ITE file" must also be uploaded for the price bid to be considered by the system. The selected items will be displayed once uploaded and the bidder can verify that all items he wishes to quote for, are present in the list
- vi) A person signing (manually or digitally) the tender form or any documents forming

part of the contract on behalf of another shall be deemed to warrant that he has authority to bind such other persons and if, on enquiry, it appears that the persons so signing had no authority to do so, the purchaser may, without prejudice to other civil and criminal remedies, cancel the contract and hold the signatory liable for all cost and damages.

- vii) A bid, which does not fulfill any of the above requirements and/ or give evasive information/reply against any such requirement, shall be liable to be ignored.
- viii) Tender sent by fax/telex/cable shall be ignored.

11. Digital Signing of Tender

The tenderers shall submit their tenders as per the instructions contained as above. Tenders shall be uploaded with all relevant tender documents in the prescribed format. The relevant tender documents should be uploaded by an authorized person having Class 3 digital signature certificate.

12. Tender currencies.

The tender shall be quoted only in INR.

13. Additional information and instruction on GST:

If the Tenderer desires to ask for GST or any other taxes to be paid extra, the same must be specifically stated. In the absence of any such stipulation, the price will be taken inclusive of such taxes and no claim for the same will be entertained later. The rate of GST quoted in the tender shall be taken for price comparison. However, the rate of GST quoted in the tender or the actual rate of GST applicable, whichever is lower shall be payable by the purchaser. The supplier can charge a higher GST than quoted in the tender only if the rate of GST was revised by the government after the tender closing date.

14. Bid Security (Earnest Money Deposit)

- i. In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit^{tl}, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender.
- ii. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration^l on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.
- iii. The bidder hereby declares that they accept the condition that if they withdraw or modify their bids during period of validity etc., they will be suspended for a period of two years from participating in any tender invited by JIPMER, Puducherry with effect from the date of their withdrawal or modification of their bid.

15. Tender opening

- i. The Tender Inviting Authority will open the e-tenders at the specified date and time and at the specified place as indicated in the NIT. In case the specified date of tender opening falls on / is subsequently declared a holiday or closed day for the Tender Inviting Authority, the tenders will be opened at the appointed time and

place on the next working day.

- ii. Authorized representatives of the tenderers, who have submitted tenders on time, may attend the tender opening provided they bring with them letters of authority from the corresponding tenderers. The tender opening official(s) will prepare a list of the representatives attending the tender opening. The list will contain the representatives' names & signatures and corresponding tenderers' names and addresses.
- iii. This being a Two-Bid Tender system, the **Techno-Commercial Tenders** are to be opened in the first instance, at the prescribed time and date as indicated in NIT. These Tenders shall be scrutinized and evaluated by the competent committee/ authority with reference to parameters prescribed in the tender document. During the Techno- Commercial Tender opening, the tender opening official(s) will read the salient features of the tenders like brief description of the goods offered, delivery period, bid security declaration and any other special features of the tenders, as deemed fit by the tender opening official(s). Thereafter, in the second stage, the Price bids of only the Techno - Commercially acceptable offers (as decided in the first stage) shall be opened for further scrutiny and evaluation on a date notified after the evaluation of the Techno-Commercial tender.

16. Scrutiny and evaluation of tenders

A. Basic Principle

Tenders will be evaluated on the basis of the terms & conditions already incorporated in the tender enquiry document, based on which tenders have been received and the terms, conditions etc. mentioned by the tenderers in their tenders. No new condition will be brought in while scrutinizing and evaluating the tenders.

B. Scrutiny of Tenders

- i. The Purchaser will examine the Tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished and, whether the documents uploaded are in legible form.
- ii. The Purchaser's determination of a Tender's responsiveness is to be based on the contents of the tender itself without recourse to extrinsic evidence.
- iii. The tenders will be scrutinized to determine whether they are complete and meet the essential and important requirements, conditions etc. as prescribed in the tender document. The tenders, which do not meet the basic requirements, are liable to be treated as non-responsive and will be summarily ignored.
- iv. The following are some of the important aspects, for which a tender shall be declared non-responsive during the evaluation and will be ignored;
 - i. Tender validity is shorter than the required period.
 - ii. Non-submission of Bid Security Declaration.
 - iii. Non-submission of receipt of tender processing fee.
 - iv. Non-submission of self-certification in format as given in Appendix-A only by original manufacturer, for determining eligibility to participate in the tender under the "Public Procurement preference to Make in India" order.
 - v. Non-submission of GFR-144 (xi) compliance certificate.

- vi. Tenderer has not agreed to give the required performance security of required amount in an acceptable form.
- vii. Non-submission of samples within ten days of the closing of online submission of bids
- viii. Poor/ unsatisfactory past performance.
- ix. Tenderers who stand de-registered/banned/blacklisted by any Central Govt. Ministries/Departments/Hospitals/Institutes.
- x. Tenderer is not eligible as per tender conditions.
- xi. Tenderer has not quoted for the entire quantity as specified in the List of Requirements/ BOQ for the item quoted.
- xii. Non-submission of all details of quoted items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark).
- xiii. Tenderer has not agreed to other essential condition(s) specially incorporated in the tender enquiry, like delivery terms, delivery schedule, terms of payment, liquidated damages clause, warranty clause, dispute resolution mechanism applicable law.

17. Minor Informality/Irregularity/Non-Conformity

If during the preliminary examination, the purchaser find any minor informality and/or irregularity and/or non-conformity in a tender, the purchaser may waive the same provided it does not constitute any material deviation and financial impact and, also, does not prejudice or affect the ranking order of the tenders. Wherever necessary, the purchaser will convey its observation on such 'minor' issues to the tenderer by registered/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.

18. Award of work

- i. The selection of the agency will be at the sole discretion of the JIPMER who reserves its right to accept or reject any or all the proposals without assigning any reason thereof.
- ii. The lowest bid will be evaluated on the basis of annual estimated amount as quoted by bidders in the quotation including taxes.
- iii. Upon evaluation of offers the decision on the award of contract will be intimated to the successful bidder.
- iv. The annual estimate is given only as an indication. The actual quantity procured may increase or decrease. No assurance is given that the quantity stated will actually be procured.

Section-IV

GENERAL CONDITIONS OF CONTRACT

1. **Price of goods**

The rate quoted in the e-tender will be fixed for the whole contract period.

2. **Technical Specifications and Standards**

The Goods & Services to be provided by the supplier under this contract shall conform to the technical specifications mentioned in “Technical Specification” under Section II.

3. **Terms of Delivery**

- i. Goods shall be delivered by the supplier within 30 days of issue of supply order. Please note that the time shall be the essence of the contract.
- ii. Any unexcused delay by the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all of the following sanctions:
 - a) Imposition of liquidated damages,
 - b) Forfeiture of its performance security and
 - c) Termination of the contract for default.

4. **Liquidated Damages**

If the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods, installation, commissioning and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 11. Since the Liquidated damages are in virtue of non-performance of services, it will attract GST or any other applicable taxes which in turn shall be deducted from the bidder.

5. **Performance Security**

The successful bidder shall have to deposit an amount 3% of the value of contract as Performance Security Deposit (PSD) within two weeks after award of contract, through SBI Collect available on JIPMER website. In the event of any failure /default of the supplier with or without any quantifiable loss to the purchaser, the amount of the performance security is liable to be forfeited.

Subject to the condition mentioned above the Performance Security will be released without any interest to the supplier on completion of the supplier's all contractual obligations

including the warranty obligations and extension of time (with or without Liquidated Damages).

6. Payment Procedure

The contractor shall submit bill in triplicate upon satisfactory supply of goods. Bill must be raised based on the rate quoted in e-tender. Every effort shall be made to ensure that the payment will be made within 45 days of submission of bill. No advance payment will be considered. TDS will be deducted as per provision of Income Tax Act, GST Acts and other statutes as relevant.

7. Risk Clause

The contractor shall at all times have standby arrangements for carrying out the work under the contract, in case of any failure of the existing arrangements. JIPMER reserves the right for termination of the contract at any time by giving 30 days written notice, if the items delivered are found to be unsatisfactory and also has the right to award the contract to the next higher bidder willing to supply the item at the cost, risk and responsibilities of contractor and excess expenditure incurred on account of this will be recovered by JIPMER from the contractor's Performance Security Deposit or pending bills or by raising a separate claim.

8. Termination clause:

During the period of agreement if it is found that the agency is not providing proper services, the JIPMER reserves rights to make the vendor forfeit the security deposit deposited with JIPMER or part thereof in favour of JIPMER and agreement will be terminated after giving 30 days' notice. Furthermore, in such situations, tender can be allotted to second lowest bidder and the difference in cost shall be recovered from the earlier vendor who is breach of the contract. In addition in case it is found that the supplier is charging by fraudulent means or indulging in criminal activities the contract will be terminated immediately.

9. Jurisdiction of the courts

Jurisdiction of the courts for settlement of disputes:- Jurisdiction for the settlements of disputes if any is Puducherry only.

In above mentioned conditions Director JIPMER reserves all the rights.

Section-V

BIDDER'S PROFILE

This form duly filled and signed by authorized representative of the bidder and the scanned copy must be uploaded online

1.	Name & Designation of the contact person	
2.	Name and Address of the Tenderer	
3.	Phone No a) Land line number (functional between 9 am and 5pm)	
4.	Mobile No of contact person (available from 9am to 6pm)	
5.	Email ID of the Tenderer	
6.	Email ID of the contact person	
7.	Local supplier/Distributor in Chennai/ Puducherry or any other place (complete address must be written)	
8.	Manufacture Name	
9.	Manufacture Address	
10	Whether Tenderer is registered MSE Manufacture the product quoted. (If registered MSE, submit copy of the Udyog Aadhaar certificate or Equivalent Certificate)	Yes / No
<p>If there is any change in the above details, I will immediately intimate you by speed post or fax or email</p> <p>I.....hereby declare that the details given above are true to the best of my knowledge and I have thoroughly read and understood the terms and conditions of the tender and shall abide by the rules,</p> <p style="text-align: right;">Signature (Name and Designation & Seal)</p> <p>Dated:</p>		

NB: This form must be duly filled in by an authorized person

Section-VI
MANUFACTURER'S AUTHORISATION FORM
(Letterhead)

The Director,
JIPMER, Puducherry

Dear Sir,

Ref: Your TE document No: _____ dated: _____

We, _____ who are proven and reputable manufacturers of _____ (name and description of the goods offered in the Tender) having factories at _____, hereby authorize Messrs. _____ (name and address of the agent) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also state that we are not participating directly in this tender for the following reason(s):
_____. (Please provide reason here).

We further confirm that no supplier or firm or individual other than Messrs _____ (name and address of the above agent) is authorised to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us. We also hereby extend our full warranty, CMC/AMC as applicable as per the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document. We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorised agent.

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly”

Yours faithfully,
[Signature with date, name, designation and Email]
for and on behalf of Messrs _____
[Name & address of the manufacturers]

Note:

- (1) This letter of authorization should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.
- (2) Original letter may be sent.
- (3) The purchaser reserves the right to verify this document with its signatory.

SECTION-VII
TENDER FORM
(On Firms' Letter Head)

To
The Director
JIPMER, Puducherry 605006

Date _____

Ref. Your TE document No. _____ dated _____

We, the undersigned have examined the above mentioned TE document, including amendment/corrigendum No. _____, dated _____ (if any), the receipt of which is hereby confirmed. We now offer to supply and deliver _____ (Description of goods and services) in conformity with your above referred document **for the sum as shown in the price schedules attached herewith and made part of this tender**. If our tender is accepted, we undertake to supply the goods and perform the services as mentioned above, in accordance with the delivery schedule specified in the List of Requirements.

We further confirm that, if our tender is accepted, we shall provide you with a performance security of required amount in an acceptable form in terms of GCC, Special Conditions of Contract", for due performance of the contract.

We have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and we certify that this bidder is not from such a country/ from such a country and has been registered with the Competent Authority and a copy of the valid registration by the Competent Authority is attached as evidence of the same (Strike out what is not applicable). In case there are Turnkey works to be carried out this bidder will not sub-contract any work to a contractor from such countries unless such contractor is registered with the Competent Authority. We hereby certify that this bidder fulfills all requirements in this regard and is eligible to be considered.

We agree to keep our tender valid for acceptance as required in the GIT, Special Instructions to Tenderers" or for subsequently extended period, if any, agreed to by us. We also accordingly confirm to abide by this tender up to the aforesaid period and this tender may be accepted any time before the expiry of the aforesaid period. We further confirm that, until a formal contract is executed, this tender read with your written acceptance thereof within the aforesaid period shall constitute a binding contract between us.

We further understand that you are not bound to accept the lowest or any tender you may receive against your above-referred tender enquiry.

We confirm that we do not stand deregistered/banned/blacklisted by any Govt. Authorities in the last 7 years.

We confirm that we fully agree to the terms and conditions specified in above mentioned TE document, including amendment/ corrigendum if any.

(Signature with date)
(Name and designation)

Duly authorized to sign tender for and on behalf of

Section-VIII

CHECK LIST FOR SUBMISSION OF TENDER

(To be filled by the tenderer and submitted along with the bid)

Sl. No.	Particular	Yes / No
1	i. Tender Processing Fee (Mandatory for all bidders) ii. Bid security declaration	
2	Copies of Supply orders/Completion certificate in support of Eligibility condition 2(ii).	
3	Copy of PAN Card	
4	Copy of ISO Certification of bidder, ISI/CE Certification of each product wherever asked	
5	Copies of last three years Income Tax Return statement with Balance sheets & Profit & Loss A/c i.e. 2017-18, 2018-19, 2019-20.	
6	Copy of GST Registration Certificate	
7	Authorization letter for signing tender documents if a person other than the Owner, Partner, Managing Director is signing/uploading the tender	
8	Tender form as in Section – VII duly signed and stamped	
9	Manufacturer's authorization form in format as in Section VI and	
10	Self-certification for supporting the claim to be a local supplier under the "Public Procurement preference to Make in India" order in format as in Appendix A	
11	A copy bidder's empanelment by the Competent Authority under GFR 144 (xi) or GFR 144 (xi) compliance certificate	
12	A PDF file containing list of all items quoted by the bidder without price bid in technical cover	
13	All details of the items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) quoted by the bidder.	
14	Any other document(s) enclosed (To be specified)	

I/We certify that the information furnished above is true and correct. The terms and conditions are acceptable to us and have the authority to bid a tender.

Signature of the owner/
Managing Partner/Director

Date:
Place:

Name:
Seal:

SECTION – IX

**Bid Security Declaration
(To printed on the Firm's letterhead)**

(In compliance with the OM No. F.9/4/2020-PPD dated 12.11.2020 on the subject —Bid Security/ Earnest Money Deposit, issued by the Procurement Policy Division, Department of Expenditure, Ministry of Finance, Government of India, there is no need for bidders to pay Bid Security/ Earnest Money Deposit for participating in this tender. However, in lieu of Bid Security, the bidder must print this —Bid Security Declaration on his/her firm's letterhead duly sign the undertaking and upload the document in the fee cover.)

I/We hereby declare that we accept the condition that if I/We withdraw or modify our bids during period of validity etc., we will be suspended for a period of two years from participating in all tenders invited by JIPMER, Puducherry with effect from the date of our withdrawal or modification of our bid.

(Signature with date)

(Name and designation)

Duly authorized to sign this tender and declaration

for and on behalf of _____ (Name of the Bidder's Firm) Seal

Appendix-A

Self-certification format for claiming purchase preference under the “Public Procurement preference to Make in India” order

As per the order issued by

(i) Department of Industrial Policy and Promotion (DIPP) vide No. P-45021/2/2017-BE-II dated 15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017-B.E.-II dated 29.05.2019, Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020 and Order No. P-45021/2/2017-PP (BE-II) dated 16.09.2020; and

(ii) Department of Pharmaceuticals vide No. F- 31026/36/2016-MD dated 18.05.2018 and the subsequent orders thereof; The purchaser reserves the right to give preference to the local supplier.

A local supplier (definition of “local supplier” is given in clause 2 of the aforesaid order of DIPP as amended from time to time) has to submit the following along with their e-tender(s) failing which their bid will be evaluated without considering such preference mentioned in the DIPP order dated 15.06.2017 further amended on 28.05.2018, 25.09.2019 and 04.06.2020:

a. The local supplier at the time of e-tender, bidding or solicitation shall be required to provide self-certification that the item offered meets the minimum local content and shall give details of the location(s) at which the local value addition is made in the format in Annexure A.

“Certified that the following items quoted has more than 50% or 20% to 50% of value addition in India at the location(s) mentioned against each and is eligible for purchase preference as per the Govt. of India “Public Procurement preference to Make in India” order Dt.15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017-B.E.-II dated 29.05.2019, and Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020.

Sl. No	Name of the item	Details of the location(s) at which the local value addition was made.

Authorized Signature:
Name:
Designation:

b. In cases of procurement for a value in excess of Rs. 10 Crore. the local supplier shall be required to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.

c. Minimum Local Content: ‘Class-I local supplier’ means a supplier or service provider, whose goods, services or works offered for procurement, has local content equal to or more than 50%, as defined

under the Order. “Class-II local supplier” means a supplier or service provider, whose goods, services or works offered for procurement, has local content more than 20% but less than 50%, as defined under this Order.

d. Margin of Purchase Preference: The margin of purchase preference shall be 20%.

e. Manufacture under license/technology collaboration agreements with phased indigenization are exempted from meeting the stipulated local content if the product is being manufactured in India under a license from a foreign manufacturer who holds intellectual property rights and where there is a technology collaboration agreement/transfer of technology agreement for indigenous manufacture of a product developed abroad with clear phasing of increase in local content

f. Decisions on complaints relating to implementation of this Order shall be taken by the competent authority which is empowered to look into procurement-related complaints relating the procuring entity.

g. A constituted committee with internal and external experts will examine for independent verification of self-declarations and auditor’s/accountant’s certificates on random basis and in the case of complaints.

h. A fees of Rs.10000/- in the form of demand draft favoring The Director, JIPMER, payable at Puducherry, is required to be deposited with complaints for verification of local content.

i. False declarations will be breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.

j. A supplier who has been debarred by any procuring entity for violation of this Order shall not be eligible for preference under this Order for procurement by any other procuring entity for the duration of the debarment. The debarment for such other procuring entities shall take effect prospectively from the date on which it comes to the notice of other procurement entities.

Officer in Charge
Purchase Section
JIPMER Puducherry
For Director, JIPMER

SECTION – X

**GFR-144 (xi) compliance certificate
(To printed on the Firm's letterhead)**

Tender No:

GFR-144(xi) compliance certificate (as per order F.No. 6/18/2019-PPD, Ministry of Finance, GOI)

I have read the clauses regarding restrictions under GFR144(xi) on procurement from a bidder of a country which shares a land border with India. I certify that, the vendor

- is not such a country
- is from a country and has been registered with a competent authority (attached evidence of valid registration).

(Select one of the above and strike off the other)

I hereby certify that we fulfill all requirement in this regard and is eligible to be considered for the procurement on CPP portal.

Thanking you.

Authorized Signatory